

Surrogate Summary

SW-846

SDG No.: Q2033

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168048BL	PB168048BL	2-Fluorophenol	150	133	88		15 (10)	110 (139)
		Phenol-d6	150	122	81		15 (10)	110 (134)
		Nitrobenzene-d5	100	76.1	76		30 (49)	130 (133)
		2-Fluorobiphenyl	100	76.8	77		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	145	97		15 (44)	110 (137)
		Terphenyl-d14	100	90.1	90		30 (48)	130 (125)
PB168048BS	PB168048BS	2-Fluorophenol	150	122	82		15 (10)	110 (139)
		Phenol-d6	150	117	78		15 (10)	110 (134)
		Nitrobenzene-d5	100	69.5	69		30 (49)	130 (133)
		2-Fluorobiphenyl	100	74.5	74		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	139	93		15 (44)	110 (137)
		Terphenyl-d14	100	76.9	77		30 (48)	130 (125)
PB168048BSD	PB168048BSD	2-Fluorophenol	150	128	85		15 (10)	110 (139)
		Phenol-d6	150	120	80		15 (10)	110 (134)
		Nitrobenzene-d5	100	75.8	76		30 (49)	130 (133)
		2-Fluorobiphenyl	100	78.1	78		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	146	97		15 (44)	110 (137)
		Terphenyl-d14	100	79.4	79		30 (48)	130 (125)
Q2033-01	TW-WTS-08	2-Fluorophenol	150	68.4	46		15 (10)	110 (139)
		Phenol-d6	150	40.9	27		15 (10)	110 (134)
		Nitrobenzene-d5	100	79.2	79		30 (49)	130 (133)
		2-Fluorobiphenyl	100	76.9	77		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	157	104		15 (44)	110 (137)
		Terphenyl-d14	100	86.2	86		30 (48)	130 (125)