

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/19/2025

Lab File ID: BP024686.D Time Analyzed: 12:48

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	148955	7.658	651543	10.43	426043	14.29
UPPER LIMIT	297910	8.158	1303090	10.928	852086	14.792
LOWER LIMIT	74477.5	7.158	325772	9.928	213022	13.792
EPA SAMPLE NO.						
01 TW-WTS-08	192352	7.66	773092	10.43	505969	14.29

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.