

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/19/2025

Lab File ID: BP024686.D Time Analyzed: 12:48

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	875686	17.092	872281	21.527	779874	24.81
UPPER LIMIT	1751370	17.592	1744560	22.027	1559750	25.31
LOWER LIMIT	437843	16.592	436141	21.027	389937	24.31
EPA SAMPLE NO.						
01 TW-WTS-08	1051810	17.09	1193980	21.52	1245480	24.80

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.