

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG No.: Q2033

Instrument ID: MSVOA_X Calibration Date/Time: 05/14/2025 09:59

Lab File ID: VX046181.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Methyl tert-butyl Ether	2.079	2.130		2.45	20
Carbon Tetrachloride	0.544	0.568		4.41	20
Chloroform	1.271	1.302		2.44	20
1,1,1-Trichloroethane	1.101	1.122		1.91	20
Benzene	1.417	1.472		3.88	20
Toluene	0.869	0.907		4.37	20
Tetrachloroethene	0.354	0.354		0	20
Ethyl Benzene	1.929	2.001		3.73	20
m/p-Xylenes	0.706	0.735		4.11	20
o-Xylene	0.688	0.718		4.36	20
1,2-Dichloroethane-d4	0.932	0.914		-1.93	20
Dibromofluoromethane	0.360	0.380		5.56	20
Toluene-d8	1.246	1.293		3.77	20
4-Bromofluorobenzene	0.478	0.507		6.07	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.