

Surrogate Summary

SDG No.: Q2033

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2033-01	TW-WTS-08	1,2-Dichloroethane-d4	50	52.8	106	70 (74)	130 (125)
		Dibromofluoromethane	50	51.8	104	70 (75)	130 (124)
		Toluene-d8	50	50.3	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.9	100	70 (77)	130 (121)
VX0514WBL01	VX0514WBL01	1,2-Dichloroethane-d4	50	53.7	107	70 (74)	130 (125)
		Dibromofluoromethane	50	51.2	102	70 (75)	130 (124)
		Toluene-d8	50	50.1	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.1	102	70 (77)	130 (121)
VX0514WBS01	VX0514WBS01	1,2-Dichloroethane-d4	50	52.1	104	70 (74)	130 (125)
		Dibromofluoromethane	50	51.2	102	70 (75)	130 (124)
		Toluene-d8	50	51.3	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.2	100	70 (77)	130 (121)
VX0514WBSD0	VX0514WBSD01	1,2-Dichloroethane-d4	50	51.3	103	70 (74)	130 (125)
		Dibromofluoromethane	50	51.6	103	70 (75)	130 (124)
		Toluene-d8	50	51.2	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.8	101	70 (77)	130 (121)

