

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033
 Lab File ID: VX046181.D Date Analyzed: 05/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:59
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	92522	5.54	155707	6.76	140519	10.05
UPPER LIMIT	185044	6.043	311414	7.257	281038	10.549
LOWER LIMIT	46261	5.043	77853.5	6.257	70259.5	9.549
EPA SAMPLE NO.						
TW-WTS-08	64370	5.55	127208	6.76	118657	10.05
VX0514WBL01	66143	5.55	133003	6.76	124085	10.05
VX0514WBS01	87880	5.55	155746	6.76	138245	10.05
VX0514WBSD01	86737	5.54	152178	6.76	134818	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.