

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033
 Lab File ID: VX046181.D Date Analyzed: 05/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:59
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	65751	12.018				
UPPER LIMIT	131502	12.518				
LOWER LIMIT	32875.5	11.518				
EPA SAMPLE NO.						
TW-WTS-08	50566	12.02				
VX0514WBL01	54666	12.02				
VX0514WBS01	63286	12.02				
VX0514WBSD01	63322	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.