

SOP ID : M1312-SPLP-10
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-7
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pipette ID : WC
Tumbler ID : T-2
TCLP Filter ID : 115525

Start Prep Date : 05/15/2025 Time : 14:00
End Prep Date : 05/16/2025 Time : 07:20
Combination Ratio : 20
ZHE Cleaning Batch : N/A JP
Initial Room Temperature: 23 °C
Final Room Temperature: 22 °C
TCLP Technician Signature : JP
Supervisor By : 12

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
SPLP FLUID	WP112802	N/A
N/A	N/A	N/A
HNO3-TCLP,1N	N/A	WP112799
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1940,W1941,W1942	W3166,W1938,W1939,
N/A	N/A	N/A
120ml Plastic bottle	N/A	2738
1:1 HNO3	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-2 checked,30 rpm.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
05/16/25 09:30	JP / TCLP Room	SKY. / HVE Lab
	Preparation Group	Analysis Group

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB168024TB	LEB024	17	N/A	2000	N/A	N/A	N/A	4.23	1.0	T-2
Q2035-01	SOIL-2	16	100.03	2000	N/A	N/A	N/A	4.5	1.5	T-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB168024TB	LEB024	N/A	N/A	N/A	N/A	N/A	N/A
Q2035-01	SOIL-2	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : WC S-1 / WC S-2

Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB168024TB	LEB024	N/A	N/A	N/A	N/A	#1	4.23
Q2035-01	SOIL-2	N/A	N/A	N/A	N/A	#1	4.23