

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen **SDG No.:** Q2035
Contract: BAPS03 **Lab Code:** CHEM **Case No.:** Q2035 **SAS No.:** Q2035
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	508	510	100	90 - 110	P	05/19/2025	12:28	LB135823
	Lead	998	1000	100	90 - 110	P	05/19/2025	12:28	LB135823

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen **SDG No.:** Q2035
Contract: BAPS03 **Lab Code:** CHEM **Case No.:** Q2035 **SAS No.:** Q2035
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	6.31	6.0	105	80 - 120	P	05/19/2025	12:38	LB135823
	Lead	11.2	12.0	93	80 - 120	P	05/19/2025	12:38	LB135823

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen **SDG No.:** Q2035
Contract: BAPS03 **Lab Code:** CHEM **Case No.:** Q2035 **SAS No.:** Q2035
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2540	2500	102	90 - 110	P	05/19/2025	13:15	LB135823
	Lead	5090	5000	102	90 - 110	P	05/19/2025	13:15	LB135823
CCV02	Cadmium	2540	2500	101	90 - 110	P	05/19/2025	14:13	LB135823
	Lead	5070	5000	101	90 - 110	P	05/19/2025	14:13	LB135823
CCV03	Cadmium	2430	2500	97	90 - 110	P	05/19/2025	15:14	LB135823
	Lead	4880	5000	98	90 - 110	P	05/19/2025	15:14	LB135823
CCV04	Cadmium	2450	2500	98	90 - 110	P	05/19/2025	16:03	LB135823
	Lead	4880	5000	98	90 - 110	P	05/19/2025	16:03	LB135823
CCV05	Cadmium	2440	2500	98	90 - 110	P	05/19/2025	16:47	LB135823
	Lead	4870	5000	97	90 - 110	P	05/19/2025	16:47	LB135823
CCV06	Cadmium	2480	2500	99	90 - 110	P	05/19/2025	17:35	LB135823
	Lead	4940	5000	99	90 - 110	P	05/19/2025	17:35	LB135823
CCV07	Cadmium	2460	2500	99	90 - 110	P	05/19/2025	18:21	LB135823
	Lead	4910	5000	98	90 - 110	P	05/19/2025	18:21	LB135823
CCV08	Cadmium	2470	2500	99	90 - 110	P	05/19/2025	19:09	LB135823
	Lead	4910	5000	98	90 - 110	P	05/19/2025	19:09	LB135823
CCV09	Cadmium	2430	2500	97	90 - 110	P	05/19/2025	19:55	LB135823
	Lead	4870	5000	98	90 - 110	P	05/19/2025	19:55	LB135823
CCV10	Cadmium	2440	2500	98	90 - 110	P	05/19/2025	20:11	LB135823
	Lead	4880	5000	98	90 - 110	P	05/19/2025	20:11	LB135823