

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: BAPS North Bergen **SDG No.:** Q2035
Contract: BAPS03 **Lab Code:** CHEM **Case No.:** Q2035 **SAS No.:** Q2035
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-4.22	1.0	422	-5	7	05/19/2025	12:52	LB135823
	Lead	-0.59			-12	12	05/19/2025	12:52	LB135823
ICSAB01	Cadmium	1030	972	106	826	1120	05/19/2025	12:57	LB135823
	Lead	49.5	49.0	101	37	61	05/19/2025	12:57	LB135823