

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: BAPS North Bergen **SDG No.:** Q2035
Contract: BAPS03 **Lab Code:** CHEM **Case No.:** Q2035 **SAS No.:** Q2035

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168044BS							
Cadmium	ug/L	200	187		94	80 - 120	P
Lead	ug/L	1000	921		92	80 - 120	P