

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2036

Project: Monthly 2025

Run No.: LB135774

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135774BS							
BOD5	mg/L	198	201		102	1	84.6-115.4	05/14/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2036

Project: Monthly 2025

Run No.: LB135782

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135782BS							
TPH	mg/L	20.0	16.9		84	1	78-114	05/15/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2036

Project: Monthly 2025

Run No.: LB135782

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135782BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	05/15/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2036

Project: Monthly 2025

Run No.: LB135797

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135797BS							
TSS	mg/L	550	532		97	1	90-110	05/16/2025