

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2037

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042538.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0514ABS01	Vinyl Chloride	10	8.60	ppbv	86			70	130	
	1,1-Dichloroethene	10	10.0	ppbv	100			70	130	
	cis-1,2-Dichloroethene	10	9.60	ppbv	96			70	130	
	1,1,1-Trichloroethane	10	8.90	ppbv	89			70	130	
	Trichloroethene	10	9.30	ppbv	93			70	130	
	Tetrachloroethene	10	9.00	ppbv	90			70	130	