

ANALYTICAL RESULTS SUMMARY

SEMI-VOLATILE ORGANICS
VOLATILE ORGANICS

PROJECT NAME : NWIRP BETHPAGE 112G08005-WE13

TETRA TECH NUS, INC.

661 Andersen Drive

Suite 200

Pittsburgh, PA - 15220-2745

Phone No: 412-921-7090

ORDER ID : Q2050

ATTENTION : Ernie Wu



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) VOCMS Group1- Case Narrative	4
2.2) SVOC-SIMGroup1- Case Narrative	6
3) Qualifier Page	8
4) QA Checklist	9
5) VOCMS Group1 Data	10
6) SVOC-SIMGroup1 Data	63
7) Shipping Document	93
7.1) CHAIN OF CUSTODY	94
7.2) Lab Certificate	95
7.3) Internal COC	96

1
2
3
4
5
6
7

Cover Page

Order ID : Q2050

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q2050-01
Q2050-02
Q2050-03
Q2050-04
Q2050-05
Q2050-06
Q2050-07
Q2050-08
Q2050-09

Client Sample Number

BP-VPB-182-TB-20250512
BP-VPB-182-GW-60-62
BP-VPB-182-GW-100-102
BP-VPB-182-GW-150-152
BP-VPB-182-GW-205-207
BP-VPB-182-GW-220-222
BP-VPB-182-DUP-20250514
BP-VPB-182-EB-20250514
BP-VPB-182-GW-240-242

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/26/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager # Ernie Wu

Order ID # Q2050

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

9 Water samples were received on 05/14/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOCMS Group1 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data Summary forms (i.e. Form Is)."

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager# Ernie Wu

Order ID # Q2050

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

9 Water samples were received on 05/14/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for BP-VPB-182-GW-150-152 [Terphenyl-d14 - 145%], BP-VPB-182-EB-20250514 [Terphenyl-d14 - 134%], The failure surrogates not associated with the client parameters list, therefore no corrective action was taken.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2050

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 05/26/2025

LAB CHRONICLE

OrderID:	Q2050	OrderDate:	5/15/2025 9:46:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	L31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2050-01	BP-VPB-182-TB-2025 0512	Water			05/12/25			05/14/25
			VOCMS Group1	8260-Low			05/15/25	
Q2050-02	BP-VPB-182-GW-60-6 2	Water			05/12/25			05/14/25
			VOCMS Group1	8260-Low			05/15/25	
Q2050-03	BP-VPB-182-GW-100- 102	Water			05/12/25			05/14/25
			VOCMS Group1	8260-Low			05/15/25	
Q2050-04	BP-VPB-182-GW-150- 152	Water			05/13/25			05/14/25
			VOCMS Group1	8260-Low			05/15/25	
Q2050-05	BP-VPB-182-GW-205- 207	Water			05/14/25			05/14/25
			VOCMS Group1	8260-Low			05/15/25	
Q2050-06	BP-VPB-182-GW-220- 222	Water			05/14/25			05/14/25
			VOCMS Group1	8260-Low			05/15/25	
Q2050-07	BP-VPB-182-DUP-202 50514	Water			05/14/25			05/14/25
			VOCMS Group1	8260-Low			05/15/25	
Q2050-08	BP-VPB-182-EB-2025 0514	Water			05/14/25			05/14/25
			VOCMS Group1	8260-Low			05/15/25	
Q2050-09	BP-VPB-182-GW-240- 242	Water			05/14/25			05/14/25
			VOCMS Group1	8260-Low			05/15/25	

Hit Summary Sheet
SW-846

SDG No.: Q2050
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID: Q2050-01	BP-VPB-182-TB-20250512 BP-VPB-182-TB-20 Water	Acetone		2.50	J	1.50	3.80	5.00	ug/L
		Total Voc :		2.50					
		Total Concentration:		2.50					
Client ID: Q2050-02	BP-VPB-182-GW-60-62 BP-VPB-182-GW-6 Water	Acetone		7.80		1.50	3.80	5.00	ug/L
Q2050-02	BP-VPB-182-GW-6 Water	Chloroform		1.10		0.25	0.50	1.00	ug/L
Q2050-02	BP-VPB-182-GW-6 Water	Tetrachloroethene		1.30		0.23	0.50	1.00	ug/L
		Total Voc :		10.2					
		Total Concentration:		10.2					
Client ID: Q2050-03	BP-VPB-182-GW-100-102 BP-VPB-182-GW-1 Water	Acetone		3.80	J	1.50	3.80	5.00	ug/L
Q2050-03	BP-VPB-182-GW-1 Water	Chloroform		0.93	J	0.25	0.50	1.00	ug/L
Q2050-03	BP-VPB-182-GW-1 Water	Tetrachloroethene		1.10		0.23	0.50	1.00	ug/L
		Total Voc :		5.83					
		Total Concentration:		5.83					
Client ID: Q2050-04	BP-VPB-182-GW-150-152 BP-VPB-182-GW-1 Water	Acetone		4.00	J	1.50	3.80	5.00	ug/L
		Total Voc :		4.00					
		Total Concentration:		4.00					
Client ID: Q2050-05	BP-VPB-182-GW-205-207 BP-VPB-182-GW-2 Water	Acetone		5.20		1.50	3.80	5.00	ug/L
		Total Voc :		5.20					
		Total Concentration:		5.20					
Client ID: Q2050-06	BP-VPB-182-GW-220-222 BP-VPB-182-GW-2 Water	Acetone		3.20	J	1.50	3.80	5.00	ug/L
		Total Voc :		3.20					
		Total Concentration:		3.20					
Client ID: Q2050-07	BP-VPB-182-DUP-20250514 BP-VPB-182-DUP- Water	Acetone		2.80	J	1.50	3.80	5.00	ug/L
		Total Voc :		2.80					
		Total Concentration:		2.80					
Client ID: Q2050-08	BP-VPB-182-EB-20250514 BP-VPB-182-EB-20 Water	Acetone		4.30	J	1.50	3.80	5.00	ug/L
		Total Voc :		4.30					
		Total Concentration:		4.30					
Client ID: Q2050-09	BP-VPB-182-GW-240-242 BP-VPB-182-GW-2 Water	Acetone		2.90	J	1.50	3.80	5.00	ug/L
		Total Voc :		2.90					
		Total Concentration:		2.90					



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/12/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-TB-20250512		SDG No.:	Q2050	
Lab Sample ID:	Q2050-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046209.D	1		05/15/25 13:24	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.50	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/12/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-TB-20250512		SDG No.:	Q2050	
Lab Sample ID:	Q2050-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046209.D	1		05/15/25 13:24	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.1		81 - 118		110%	SPK: 50
1868-53-7	Dibromofluoromethane	52.2		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	51.0		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.7		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	62700	5.544				
540-36-3	1,4-Difluorobenzene	125000	6.757				
3114-55-4	Chlorobenzene-d5	120000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	51100	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/12/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-TB-20250512		SDG No.:	Q2050	
Lab Sample ID:	Q2050-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046209.D	1		05/15/25 13:24	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/12/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-60-62		SDG No.:	Q2050	
Lab Sample ID:	Q2050-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046208.D	1		05/15/25 13:01	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	7.80		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	1.10		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/12/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-60-62		SDG No.:	Q2050	
Lab Sample ID:	Q2050-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046208.D	1		05/15/25 13:01	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	1.30		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.7		81 - 118		109%	SPK: 50
1868-53-7	Dibromofluoromethane	51.3		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	49.8		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.0		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	61900	5.55				
540-36-3	1,4-Difluorobenzene	123000	6.757				
3114-55-4	Chlorobenzene-d5	115000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	50500	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/12/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-60-62		SDG No.:	Q2050	
Lab Sample ID:	Q2050-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046208.D	1		05/15/25 13:01	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/12/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-100-102		SDG No.:	Q2050	
Lab Sample ID:	Q2050-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046211.D	1		05/15/25 14:11	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.93	J	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/12/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-100-102		SDG No.:	Q2050	
Lab Sample ID:	Q2050-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046211.D	1		05/15/25 14:11	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	1.10		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.9		81 - 118		110%	SPK: 50
1868-53-7	Dibromofluoromethane	53.0		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	51.3		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.6		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	61200	5.55				
540-36-3	1,4-Difluorobenzene	121000	6.757				
3114-55-4	Chlorobenzene-d5	115000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	47600	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/12/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-100-102		SDG No.:	Q2050	
Lab Sample ID:	Q2050-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046211.D	1		05/15/25 14:11	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/13/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-150-152		SDG No.:	Q2050	
Lab Sample ID:	Q2050-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046212.D	1		05/15/25 14:34	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	4.00	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/13/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-150-152		SDG No.:	Q2050	
Lab Sample ID:	Q2050-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046212.D	1		05/15/25 14:34	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.7		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.4		80 - 119		105%	SPK: 50
2037-26-5	Toluene-d8	50.5		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.1		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	63700	5.55				
540-36-3	1,4-Difluorobenzene	128000	6.757				
3114-55-4	Chlorobenzene-d5	120000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	51500	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/13/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-150-152		SDG No.:	Q2050	
Lab Sample ID:	Q2050-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046212.D	1		05/15/25 14:34	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-205-207			SDG No.:	Q2050	
Lab Sample ID:	Q2050-05			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046213.D	1		05/15/25 14:58	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	5.20		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/14/25		
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	05/14/25		
Client Sample ID:	BP-VPB-182-GW-205-207			SDG No.:	Q2050		
Lab Sample ID:	Q2050-05			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046213.D	1		05/15/25 14:58	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.4		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	51.5		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.5		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	60500	5.55				
540-36-3	1,4-Difluorobenzene	121000	6.757				
3114-55-4	Chlorobenzene-d5	113000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	47400	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-205-207		SDG No.:	Q2050	
Lab Sample ID:	Q2050-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046213.D	1		05/15/25 14:58	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-220-222		SDG No.:	Q2050	
Lab Sample ID:	Q2050-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046214.D	1		05/15/25 15:21	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.20	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-220-222		SDG No.:	Q2050	
Lab Sample ID:	Q2050-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046214.D	1		05/15/25 15:21	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.6		81 - 118		109%	SPK: 50
1868-53-7	Dibromofluoromethane	51.6		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	50.8		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.4		85 - 114		97%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	60700	5.55				
540-36-3	1,4-Difluorobenzene	121000	6.763				
3114-55-4	Chlorobenzene-d5	112000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	45000	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-220-222		SDG No.:	Q2050	
Lab Sample ID:	Q2050-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046214.D	1		05/15/25 15:21	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-DUP-20250514			SDG No.:	Q2050	
Lab Sample ID:	Q2050-07			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046215.D	1		05/15/25 15:44	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.80	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-DUP-20250514		SDG No.:	Q2050	
Lab Sample ID:	Q2050-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046215.D	1		05/15/25 15:44	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.6		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	50.6		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.4		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	61100	5.544				
540-36-3	1,4-Difluorobenzene	122000	6.757				
3114-55-4	Chlorobenzene-d5	117000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	50000	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-DUP-20250514		SDG No.:	Q2050	
Lab Sample ID:	Q2050-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046215.D	1		05/15/25 15:44	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-EB-20250514		SDG No.:	Q2050	
Lab Sample ID:	Q2050-08		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046210.D	1		05/15/25 13:47	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	4.30	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-EB-20250514		SDG No.:	Q2050	
Lab Sample ID:	Q2050-08		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046210.D	1		05/15/25 13:47	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.1		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.8		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	50.5		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.9		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	61600	5.544				
540-36-3	1,4-Difluorobenzene	122000	6.757				
3114-55-4	Chlorobenzene-d5	116000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	50300	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-EB-20250514		SDG No.:	Q2050	
Lab Sample ID:	Q2050-08		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046210.D	1		05/15/25 13:47	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-240-242			SDG No.:	Q2050	
Lab Sample ID:	Q2050-09			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046216.D	1		05/15/25 16:08	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.90	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/14/25		
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	05/14/25		
Client Sample ID:	BP-VPB-182-GW-240-242			SDG No.:	Q2050		
Lab Sample ID:	Q2050-09			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046216.D	1		05/15/25 16:08	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.0		81 - 118		110%	SPK: 50
1868-53-7	Dibromofluoromethane	52.2		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	50.6		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	62400	5.55				
540-36-3	1,4-Difluorobenzene	125000	6.757				
3114-55-4	Chlorobenzene-d5	120000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	50200	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/14/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/14/25	
Client Sample ID:	BP-VPB-182-GW-240-242		SDG No.:	Q2050	
Lab Sample ID:	Q2050-09		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046216.D	1		05/15/25 16:08	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q2050**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2050-01	BP-VPB-182-TB-20250512	1,2-Dichloroethane-d4	50	55.1	110	81	118
		Dibromofluoromethane	50	52.3	104	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	50.7	101	85	114
Q2050-02	BP-VPB-182-GW-60-62	1,2-Dichloroethane-d4	50	54.7	109	81	118
		Dibromofluoromethane	50	51.3	103	80	119
		Toluene-d8	50	49.8	100	89	112
		4-Bromofluorobenzene	50	50.0	100	85	114
Q2050-03	BP-VPB-182-GW-100-102	1,2-Dichloroethane-d4	50	54.9	110	81	118
		Dibromofluoromethane	50	53.0	106	80	119
		Toluene-d8	50	51.3	103	89	112
		4-Bromofluorobenzene	50	49.6	99	85	114
Q2050-04	BP-VPB-182-GW-150-152	1,2-Dichloroethane-d4	50	55.6	111	81	118
		Dibromofluoromethane	50	52.4	105	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	50.0	100	85	114
Q2050-05	BP-VPB-182-GW-205-207	1,2-Dichloroethane-d4	50	55.5	111	81	118
		Dibromofluoromethane	50	52.0	104	80	119
		Toluene-d8	50	51.5	103	89	112
		4-Bromofluorobenzene	50	49.5	99	85	114
Q2050-06	BP-VPB-182-GW-220-222	1,2-Dichloroethane-d4	50	54.6	109	81	118
		Dibromofluoromethane	50	51.6	103	80	119
		Toluene-d8	50	50.8	102	89	112
		4-Bromofluorobenzene	50	48.4	97	85	114
Q2050-07	BP-VPB-182-DUP-20250514	1,2-Dichloroethane-d4	50	55.6	111	81	118
		Dibromofluoromethane	50	52.0	104	80	119
		Toluene-d8	50	50.6	101	89	112
		4-Bromofluorobenzene	50	51.4	103	85	114
Q2050-08	BP-VPB-182-EB-20250514	1,2-Dichloroethane-d4	50	54.1	108	81	118
		Dibromofluoromethane	50	51.8	104	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	51.9	104	85	114
Q2050-09	BP-VPB-182-GW-240-242	1,2-Dichloroethane-d4	50	55.0	110	81	118
		Dibromofluoromethane	50	52.1	104	80	119
		Toluene-d8	50	50.6	101	89	112
		4-Bromofluorobenzene	50	50.4	101	85	114
VX0515WBL01	VX0515WBL01	1,2-Dichloroethane-d4	50	54.2	108	81	118
		Dibromofluoromethane	50	51.5	103	80	119
		Toluene-d8	50	50.2	100	89	112
		4-Bromofluorobenzene	50	49.0	98	85	114
VX0515WBS01	VX0515WBS01	1,2-Dichloroethane-d4	50	50.7	101	81	118
		Dibromofluoromethane	50	50.0	100	80	119
		Toluene-d8	50	49.7	99	89	112
		4-Bromofluorobenzene	50	48.9	98	85	114
VX0515WBSD0	VX0515WBSD01	1,2-Dichloroethane-d4	50	49.4	99	81	118
		Dibromofluoromethane	50	49.3	99	80	119
		Toluene-d8	50	48.5	97	89	112
		4-Bromofluorobenzene	50	47.9	96	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2050

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX046206.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0515WBS01	Chloromethane	20	19.5	ug/L	98			50	139	
	Vinyl chloride	20	19.1	ug/L	96			58	137	
	Bromomethane	20	18.9	ug/L	95			53	141	
	Chloroethane	20	22.5	ug/L	113			60	138	
	Trichlorofluoromethane	20	20.4	ug/L	102			65	141	
	1,1,2-Trichlorotrifluoroethane	20	19.9	ug/L	100			70	136	
	1,1-Dichloroethene	20	19.0	ug/L	95			71	131	
	Acetone	100	100	ug/L	100			39	160	
	Carbon disulfide	20	16.7	ug/L	84			64	133	
	Methyl tert-butyl Ether	20	19.9	ug/L	100			71	124	
	Methylene Chloride	20	19.0	ug/L	95			74	124	
	trans-1,2-Dichloroethene	20	19.5	ug/L	98			75	124	
	1,1-Dichloroethane	20	20.2	ug/L	101			77	125	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	19.5	ug/L	98			72	136	
	cis-1,2-Dichloroethene	20	20.2	ug/L	101			78	123	
	Chloroform	20	21.0	ug/L	105			79	124	
	1,1,1-Trichloroethane	20	20.2	ug/L	101			74	131	
	Methylcyclohexane	20	18.3	ug/L	92			72	132	
	Benzene	20	20.1	ug/L	101			79	120	
	1,2-Dichloroethane	20	20.6	ug/L	103			73	128	
	Trichloroethene	20	19.3	ug/L	97			79	123	
	1,2-Dichloropropane	20	20.9	ug/L	104			78	122	
	Bromodichloromethane	20	20.6	ug/L	103			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	20.0	ug/L	100			80	121	
	t-1,3-Dichloropropene	20	18.5	ug/L	93			73	127	
	cis-1,3-Dichloropropene	20	19.5	ug/L	98			75	124	
	1,1,2-Trichloroethane	20	20.7	ug/L	104			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	20.7	ug/L	104			74	126	
	Tetrachloroethene	20	19.5	ug/L	98			74	129	
	Chlorobenzene	20	19.5	ug/L	98			82	118	
	Ethyl Benzene	20	19.8	ug/L	99			79	121	
	m/p-Xylenes	40	39.4	ug/L	99			80	121	
	o-Xylene	20	20.3	ug/L	102			78	122	
	Styrene	20	20.5	ug/L	103			78	123	
	Bromoform	20	19.1	ug/L	96			66	130	
	Isopropylbenzene	20	20.1	ug/L	101			72	131	
	1,1,2,2-Tetrachloroethane	20	19.8	ug/L	99			71	121	
	1,3-Dichlorobenzene	20	19.7	ug/L	99			80	119	
	1,4-Dichlorobenzene	20	19.3	ug/L	97			79	118	
	1,2-Dichlorobenzene	20	20.3	ug/L	102			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2050

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX046207.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0515WBSD01	Chloromethane	20	19.3	ug/L	97	1		50	139	20
	Vinyl chloride	20	18.2	ug/L	91	5		58	137	20
	Bromomethane	20	17.8	ug/L	89	7		53	141	20
	Chloroethane	20	19.8	ug/L	99	13		60	138	20
	Trichlorofluoromethane	20	19.8	ug/L	99	3		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	19.3	ug/L	97	3		70	136	20
	1,1-Dichloroethene	20	17.9	ug/L	90	5		71	131	20
	Acetone	100	100	ug/L	100	0		39	160	20
	Carbon disulfide	20	16.1	ug/L	81	4		64	133	20
	Methyl tert-butyl Ether	20	20.1	ug/L	101	1		71	124	20
	Methylene Chloride	20	18.7	ug/L	94	1		74	124	20
	trans-1,2-Dichloroethene	20	19.0	ug/L	95	3		75	124	20
	1,1-Dichloroethane	20	20.0	ug/L	100	1		77	125	20
	2-Butanone	100	110	ug/L	110	0		56	143	20
	Carbon Tetrachloride	20	19.2	ug/L	96	2		72	136	20
	cis-1,2-Dichloroethene	20	19.3	ug/L	97	4		78	123	20
	Chloroform	20	20.6	ug/L	103	2		79	124	20
	1,1,1-Trichloroethane	20	19.8	ug/L	99	2		74	131	20
	Methylcyclohexane	20	17.9	ug/L	90	2		72	132	20
	Benzene	20	19.7	ug/L	99	2		79	120	20
	1,2-Dichloroethane	20	20.7	ug/L	104	1		73	128	20
	Trichloroethene	20	18.8	ug/L	94	3		79	123	20
	1,2-Dichloropropane	20	21.3	ug/L	106	2		78	122	20
	Bromodichloromethane	20	20.3	ug/L	102	1		79	125	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	0		67	130	20
	Toluene	20	20.1	ug/L	101	1		80	121	20
	t-1,3-Dichloropropene	20	18.8	ug/L	94	1		73	127	20
	cis-1,3-Dichloropropene	20	19.6	ug/L	98	0		75	124	20
	1,1,2-Trichloroethane	20	21.5	ug/L	108	4		80	119	20
	2-Hexanone	100	110	ug/L	110	0		57	139	20
	Dibromochloromethane	20	20.7	ug/L	104	0		74	126	20
	Tetrachloroethene	20	19.3	ug/L	97	1		74	129	20
	Chlorobenzene	20	19.5	ug/L	98	0		82	118	20
	Ethyl Benzene	20	19.7	ug/L	99	0		79	121	20
	m/p-Xylenes	40	39.6	ug/L	99	0		80	121	20
	o-Xylene	20	20.0	ug/L	100	2		78	122	20
	Styrene	20	20.5	ug/L	103	0		78	123	20
	Bromoform	20	19.9	ug/L	100	4		66	130	20
	Isopropylbenzene	20	20.1	ug/L	101	0		72	131	20
	1,1,2,2-Tetrachloroethane	20	20.6	ug/L	103	4		71	121	20
	1,3-Dichlorobenzene	20	19.5	ug/L	98	1		80	119	20
	1,4-Dichlorobenzene	20	18.8	ug/L	94	3		79	118	20
	1,2-Dichlorobenzene	20	20.5	ug/L	103	1		80	119	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0515WBL01

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q2050

SAS No.: Q2050 SDG NO.: Q2050

Lab File ID: VX046202.D

Lab Sample ID: VX0515WBL01

Date Analyzed: 05/15/2025

Time Analyzed: 10:38

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0515WBS01	VX0515WBS01	VX046206.D	05/15/2025
VX0515WBSD01	VX0515WBSD01	VX046207.D	05/15/2025
BP-VPB-182-GW-60-62	Q2050-02	VX046208.D	05/15/2025
BP-VPB-182-TB-20250512	Q2050-01	VX046209.D	05/15/2025
BP-VPB-182-EB-20250514	Q2050-08	VX046210.D	05/15/2025
BP-VPB-182-GW-100-102	Q2050-03	VX046211.D	05/15/2025
BP-VPB-182-GW-150-152	Q2050-04	VX046212.D	05/15/2025
BP-VPB-182-GW-205-207	Q2050-05	VX046213.D	05/15/2025
BP-VPB-182-GW-220-222	Q2050-06	VX046214.D	05/15/2025
BP-VPB-182-DUP-20250514	Q2050-07	VX046215.D	05/15/2025
BP-VPB-182-GW-240-242	Q2050-09	VX046216.D	05/15/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050
Lab File ID: VX046038.D BFB Injection Date: 05/05/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:37
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	68.8
175	5.0 - 9.0% of mass 174	5 (7.3) 1
176	95.0 - 101.0% of mass 174	66.7 (97) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC020	VSTDIC020	VX046041.D	05/05/2025	11:35
VSTDIC050	VSTDIC050	VX046042.D	05/05/2025	11:58
VSTDIC100	VSTDIC100	VX046043.D	05/05/2025	12:21
VSTDIC150	VSTDIC150	VX046044.D	05/05/2025	12:45
VSTDIC005	VSTDIC005	VX046046.D	05/05/2025	16:04
VSTDIC001	VSTDIC001	VX046047.D	05/05/2025	16:27

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050
Lab File ID: VX046199.D BFB Injection Date: 05/15/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:47
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.9
75	30.0 - 60.0% of mass 95	58.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.7 (1.1) 1
174	50.0 - 100.0% of mass 95	67.5
175	5.0 - 9.0% of mass 174	5.3 (7.8) 1
176	95.0 - 101.0% of mass 174	65.6 (97.2) 1
177	5.0 - 9.0% of mass 176	4 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046200.D	05/15/2025	09:52
VX0515WBL01	VX0515WBL01	VX046202.D	05/15/2025	10:38
VX0515WBS01	VX0515WBS01	VX046206.D	05/15/2025	12:11
VX0515WBSD01	VX0515WBSD01	VX046207.D	05/15/2025	12:38
BP-VPB-182-GW-60-62	Q2050-02	VX046208.D	05/15/2025	13:01
BP-VPB-182-TB-20250512	Q2050-01	VX046209.D	05/15/2025	13:24
BP-VPB-182-EB-20250514	Q2050-08	VX046210.D	05/15/2025	13:47
BP-VPB-182-GW-100-102	Q2050-03	VX046211.D	05/15/2025	14:11
BP-VPB-182-GW-150-152	Q2050-04	VX046212.D	05/15/2025	14:34
BP-VPB-182-GW-205-207	Q2050-05	VX046213.D	05/15/2025	14:58
BP-VPB-182-GW-220-222	Q2050-06	VX046214.D	05/15/2025	15:21
BP-VPB-182-DUP-20250514	Q2050-07	VX046215.D	05/15/2025	15:44
BP-VPB-182-GW-240-242	Q2050-09	VX046216.D	05/15/2025	16:08
VSTDCCC050EC	VSTDCCC050	VX046226.D	05/15/2025	20:01

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050
Lab File ID: VX046200.D Date Analyzed: 05/15/2025
Instrument ID: MSVOA_X Time Analyzed: 09:52
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	79845	5.54	140483	6.75	127630	10.05
UPPER LIMIT	159690	6.044	280966	7.251	255260	10.549
LOWER LIMIT	39922.5	5.044	70241.5	6.251	63815	9.549
EPA SAMPLE NO.						
BP-VPB-182-TB-20250512	62722	5.54	124979	6.76	119897	10.05
BP-VPB-182-GW-60-62	61866	5.55	123143	6.76	115076	10.06
BP-VPB-182-GW-100-102	61221	5.55	120798	6.76	115485	10.06
BP-VPB-182-GW-150-152	63694	5.55	128038	6.76	119578	10.05
BP-VPB-182-GW-205-207	60533	5.55	121144	6.76	113426	10.06
BP-VPB-182-GW-220-222	60707	5.55	121001	6.76	112446	10.06
BP-VPB-182-DUP-20250514	61073	5.54	122275	6.76	116802	10.05
BP-VPB-182-EB-20250514	61610	5.54	121862	6.76	116490	10.05
BP-VPB-182-GW-240-242	62386	5.55	125278	6.76	120203	10.06
VX0515WBL01	65765	5.54	130899	6.76	122283	10.05
VX0515WBS01	89041	5.54	157824	6.76	139362	10.05
VX0515WBSD01	86433	5.54	151750	6.76	132818	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050
Lab File ID: VX046200.D Date Analyzed: 05/15/2025
Instrument ID: MSVOA_X Time Analyzed: 09:52
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	62876	12.018				
UPPER LIMIT	125752	12.518				
LOWER LIMIT	31438	11.518				
EPA SAMPLE NO.						
BP-VPB-182-TB-20250512	51079	12.02				
BP-VPB-182-GW-60-62	50521	12.02				
BP-VPB-182-GW-100-102	47647	12.02				
BP-VPB-182-GW-150-152	51539	12.02				
BP-VPB-182-GW-205-207	47407	12.02				
BP-VPB-182-GW-220-222	44995	12.02				
BP-VPB-182-DUP-20250514	50006	12.02				
BP-VPB-182-EB-20250514	50288	12.02				
BP-VPB-182-GW-240-242	50190	12.02				
VX0515WBL01	51360	12.02				
VX0515WBS01	65777	12.02				
VX0515WBSD01	63406	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0515WBL01		SDG No.:	Q2050
Lab Sample ID:	VX0515WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046202.D	1		05/15/25 10:38	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0515WBL01	SDG No.:	Q2050	
Lab Sample ID:	VX0515WBL01	Matrix:	Water	
Analytical Method:	8260D	% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	
Soil Aliquot Vol:			uL	
GC Column:	DB-624UI	ID :	0.18	
Prep Method :		Level :	LOW	
		Final Vol:	5000	uL
		Test:	VOCMS Group1	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046202.D	1		05/15/25 10:38	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.2		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.5		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.9		85 - 114		98%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	65800	5.544				
540-36-3	1,4-Difluorobenzene	131000	6.757				
3114-55-4	Chlorobenzene-d5	122000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	51400	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0515WBS01		SDG No.:	Q2050
Lab Sample ID:	VX0515WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046206.D	1		05/15/25 12:11	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	19.5		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	19.1		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	18.9		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	22.5		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	20.4		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.9		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.0		0.23	0.75	1.00	ug/L
67-64-1	Acetone	100		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	16.7		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.9		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	19.0		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.5		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.2		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.5		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.2		0.19	0.75	1.00	ug/L
67-66-3	Chloroform	21.0		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.2		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.3		0.16	0.50	1.00	ug/L
71-43-2	Benzene	20.1		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.6		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.3		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.9		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.6		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	20.0		0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	18.5		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.5		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.7		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0515WBS01		SDG No.:	Q2050
Lab Sample ID:	VX0515WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046206.D	1		05/15/25 12:11	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	20.7		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.5		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.5		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.8		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.4		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	20.3		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.5		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	19.1		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.1		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.8		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.7		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.3		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.3		0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.7		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	50.0		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.7		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.9		85 - 114		98%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	89000	5.544				
540-36-3	1,4-Difluorobenzene	158000	6.757				
3114-55-4	Chlorobenzene-d5	139000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	65800	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0515WBSD01		SDG No.:	Q2050
Lab Sample ID:	VX0515WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046207.D	1		05/15/25 12:38	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	19.3		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.2		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	17.8		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.8		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.8		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.9		0.23	0.75	1.00	ug/L
67-64-1	Acetone	100		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	16.1		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.1		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	18.7		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.0		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.0		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.2		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.3		0.19	0.75	1.00	ug/L
67-66-3	Chloroform	20.6		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.8		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	17.9		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.7		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.7		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	18.8		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	21.3		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.3		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	20.1		0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	18.8		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.6		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0515WBSD01		SDG No.:	Q2050
Lab Sample ID:	VX0515WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046207.D	1		05/15/25 12:38	VX051525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	20.7		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.3		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.5		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.7		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.6		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	20.0		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.5		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	19.9		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.1		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.6		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.5		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.8		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.5		0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	49.4		81 - 118		99%	SPK: 50
1868-53-7	Dibromofluoromethane	49.3		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	48.5		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.9		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	86400	5.544				
540-36-3	1,4-Difluorobenzene	152000	6.757				
3114-55-4	Chlorobenzene-d5	133000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	63400	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG No.: Q2050

Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF020 = VX046041.D	RRF050 = VX046042.D	RRF100 = VX046043.D	RRF150 = VX046044.D	RRF005 = VX046046.D	RRF001 = VX046047.D		
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD
Chloromethane	0.727	0.775	0.787	0.791	0.679	0.694	0.742	6.6
Vinyl Chloride	0.660	0.710	0.727	0.755	0.619	0.673	0.691	7.2
Bromomethane	0.296	0.326	0.340	0.334	0.305		0.320	5.8
Chloroethane	0.354	0.378	0.329	0.317	0.368	0.467	0.369	14.4
Trichlorofluoromethane	1.035	1.068	0.983	0.985	0.990	1.064	1.021	3.9
1,1,2-Trichlorotrifluoroethane	0.628	0.641	0.629	0.648	0.610	0.633	0.632	2.1
1,1-Dichloroethene	0.565	0.601	0.607	0.625	0.567	0.594	0.593	3.9
Acetone	0.361	0.362	0.361	0.370	0.408	0.380	0.374	4.9
Carbon Disulfide	1.295	1.455	1.522	1.597	1.141	1.423	1.406	11.7
Methyl tert-butyl Ether	2.044	2.160	2.172	2.239	1.908	1.949	2.079	6.4
Methylene Chloride	0.689	0.684	0.691	0.691	0.689	0.853	0.716	9.4
trans-1,2-Dichloroethene	0.573	0.610	0.612	0.622	0.557	0.604	0.596	4.3
1,1-Dichloroethane	1.233	1.263	1.263	1.286	1.154	1.116	1.219	5.6
2-Butanone	0.540	0.555	0.558	0.569	0.539	0.495	0.543	4.8
Carbon Tetrachloride	0.528	0.558	0.552	0.577	0.505	0.541	0.544	4.6
cis-1,2-Dichloroethene	0.716	0.737	0.738	0.755	0.642	0.719	0.718	5.5
Chloroform	1.287	1.296	1.277	1.300	1.199	1.265	1.271	3
1,1,1-Trichloroethane	1.106	1.131	1.155	1.188	1.013	1.015	1.101	6.6
Methylcyclohexane	0.596	0.641	0.627	0.658	0.587	0.627	0.623	4.3
Benzene	1.426	1.474	1.441	1.477	1.337	1.348	1.417	4.3
1,2-Dichloroethane	0.632	0.627	0.611	0.625	0.594	0.579	0.612	3.5
Trichloroethene	0.344	0.355	0.345	0.362	0.315	0.324	0.341	5.3
1,2-Dichloropropane	0.356	0.371	0.368	0.378	0.324	0.317	0.352	7.4
Bromodichloromethane	0.557	0.577	0.573	0.594	0.498	0.485	0.547	8.2
4-Methyl-2-Pentanone	0.620	0.634	0.630	0.631	0.555	0.561	0.605	6
Toluene	0.884	0.898	0.885	0.904	0.838	0.803	0.869	4.5
t-1,3-Dichloropropene	0.468	0.528	0.555	0.591	0.406	0.371	0.487	17.9
cis-1,3-Dichloropropene	0.531	0.578	0.602	0.623	0.469	0.423	0.538	14.6
1,1,2-Trichloroethane	0.349	0.354	0.351	0.356	0.337	0.308	0.343	5.3
2-Hexanone	0.466	0.473	0.477	0.473	0.414	0.385	0.448	8.7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG No.: Q2050

Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D			
		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D			
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD	
Dibromochloromethane	0.378	0.400	0.415	0.431	0.326	0.306	0.376	13.3	
Tetrachloroethene	0.390	0.375	0.345	0.344	0.323	0.347	0.354	6.8	
Chlorobenzene	1.093	1.098	1.085	1.114	1.046	1.131	1.094	2.7	
Ethyl Benzene	1.919	2.022	1.979	2.036	1.816	1.803	1.929	5.2	
m/p-Xylenes	0.706	0.740	0.721	0.740	0.678	0.648	0.706	5.2	
o-Xylene	0.688	0.727	0.706	0.726	0.639	0.642	0.688	5.7	
Styrene	1.135	1.219	1.214	1.230	1.012	0.951	1.127	10.6	
Bromoform	0.270	0.304	0.312	0.327	0.236	0.234	0.281	14.2	
Isopropylbenzene	3.843	4.130	3.876	4.156	3.562	3.789	3.893	5.7	
1,1,2,2-Tetrachloroethane	1.315	1.338	1.284	1.345	1.350	1.552	1.364	7	
1,3-Dichlorobenzene	1.633	1.701	1.656	1.730	1.558	1.619	1.649	3.7	
1,4-Dichlorobenzene	1.629	1.693	1.639	1.722	1.606	1.817	1.684	4.6	
1,2-Dichlorobenzene	1.613	1.696	1.634	1.702	1.577	1.710	1.655	3.3	
1,2-Dichloroethane-d4	0.953	0.910	0.930	0.932	0.935		0.932	1.6	
Dibromofluoromethane	0.359	0.355	0.364	0.368	0.354		0.360	1.7	
Toluene-d8	1.246	1.223	1.266	1.275	1.221		1.246	2	
4-Bromofluorobenzene	0.455	0.470	0.500	0.500	0.464		0.478	4.4	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG No.: Q2050

Instrument ID: MSVOA_X Calibration Date/Time: 05/15/2025 09:52

Lab File ID: VX046200.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.742	0.714	0.1	-3.77	20
Vinyl Chloride	0.691	0.611		-11.58	20
Bromomethane	0.320	0.287		-10.31	20
Chloroethane	0.369	0.363		-1.63	20
Trichlorofluoromethane	1.021	0.958		-6.17	20
1,1,2-Trichlorotrifluoroethane	0.632	0.579		-8.39	20
1,1-Dichloroethene	0.593	0.532		-10.29	20
Acetone	0.374	0.380		1.6	20
Carbon Disulfide	1.406	1.195		-15.01	20
Methyl tert-butyl Ether	2.079	2.213		6.45	20
Methylene Chloride	0.716	0.687		-4.05	20
trans-1,2-Dichloroethene	0.596	0.572		-4.03	20
1,1-Dichloroethane	1.219	1.221	0.1	0.16	20
2-Butanone	0.543	0.583		7.37	20
Carbon Tetrachloride	0.544	0.510		-6.25	20
cis-1,2-Dichloroethene	0.718	0.727		1.25	20
Chloroform	1.271	1.299		2.2	20
1,1,1-Trichloroethane	1.101	1.079		-2	20
Methylcyclohexane	0.623	0.559		-10.27	20
Benzene	1.417	1.399		-1.27	20
1,2-Dichloroethane	0.612	0.634		3.6	20
Trichloroethene	0.341	0.328		-3.81	20
1,2-Dichloropropane	0.352	0.371		5.4	20
Bromodichloromethane	0.547	0.585		6.95	20
4-Methyl-2-Pentanone	0.605	0.671		10.91	20
Toluene	0.869	0.871		0.23	20
t-1,3-Dichloropropene	0.487	0.535		9.86	20
cis-1,3-Dichloropropene	0.538	0.583		8.36	20
1,1,2-Trichloroethane	0.343	0.367		7	20
2-Hexanone	0.448	0.501		11.83	20
Dibromochloromethane	0.376	0.418		11.17	20
Tetrachloroethene	0.354	0.326		-7.91	20
Chlorobenzene	1.094	1.065	0.3	-2.65	20
Ethyl Benzene	1.929	1.871		-3.01	20
m/p-Xylenes	0.706	0.693		-1.84	20
o-Xylene	0.688	0.686		-0.29	20
Styrene	1.127	1.188		5.41	20
Bromoform	0.281	0.301	0.1	7.12	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG No.: Q2050

Instrument ID: MSVOA_X Calibration Date/Time: 05/15/2025 09:52

Lab File ID: VX046200.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.893	3.674		-5.63	20
1,1,2,2-Tetrachloroethane	1.364	1.329	0.3	-2.57	20
1,3-Dichlorobenzene	1.649	1.630		-1.15	20
1,4-Dichlorobenzene	1.684	1.628		-3.33	20
1,2-Dichlorobenzene	1.655	1.656		0.06	20
1,2-Dichloroethane-d4	0.932	0.941		0.97	20
Dibromofluoromethane	0.360	0.366		1.67	20
Toluene-d8	1.246	1.215		-2.49	20
4-Bromofluorobenzene	0.478	0.492		2.93	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG No.: Q2050

Instrument ID: MSVOA_X Calibration Date/Time: 05/15/2025 20:01

Lab File ID: VX046226.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.742	0.751	0.1	1.21	50
Vinyl Chloride	0.691	0.688		-0.43	50
Bromomethane	0.320	0.304		-5	50
Chloroethane	0.369	0.380		2.98	50
Trichlorofluoromethane	1.021	1.042		2.06	50
1,1,2-Trichlorotrifluoroethane	0.632	0.632		0	50
1,1-Dichloroethene	0.593	0.583		-1.69	50
Acetone	0.374	0.390		4.28	50
Carbon Disulfide	1.406	1.298		-7.68	50
Methyl tert-butyl Ether	2.079	2.200		5.82	50
Methylene Chloride	0.716	0.691		-3.49	50
trans-1,2-Dichloroethene	0.596	0.592		-0.67	50
1,1-Dichloroethane	1.219	1.261	0.1	3.44	50
2-Butanone	0.543	0.593		9.21	50
Carbon Tetrachloride	0.544	0.552		1.47	50
cis-1,2-Dichloroethene	0.718	0.724		0.84	50
Chloroform	1.271	1.306		2.75	50
1,1,1-Trichloroethane	1.101	1.149		4.36	50
Methylcyclohexane	0.623	0.600		-3.69	50
Benzene	1.417	1.453		2.54	50
1,2-Dichloroethane	0.612	0.624		1.96	50
Trichloroethene	0.341	0.339		-0.59	50
1,2-Dichloropropane	0.352	0.379		7.67	50
Bromodichloromethane	0.547	0.579		5.85	50
4-Methyl-2-Pentanone	0.605	0.676		11.74	50
Toluene	0.869	0.895		2.99	50
t-1,3-Dichloropropene	0.487	0.511		4.93	50
cis-1,3-Dichloropropene	0.538	0.563		4.65	50
1,1,2-Trichloroethane	0.343	0.367		7	50
2-Hexanone	0.448	0.498		11.16	50
Dibromochloromethane	0.376	0.408		8.51	50
Tetrachloroethene	0.354	0.324		-8.48	50
Chlorobenzene	1.094	1.101	0.3	0.64	50
Ethyl Benzene	1.929	2.000		3.68	50
m/p-Xylenes	0.706	0.733		3.82	50
o-Xylene	0.688	0.721		4.8	50
Styrene	1.127	1.224		8.61	50
Bromoform	0.281	0.290	0.1	3.2	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG No.: Q2050

Instrument ID: MSVOA_X Calibration Date/Time: 05/15/2025 20:01

Lab File ID: VX046226.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.893	4.011		3.03	50
1,1,2,2-Tetrachloroethane	1.364	1.352	0.3	-0.88	50
1,3-Dichlorobenzene	1.649	1.651		0.12	50
1,4-Dichlorobenzene	1.684	1.597		-5.17	50
1,2-Dichlorobenzene	1.655	1.632		-1.39	50
1,2-Dichloroethane-d4	0.932	0.923		-0.97	50
Dibromofluoromethane	0.360	0.369		2.5	50
Toluene-d8	1.246	1.242		-0.32	50
4-Bromofluorobenzene	0.478	0.482		0.84	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

LAB CHRONICLE

OrderID:	Q2050	OrderDate:	5/15/2025 9:46:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	L31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2050-02	BP-VPB-182-GW-60-6 2	Water			05/12/25			05/14/25
			SVOC-SIMGroup1	8270-Modified		05/16/25	05/19/25	
Q2050-04	BP-VPB-182-GW-150- 152	Water			05/13/25			05/14/25
			SVOC-SIMGroup1	8270-Modified		05/16/25	05/19/25	
Q2050-06	BP-VPB-182-GW-220- 222	Water			05/14/25			05/14/25
			SVOC-SIMGroup1	8270-Modified		05/16/25	05/19/25	
Q2050-08	BP-VPB-182-EB-2025 0514	Water			05/14/25			05/14/25
			SVOC-SIMGroup1	8270-Modified		05/16/25	05/19/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q2050
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :	BP-VPB-182-GW-220-222							
Q2050-06	BP-VPB-182-GW-220-22 WATER	1,4-Dioxane	0.130	J	0.08	0.23	0.23	ug/L
		Total Svoc :	0.13					
		Total Concentration:	0.13					



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/12/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/14/25
Client Sample ID:	BP-VPB-182-GW-60-62	SDG No.:	Q2050
Lab Sample ID:	Q2050-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	550 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037048.D	1	05/16/25 08:42	05/19/25 16:51	PB168032

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.36	U	0.12	0.36	0.36	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.24		30 - 150		60%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.39		30 - 150		98%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.26		55 - 111		65%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.28		53 - 106		71%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.52		58 - 132		129%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2090	7.611				
1146-65-2	Naphthalene-d8	5700	10.394				
15067-26-2	Acenaphthene-d10	3310	14.256				
1517-22-2	Phenanthrene-d10	6660	17.009				
1719-03-5	Chrysene-d12	6100	21.206				
1520-96-3	Perylene-d12	5200	23.409				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/13/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/14/25
Client Sample ID:	BP-VPB-182-GW-150-152	SDG No.:	Q2050
Lab Sample ID:	Q2050-04	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	910 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037049.D	1	05/16/25 08:42	05/19/25 17:27	PB168032

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.22	U	0.070	0.22	0.22	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.28		30 - 150		69%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		93%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.28		55 - 111		71%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		83%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.58	*	58 - 132		145%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2170	7.611				
1146-65-2	Naphthalene-d8	5840	10.394				
15067-26-2	Acenaphthene-d10	3380	14.267				
1517-22-2	Phenanthrene-d10	6630	17.009				
1719-03-5	Chrysene-d12	5280	21.207				
1520-96-3	Perylene-d12	4530	23.407				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/14/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/14/25
Client Sample ID:	BP-VPB-182-GW-220-222	SDG No.:	Q2050
Lab Sample ID:	Q2050-06	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	870 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037050.D	1	05/16/25 08:42	05/19/25 18:03	PB168032

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.13	J	0.080	0.23	0.23	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.27		30 - 150		68%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.36		30 - 150		90%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.27		55 - 111		68%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.32		53 - 106		80%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.49		58 - 132		122%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1930	7.611				
1146-65-2	Naphthalene-d8	5280	10.394				
15067-26-2	Acenaphthene-d10	3090	14.256				
1517-22-2	Phenanthrene-d10	6260	17.009				
1719-03-5	Chrysene-d12	4980	21.207				
1520-96-3	Perylene-d12	4280	23.41				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/14/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/14/25
Client Sample ID:	BP-VPB-182-EB-20250514	SDG No.:	Q2050
Lab Sample ID:	Q2050-08	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	950 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037051.D	1	05/16/25 08:42	05/19/25 18:40	PB168032

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.21	U	0.070	0.21	0.21	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.27		30 - 150		68%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.36		30 - 150		90%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.28		55 - 111		69%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		53 - 106		84%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.54	*	58 - 132		134%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1900	7.618				
1146-65-2	Naphthalene-d8	5250	10.394				
15067-26-2	Acenaphthene-d10	3120	14.267				
1517-22-2	Phenanthrene-d10	6080	17.009				
1719-03-5	Chrysene-d12	4710	21.207				
1520-96-3	Perylene-d12	4050	23.412				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2050

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168032BL	PB168032BL	2-Methylnaphthalene-d10	0.4	0.36	90		30	150
		Fluoranthene-d10	0.4	0.34	84		30	150
		Nitrobenzene-d5	0.4	0.32	81		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.42	104		58	132
PB168032BS	PB168032BS	2-Methylnaphthalene-d10	0.4	0.39	96		30	150
		Fluoranthene-d10	0.4	0.33	81		30	150
		Nitrobenzene-d5	0.4	0.32	81		55	111
		2-Fluorobiphenyl	0.4	0.33	83		53	106
		Terphenyl-d14	0.4	0.39	97		58	132
PB168032BSD	PB168032BSD	2-Methylnaphthalene-d10	0.4	0.39	98		30	150
		Fluoranthene-d10	0.4	0.33	83		30	150
		Nitrobenzene-d5	0.4	0.35	88		55	111
		2-Fluorobiphenyl	0.4	0.33	83		53	106
		Terphenyl-d14	0.4	0.42	104		58	132
Q2050-02	BP-VPB-182-GW-60-62	2-Methylnaphthalene-d10	0.4	0.24	60		30	150
		Fluoranthene-d10	0.4	0.39	98		30	150
		Nitrobenzene-d5	0.4	0.26	65		55	111
		2-Fluorobiphenyl	0.4	0.28	71		53	106
		Terphenyl-d14	0.4	0.52	129		58	132
Q2050-04	BP-VPB-182-GW-150-152	2-Methylnaphthalene-d10	0.4	0.28	69		30	150
		Fluoranthene-d10	0.4	0.37	93		30	150
		Nitrobenzene-d5	0.4	0.28	71		55	111
		2-Fluorobiphenyl	0.4	0.33	83		53	106
		Terphenyl-d14	0.4	0.58	145	*	58	132
Q2050-06	BP-VPB-182-GW-220-222	2-Methylnaphthalene-d10	0.4	0.27	68		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.27	68		55	111
		2-Fluorobiphenyl	0.4	0.32	80		53	106
		Terphenyl-d14	0.4	0.49	122		58	132
Q2050-08	BP-VPB-182-EB-20250514	2-Methylnaphthalene-d10	0.4	0.27	68		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.28	69		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.54	134	*	58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2050

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN037080.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168032BS	1,4-Dioxane	0.4	0.31	ug/L	78				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2050

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN037081.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168032BSD	1,4-Dioxane	0.4	0.32	ug/L	80	3			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168032BL

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q2050

SAS No.: Q2050 SDG NO.: Q2050

Lab File ID: BN037075.D

Lab Sample ID: PB168032BL

Instrument ID: BNA_N

Date Extracted: 05/16/2025

Matrix: (soil/water) Water

Date Analyzed: 05/20/2025

Level: (low/med) LOW

Time Analyzed: 15:45

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB168032BS	PB168032BS	BN037080.D	05/20/2025
PB168032BSD	PB168032BSD	BN037081.D	05/20/2025
BP-VPB-182-GW-60-62	Q2050-02	BN037048.D	05/19/2025
BP-VPB-182-GW-150-152	Q2050-04	BN037049.D	05/19/2025
BP-VPB-182-GW-220-222	Q2050-06	BN037050.D	05/19/2025
BP-VPB-182-EB-20250514	Q2050-08	BN037051.D	05/19/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2050

SDG NO.: Q2050

Lab File ID: BN036998.D

DFTPP Injection Date: 05/13/2025

Instrument ID: BNA_N

DFTPP Injection Time: 17:02

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	62.8
68	Less than 2.0% of mass 69	0.8 (1.4) 1
69	Mass 69 relative abundance	55.6
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	52.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	8.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.4 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036999.D	05/13/2025	17:41
SSTDICC0.2	SSTDICC0.2	BN037000.D	05/13/2025	18:17
SSTDICCC0.4	SSTDICCC0.4	BN037001.D	05/13/2025	18:53
SSTDICC0.8	SSTDICC0.8	BN037002.D	05/13/2025	19:29
SSTDICC1.6	SSTDICC1.6	BN037003.D	05/13/2025	20:05
SSTDICC3.2	SSTDICC3.2	BN037004.D	05/13/2025	20:41
SSTDICC5.0	SSTDICC5.0	BN037005.D	05/13/2025	21:17

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2050

SDG NO.: Q2050

Lab File ID: BN037037.D

DFTPP Injection Date: 05/19/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	62.3
68	Less than 2.0% of mass 69	0.8 (1.4) 1
69	Mass 69 relative abundance	55.6
70	Less than 2.0% of mass 69	0.4 (0.7) 1
127	10.0 - 80.0% of mass 198	51.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	24.7
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	8.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.1 (18) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037038.D	05/19/2025	10:47
BP-VPB-182-GW-60-62	Q2050-02	BN037048.D	05/19/2025	16:51
BP-VPB-182-GW-150-152	Q2050-04	BN037049.D	05/19/2025	17:27
BP-VPB-182-GW-220-222	Q2050-06	BN037050.D	05/19/2025	18:03
BP-VPB-182-EB-20250514	Q2050-08	BN037051.D	05/19/2025	18:40
SSTDCCC0.4EC	SSTDCCC0.4	BN037055.D	05/19/2025	21:05

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2050

SDG NO.: Q2050

Lab File ID: BN037073.D

DFTPP Injection Date: 05/20/2025

Instrument ID: BNA_N

DFTPP Injection Time: 14:30

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	69.3
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	59.3
70	Less than 2.0% of mass 69	0.3 (0.4) 1
127	10.0 - 80.0% of mass 198	52.8
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	25.4
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	8.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.5 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037074.D	05/20/2025	15:09
PB168032BL	PB168032BL	BN037075.D	05/20/2025	15:45
PB168032BS	PB168032BS	BN037080.D	05/20/2025	18:47
PB168032BSD	PB168032BSD	BN037081.D	05/20/2025	19:23
SSTDCCC0.4EC	SSTDCCC0.4	BN037082.D	05/20/2025	19:59

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/19/2025

Lab File ID: BN037038.D Time Analyzed: 10:47

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2152	7.611	5916	10.39	3512	14.26
UPPER LIMIT	4304	8.111	11832	10.894	7024	14.756
LOWER LIMIT	1076	7.111	2958	9.894	1756	13.756
EPA SAMPLE NO.						
01 BP-VPB-182-GW-60-62	2086	7.61	5701	10.39	3309	14.26
02 BP-VPB-182-GW-150-152	2171	7.61	5837	10.39	3378	14.27
03 BP-VPB-182-GW-220-222	1931	7.61	5282	10.39	3090	14.26
04 BP-VPB-182-EB-20250514	1904	7.62	5247	10.39	3115	14.27

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/19/2025

Lab File ID: BN037038.D Time Analyzed: 10:47

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	7043	17.009	5607	21.207	4830	23.41
UPPER LIMIT	14086	17.509	11214	21.707	9660	23.91
LOWER LIMIT	3521.5	16.509	2803.5	20.707	2415	22.91
EPA SAMPLE NO.						
01 BP-VPB-182-GW-60-62	6655	17.01	6097	21.21	5198	23.41
02 BP-VPB-182-GW-150-152	6626	17.01	5279	21.21	4526	23.41
03 BP-VPB-182-GW-220-222	6262	17.01	4980	21.21	4275	23.41
04 BP-VPB-182-EB-20250514	6080	17.01	4710	21.21	4050	23.41

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/20/2025

Lab File ID: BN037074.D Time Analyzed: 15:09

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2240	7.611	6230	10.39	3461	14.26
UPPER LIMIT	4480	8.111	12460	10.894	6922	14.756
LOWER LIMIT	1120	7.111	3115	9.894	1730.5	13.756
EPA SAMPLE NO.						
01 PB168032BL	1695	7.61	4260	10.39	2468	14.26
02 PB168032BS	1842	7.61	4702	10.39	2635	14.26
03 PB168032BSD	1841	7.61	4693	10.39	2613	14.26

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/20/2025

Lab File ID: BN037074.D Time Analyzed: 15:09

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6762	17.009	4645	21.207	3820	23.407
UPPER LIMIT	13524	17.509	9290	21.707	7640	23.907
LOWER LIMIT	3381	16.509	2322.5	20.707	1910	22.907
EPA SAMPLE NO.						
01 PB168032BL	5072	17.01	3622	21.21	3128	23.41
02 PB168032BS	5352	17.01	4047	21.21	3455	23.40
03 PB168032BSD	5111	17.01	3649	21.20	3038	23.40

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB168032BL		SDG No.:	Q2050
Lab Sample ID:	PB168032BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037075.D	1	05/16/25 08:42	05/20/25 15:45	PB168032

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		90%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.34		30 - 150		84%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		81%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		53 - 106		84%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		58 - 132		104%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1700	7.611				
1146-65-2	Naphthalene-d8	4260	10.393				
15067-26-2	Acenaphthene-d10	2470	14.256				
1517-22-2	Phenanthrene-d10	5070	17.008				
1719-03-5	Chrysene-d12	3620	21.206				
1520-96-3	Perylene-d12	3130	23.409				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:		
Client Sample ID:	PB168032BS		SDG No.:	Q2050	
Lab Sample ID:	PB168032BS		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037080.D	1	05/16/25 08:42	05/20/25 18:47	PB168032

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.31		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.39		30 - 150		96%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		81%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		81%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		83%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.39		58 - 132		97%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1840	7.611				
1146-65-2	Naphthalene-d8	4700	10.394				
15067-26-2	Acenaphthene-d10	2640	14.256				
1517-22-2	Phenanthrene-d10	5350	17.009				
1719-03-5	Chrysene-d12	4050	21.206				
1520-96-3	Perylene-d12	3460	23.404				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB168032BSD		SDG No.:	Q2050
Lab Sample ID:	PB168032BSD		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037081.D	1	05/16/25 08:42	05/20/25 19:23	PB168032

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.32		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.39		30 - 150		98%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		83%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.35		55 - 111		88%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		83%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		58 - 132		104%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1840	7.611				
1146-65-2	Naphthalene-d8	4690	10.394				
15067-26-2	Acenaphthene-d10	2610	14.256				
1517-22-2	Phenanthrene-d10	5110	17.009				
1719-03-5	Chrysene-d12	3650	21.198				
1520-96-3	Perylene-d12	3040	23.401				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN051425.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 14 11:26:32 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN036999.D 0.2 =BN037000.D 0.4 =BN037001.D 0.8 =BN037002.D 1.6 =BN037003.D 3.2 =BN037004.D 5.0 =BN037005.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane		0.510	0.512	0.487	0.514	0.467	0.454	0.491	5.25
3)	n-Nitrosodimet...	1.465	0.974	0.980	0.971	1.075	0.967	0.950	1.054	17.59
4) S	2-Fluorophenol		1.101	1.134	1.024	1.093	0.964	0.971	1.048	6.87
5) S	Phenol-d6		1.304	1.385	1.236	1.392	1.259	1.292	1.311	4.91
6)	bis(2-Chloroet...	1.441	1.163	1.153	1.135	1.240	1.168	1.148	1.207	9.02
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.546	0.383	0.398	0.400	0.452	0.426	0.442	0.436	12.60
9)	Naphthalene	1.326	1.140	1.144	1.122	1.226	1.152	1.165	1.182	6.05
10)	Hexachlorobuta...	0.286	0.248	0.244	0.235	0.256	0.236	0.233	0.248	7.47
11) SURR2	Methylnaphth...	0.529	0.547	0.552	0.548	0.603	0.574	0.588	0.563	4.65
12)	2-Methylnaphth...	0.754	0.724	0.736	0.733	0.814	0.770	0.790	0.760	4.34
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.174	0.168	0.178	0.160	0.186	0.175	0.189	0.176	5.77
15) S	2-Fluorobiphenyl	1.912	1.801	1.901	1.802	1.927	1.672	1.807	1.832	4.90
16)	Acenaphthylene	1.906	1.838	1.894	1.849	2.071	1.997	2.075	1.947	5.14
17)	Acenaphthene	1.255	1.229	1.243	1.217	1.350	1.298	1.315	1.272	3.89
18)	Fluorene	1.602	1.581	1.635	1.611	1.779	1.721	1.752	1.669	4.80
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...		0.060	0.073	0.079	0.102	0.103	0.124	0.090	26.02
21)	4-Bromophenyl-...	0.243	0.246	0.250	0.247	0.262	0.261	0.259	0.253	3.13
22)	Hexachlorobenzene	0.267	0.269	0.281	0.259	0.281	0.270	0.267	0.270	3.03
23)	Atrazine	0.199	0.207	0.211	0.213	0.237	0.234	0.242	0.220	7.64
24)	Pentachlorophenol	0.133	0.134	0.141	0.137	0.159	0.162	0.177	0.149	11.45
25)	Phenanthrene	1.259	1.272	1.292	1.263	1.367	1.337	1.361	1.307	3.56
26)	Anthracene	1.099	1.104	1.166	1.130	1.269	1.259	1.300	1.190	7.13
27) SURR	Fluoranthene-d10	1.033	1.033	1.078	1.042	1.153	1.161	1.178	1.097	5.95
28)	Fluoranthene	1.461	1.439	1.500	1.496	1.670	1.672	1.693	1.562	7.13
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.744	1.708	1.727	1.656	1.790	1.641	1.711	1.711	2.96
31) S	Terphenyl-d14	0.897	0.844	0.871	0.822	0.891	0.816	0.848	0.856	3.73
32)	Benzo(a)anthra...	1.463	1.432	1.485	1.438	1.594	1.521	1.609	1.506	4.77
33)	Chrysene	1.655	1.559	1.616	1.532	1.653	1.560	1.576	1.593	3.05
34)	Bis(2-ethylhex...	0.955	0.919	0.906	0.855	0.941	0.903	1.011	0.927	5.27
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN051425.M

36)	Indeno(1,2,3-c...	1.511	1.613	1.645	1.568	1.687	1.732	1.680	1.634	4.65
37)	Benzo(b)fluora...	1.631	1.570	1.602	1.599	1.749	1.698	1.765	1.659	4.71
38)	Benzo(k)fluora...	1.539	1.538	1.642	1.601	1.770	1.661	1.719	1.639	5.34
39) C	Benzo(a)pyrene	1.380	1.343	1.381	1.331	1.486	1.444	1.486	1.407	4.59
40)	Dibenzo(a,h)an...	1.116	1.232	1.273	1.237	1.340	1.376	1.334	1.272	6.90
41)	Benzo(g,h,i)pe...	1.299	1.407	1.424	1.330	1.403	1.439	1.376	1.383	3.72

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG No.: Q2050

Instrument ID: BNA_N Calibration Date/Time: 05/19/2025 10:47

Lab File ID: BN037038.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.598		6.2	20.0
Fluoranthene-d10	1.097	1.086		-1.0	20.0
2-Fluorophenol	1.048	0.944		-9.9	20.0
Phenol-d6	1.311	1.169		-10.8	20.0
Nitrobenzene-d5	0.436	0.392		-10.1	20.0
2-Fluorobiphenyl	1.832	1.706		-6.9	20.0
2,4,6-Tribromophenol	0.176	0.147		-16.5	20.0
Terphenyl-d14	0.856	0.915		6.9	20.0
1,4-Dioxane	0.491	0.461		-6.1	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG No.: Q2050
 Instrument ID: BNA_N Calibration Date/Time: 05/19/2025 21:05
 Lab File ID: BN037055.D Init. Calib. Date(s): 05/13/2025 05/13/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 17:41 21:17
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.581		3.2	50.0
Fluoranthene-d10	1.097	1.058		-3.6	50.0
2-Fluorophenol	1.048	0.956		-8.8	50.0
Phenol-d6	1.311	1.171		-10.7	50.0
Nitrobenzene-d5	0.436	0.406		-6.9	50.0
2-Fluorobiphenyl	1.832	1.804		-1.5	50.0
2,4,6-Tribromophenol	0.176	0.160		-9.1	50.0
Terphenyl-d14	0.856	0.927		8.3	50.0
1,4-Dioxane	0.491	0.456		-7.1	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG No.: Q2050

Instrument ID: BNA_N Calibration Date/Time: 05/20/2025 15:09

Lab File ID: BN037074.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.573		1.8	20.0
Fluoranthene-d10	1.097	1.010		-7.9	20.0
2-Fluorophenol	1.048	0.964		-8.0	20.0
Phenol-d6	1.311	1.198		-8.6	20.0
Nitrobenzene-d5	0.436	0.402		-7.8	20.0
2-Fluorobiphenyl	1.832	1.798		-1.9	20.0
2,4,6-Tribromophenol	0.176	0.153		-13.1	20.0
Terphenyl-d14	0.856	0.940		9.8	20.0
1,4-Dioxane	0.491	0.488		-0.6	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG No.: Q2050

Instrument ID: BNA_N Calibration Date/Time: 05/20/2025 19:59

Lab File ID: BN037082.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.581		3.2	50.0
Fluoranthene-d10	1.097	1.017		-7.3	50.0
2-Fluorophenol	1.048	0.946		-9.7	50.0
Phenol-d6	1.311	1.150		-12.3	50.0
Nitrobenzene-d5	0.436	0.417		-4.4	50.0
2-Fluorobiphenyl	1.832	1.796		-2.0	50.0
2,4,6-Tribromophenol	0.176	0.165		-6.3	50.0
Terphenyl-d14	0.856	0.957		11.8	50.0
1,4-Dioxane	0.491	0.473		-3.7	50.0

All other compounds must meet a minimum RRF of 0.010.



SHIPPING DOCUMENTS

CLIENT INFORMATION						PROJECT INFORMATION							BILLING INFORMATION																	
COMPANY: Tetra Tech ADDRESS: 4433 Corporation Lane Suite 300 CITY: Virginia Beach STATE: VA ZIP: 23462 ATTENTION: Ernie Wu PHONE: 757-466-4901 FAX: 757-461-4148						PROJECT NAME: NWIRP Bethpage PROJECT #: 112G08005-WF13 LOCATION: VPB-182 PROJECT MANAGER: Ernie Wu E-MAIL: ernie.wu@tetratech.com PHONE: 757-466-4901 FAX: 757-461-4148							BILL TO: SEE CONTRACT PO# ADDRESS: CITY: STATE: ZIP: ATTENTION: PHONE:																	
DATA TURNAROUND INFORMATION						DATA DELIVERABLE INFORMATION							ANALYSIS																	
FAX: _____ 2 & 10 DAYS* HARD COPY: _____ 2 & 10 DAYS* EDD _____ 2 & 10 DAYS* * TO BE APPROVED BY CHEMTECH STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS						☐ RESEULTS ONLY ☐ USEPA CLP ☐ RESULTS + QC ☐ New York State ASP "B" ☐ New Jersey REDUCED ☐ New York State ASP "A" ☐ New Jersey CLP ☐ Other _____ ☐ EDD Format _____							VOC(SW846-32005) 1,4 Dioxane (8270 SIM)																	
													PRESERVATIVES																	
													COMMENTS																	
CHEMTECH SAMPLE ID						PROJECT SAMPLE IDENTIFICATION						SAMPLE MATRIX		SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles		Specify Preservatives										
												COMP		GRAB		DATE		TIME				<- Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-ICE F-Other								
1.						BP-VPB-182-TB-20250512						QA				X		5/12/25 9:00		2		Trip Blank								
2.						BP-VPB-182-GW-60-62						AQ				X		5/12/25 11:50		3										
3.						BP-VPB-182-GW-100-102						AQ				X		5/12/25 14:00		3										
4.						BP-VPB-182-GW-150-152						AQ				X		5/13/25 10:25		3										
5.						BP-VPB-182-GW-205-207						AQ				X		5/14/25 9:50		3		Extra VOA vial								
6.						BP-VPB-182-GW-220-222						AQ				X		5/14/25 11:55		3										
7.						BP-VPB-182-DUP-20250514						QA				X		5/14/25 12:00		2		8260B Duplicate								
8.						BP-VPB-182-EB-20250514						QA				X		5/14/25 12:40		3		Equipment Blank								
9.						BP-VPB-182-GW-240-242						AQ				X		5/14/25 14:05		2										
10.																														
SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY																														
RELINQUISHED BY SAMPLER						DATE/TIME						RECEIVED BY						Conditions of bottles or coolers at receipt: q Compliant q Non Compliant q Cooler Temp 2.5°C												
1. [Signature]						5/14/25 1600						[Signature] 5-14-25						MeOH extraction requires an additional 4oz. Jar for percent solid q Ice in												
RELINQUISHED BY						DATE/TIME						RECEIVED BY						Cooler?:												
2.												2.						Comments: 48hr TAT - For VOC's see worksheet #15 of SAP 2018 for VPB program VOC list												
RELINQUISHED BY						DATE/TIME						RECEIVED FOR LAB BY						10-DAY TAT - For 1,4 Dioxane (8270 SIM)												
3. [Signature]						5-14-25 1935						3.						SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight ☐ Picked Up ☐ Overnight												
																		Page 1 of 1												
																		SHIPMENT Complete ☐ YES ☐ NO												
WHITE - CHEMTECH COPYFOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY																														

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2050 TETR06

Order Date : 5/15/2025 9:46:00 AM

Project Mgr :

Client Name : Tetra Tech NUS, Inc.

Project Name : NWIRP Bethpage 112G080

Report Type : Level 4

Client Contact : Ernie Wu

Receive DateTime : 5/14/2025 7:35:00 PM

EDD Type : ADAPT

Invoice Name : Tetra Tech NUS, Inc.

Purchase Order :

Hard Copy Date :

Invoice Contact : Ernie Wu

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2050-01	BP-TB-20250512	Water	05/12/2025	09:00	VOCMS Group1		8260-Low	10 Bus. Days	2 Days
Q2050-02	BP-VPB-182-GW-60-62	Water	05/12/2025	11:50	VOCMS Group1		8260-Low	10 Bus. Days	
Q2050-03	BP-VPB-182-GW-100-102	Water	05/12/2025	14:00	VOCMS Group1		8260-Low	10 Bus. Days	
Q2050-04	BP-VPB-182-GW-150-152	Water	05/13/2025	10:25	VOCMS Group1		8260-Low	10 Bus. Days	
Q2050-06	BP-VPB-182-GW-220-222	Water	05/14/2025	11:55	VOCMS Group1		8260-Low	10 Bus. Days	
Q2050-07	BP-VPB-182-DUP-20250514	Water	05/14/2025	12:00	VOCMS Group1		8260-Low	10 Bus. Days	
Q2050-08	BP-VPB-182-EB-20250514	Water	05/14/2025	12:40	VOCMS Group1		8260-Low	10 Bus. Days	
Q2050-09	BP-VPB-182-GW-240-242	Water	05/14/2025	14:05	VOCMS Group1		8260-Low	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2050	TETR06	Order Date : 5/15/2025 9:46:00 AM	Project Mgr :
Client Name : Tetra Tech NUS, Inc.		Project Name : NWIRP Bethpage 112G080	Report Type : Level 4
Client Contact : Ernie Wu		Receive DateTime : 5/14/2025 7:35:00 PM	EDD Type : ADAPT
Invoice Name : Tetra Tech NUS, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Ernie Wu			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
--------	-----------	--------	-------------	-------------	------	------------	--------	----------	-----------

VOCMS Group1

8260-Low

~~10~~ Bus. Days

2 Days

Relinquished By :

Date / Time : 5/15/25 1045

Received By :

Date / Time :

Storage Area : VOA Refridgerator Room

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2050 TETR06

Order Date : 5/15/2025 9:46:00 AM

Project Mgr :

Client Name : Tetra Tech NUS, Inc.

Project Name : NWIRP Bethpage 112G080

Report Type : Level 4

Client Contact : Ernie Wu

Receive DateTime : 5/14/2025 7:35:00 PM

EDD Type : ADAPT

Invoice Name : Tetra Tech NUS, Inc.

Purchase Order :

Hard Copy Date :

Invoice Contact : Ernie Wu

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2050-05	BP-VPB-182-GW-205-207	Water	05/14/2025	09:50					

VOCMS Group1

8260-Low

~~10~~ Bus. Days

2 Days

Ref #04 VOA

Relinquished By :

Date / Time : 5/15/25 11:20

Received By :

Date / Time : 5/15/25 11:20

Storage Area : VOA Refridgerator Room