

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/20/2025

Lab File ID: BN037074.D Time Analyzed: 15:09

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2240	7.611	6230	10.39	3461	14.26
UPPER LIMIT	4480	8.111	12460	10.894	6922	14.756
LOWER LIMIT	1120	7.111	3115	9.894	1730.5	13.756
EPA SAMPLE NO.						
01 PB168032BL	1695	7.61	4260	10.39	2468	14.26
02 PB168032BS	1842	7.61	4702	10.39	2635	14.26
03 PB168032BSD	1841	7.61	4693	10.39	2613	14.26

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.