

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/20/2025

Lab File ID: BN037074.D Time Analyzed: 15:09

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6762	17.009	4645	21.207	3820	23.407
UPPER LIMIT	13524	17.509	9290	21.707	7640	23.907
LOWER LIMIT	3381	16.509	2322.5	20.707	1910	22.907
EPA SAMPLE NO.						
01 PB168032BL	5072	17.01	3622	21.21	3128	23.41
02 PB168032BS	5352	17.01	4047	21.21	3455	23.40
03 PB168032BSD	5111	17.01	3649	21.20	3038	23.40

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.