

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06  
Lab Code: CHEM Case No.: Q2050 SAS No.: Q2050 SDG NO.: Q2050  
Lab File ID: VX046200.D Date Analyzed: 05/15/2025  
Instrument ID: MSVOA\_X Time Analyzed: 09:52  
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                         | IS4<br>AREA # | RT #   |  |  |  |  |
|-------------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD             | 62876         | 12.018 |  |  |  |  |
| UPPER LIMIT             | 125752        | 12.518 |  |  |  |  |
| LOWER LIMIT             | 31438         | 11.518 |  |  |  |  |
| EPA SAMPLE NO.          |               |        |  |  |  |  |
| BP-VPB-182-TB-20250512  | 51079         | 12.02  |  |  |  |  |
| BP-VPB-182-GW-60-62     | 50521         | 12.02  |  |  |  |  |
| BP-VPB-182-GW-100-102   | 47647         | 12.02  |  |  |  |  |
| BP-VPB-182-GW-150-152   | 51539         | 12.02  |  |  |  |  |
| BP-VPB-182-GW-205-207   | 47407         | 12.02  |  |  |  |  |
| BP-VPB-182-GW-220-222   | 44995         | 12.02  |  |  |  |  |
| BP-VPB-182-DUP-20250514 | 50006         | 12.02  |  |  |  |  |
| BP-VPB-182-EB-20250514  | 50288         | 12.02  |  |  |  |  |
| BP-VPB-182-GW-240-242   | 50190         | 12.02  |  |  |  |  |
| VX0515WBL01             | 51360         | 12.02  |  |  |  |  |
| VX0515WBS01             | 65777         | 12.02  |  |  |  |  |
| VX0515WBSD01            | 63406         | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.