

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051925\
 Data File : BN037055.D
 Acq On : 19 May 2025 21:05
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 19 23:47:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.491	0.456	7.1	114	0.00
3	n-Nitrosodimethylamine	1.054	1.031	2.2	135	0.00
4 S	2-Fluorophenol	1.048	0.956	8.8	108	0.00
5 S	Phenol-d6	1.311	1.171	10.7	109	0.00
6	bis(2-Chloroethyl)ether	1.207	1.137	5.8	127	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	131	-0.01
8 S	Nitrobenzene-d5	0.436	0.406	6.9	133	-0.01
9	Naphthalene	1.182	1.110	6.1	127	0.00
10	Hexachlorobutadiene	0.248	0.254	-2.4	136	-0.01
11 SURR	2-Methylnaphthalene-d10	0.563	0.581	-3.2	138	0.00
12	2-Methylnaphthalene	0.760	0.711	6.4	126	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	133	-0.01
14 S	2,4,6-Tribromophenol	0.176	0.160	9.1	120	-0.01
15 S	2-Fluorobiphenyl	1.832	1.804	1.5	126	0.00
16	Acenaphthylene	1.947	1.766	9.3	124	-0.01
17	Acenaphthene	1.272	1.184	6.9	127	-0.01
18	Fluorene	1.669	1.574	5.7	128	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	133	0.00
20	4,6-Dinitro-2-methylphenol	0.090	0.074	17.8	135	-0.01
21	4-Bromophenyl-phenylether	0.253	0.245	3.2	131	0.00
22	Hexachlorobenzene	0.270	0.280	-3.7	133	0.00
23	Atrazine	0.220	0.198	10.0	125	0.00
24	Pentachlorophenol	0.149	0.124	16.8	117	-0.01
25	Phenanthrene	1.307	1.242	5.0	128	-0.01
26	Anthracene	1.190	1.088	8.6	124	0.00
27 SURR	Fluoranthene-d10	1.097	1.058	3.6	131	0.00
28	Fluoranthene	1.562	1.391	10.9	124	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
30	Pyrene	1.711	1.832	-7.1	121	0.00
31 S	Terphenyl-d14	0.856	0.927	-8.3	122	0.00
32	Benzo(a)anthracene	1.506	1.362	9.6	105	0.00
33	Chrysene	1.593	1.539	3.4	109	0.00
34	Bis(2-ethylhexyl)phthalate	0.927	0.838	9.6	106	0.00
35 I	Perylene-d12	1.000	1.000	0.0	97	0.00
36	Indeno(1,2,3-cd)pyrene	1.634	1.531	6.3	90	-0.01
37	Benzo(b)fluoranthene	1.659	1.532	7.7	93	0.00
38	Benzo(k)fluoranthene	1.639	1.601	2.3	94	0.00
39 C	Benzo(a)pyrene	1.407	1.272	9.6	89	0.00
40	Dibenzo(a,h)anthracene	1.272	1.166	8.3	89	-0.02
41	Benzo(g,h,i)perylene	1.383	1.342	3.0	91	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4