

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051425\
 Data File : BN036999.D
 Acq On : 13 May 2025 17:41
 Operator : RC/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: May 14 10:59:44 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 10:57:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.618	152	2512	0.400	ng	0.00
7) Naphthalene-d8	10.404	136	6713	0.400	ng	0.00
13) Acenaphthene-d10	14.267	164	3705	0.400	ng	0.00
19) Phenanthrene-d10	17.009	188	7492	0.400	ng	0.00
29) Chrysene-d12	21.216	240	6297	0.400	ng	# 0.00
35) Perylene-d12	23.418	264	6037	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	0.000	112	0d	0.000	ng	
5) Phenol-d6	0.000	99	0d	0.000	ng	
8) Nitrobenzene-d5	8.771	82	917	0.125	ng	0.00
11) 2-Methylnaphthalene-d10	12.001	152	888	0.094	ng	0.00
14) 2,4,6-Tribromophenol	15.767	330	161	0.099	ng	0.00
15) 2-Fluorobiphenyl	12.889	172	1771	0.104	ng	0.00
27) Fluoranthene-d10	19.049	212	1934	0.094	ng	0.00
31) Terphenyl-d14	19.658	244	1412	0.105	ng	0.00
Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.466	42	920	0.139	ng	# 94
6) bis(2-Chloroethyl)ether	7.048	93	905	0.119	ng	98
9) Naphthalene	10.447	128	2225	0.112	ng	97
10) Hexachlorobutadiene	10.735	225	480	0.115	ng	# 100
12) 2-Methylnaphthalene	12.072	142	1266	0.099	ng	98
16) Acenaphthylene	13.989	152	1765	0.098	ng	99
17) Acenaphthene	14.331	154	1162	0.099	ng	99
18) Fluorene	15.325	166	1484	0.096	ng	99
21) 4-Bromophenyl-phenylether	16.214	248	455	0.096	ng	98
22) Hexachlorobenzene	16.326	284	500	0.099	ng	97
23) Atrazine	16.487	200	372	0.090	ng	# 93
24) Pentachlorophenol	16.674	266	249	0.088	ng	97
25) Phenanthrene	17.058	178	2359	0.096	ng	99
26) Anthracene	17.145	178	2059	0.092	ng	99
28) Fluoranthene	19.082	202	2737	0.094	ng	100
30) Pyrene	19.444	202	2745	0.102	ng	100
32) Benzo(a)anthracene	21.198	228	2303	0.097	ng	96
33) Chrysene	21.251	228	2605	0.104	ng	97
34) Bis(2-ethylhexyl)phtha...	21.135	149	1504	0.103	ng	99
36) Indeno(1,2,3-cd)pyrene	25.629	276	2280	0.092	ng	97
37) Benzo(b)fluoranthene	22.758	252	2461	0.098	ng	# 75
38) Benzo(k)fluoranthene	22.801	252	2322	0.094	ng	# 73
39) Benzo(a)pyrene	23.322	252	2083	0.098	ng	# 62
40) Dibenzo(a,h)anthracene	25.646	278	1685	0.088	ng	# 78
41) Benzo(g,h,i)perylene	26.304	276	1960	0.094	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051425\
Data File : BN036999.D
Acq On : 13 May 2025 17:41
Operator : RC/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: May 14 10:59:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 14 10:57:36 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.618 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 2512
Ion Ratio Lower Upper
152 100
150 150.9 123.9 185.9
115 68.7 55.8 83.8



#3
n-Nitrosodimethylamine
Concen: 0.139 ng
RT: 3.466 min Scan# 81
Delta R.T. 0.007 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion: 42 Resp: 920
Ion Ratio Lower Upper
42 100
74 76.3 59.8 89.6
44 4.1 11.9 17.9#







#4
 2-Fluorophenol
 Concen: 0.000 ng
 Expected RT: 5.21 min
 Lab File: BN036999.D
 Acq: 13 May 2025 17:41

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion: 112


Sig	Exp Ratio
112	100
64	69.6
63	43.2




#5
 Phenol-d6
 Concen: 0.000 ng
 Expected RT: 6.79 min
 Lab File: BN036999.D
 Acq: 13 May 2025 17:41

Tgt Ion: 99

Sig	Exp Ratio
99	100
42	36.6
71	44.6









#6
bis(2-Chloroethyl)ether
Concen: 0.119 ng
RT: 7.048 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	93	Resp:	905
Ion	Ratio	Lower	Upper
93	100		
63	90.2	70.1	105.1
95	32.2	26.2	39.2



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.404 min Scan# 948
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion:	136	Resp:	6713
Ion	Ratio	Lower	Upper
136	100		
137	12.0	10.4	15.6
54	10.0	8.5	12.7
68	6.8	5.1	7.7









#8
Nitrobenzene-d5
Concen: 0.125 ng
RT: 8.771 min Scan# 795
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 82 Resp: 917
Ion Ratio Lower Upper
82 100
128 43.0 34.0 51.0
54 70.1 55.0 82.4



#9
Naphthalene
Concen: 0.112 ng
RT: 10.447 min Scan# 952
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41


Tgt Ion: 128 Resp: 2225
Ion Ratio Lower Upper
128 100
129 14.1 9.7 14.5
127 15.8 12.4 18.6








#10
Hexachlorobutadiene
Concen: 0.115 ng
RT: 10.735 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1







Tgt Ion:225 Resp: 480
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 50.9 76.3



#11
2-Methylnaphthalene-d10
Concen: 0.094 ng
RT: 12.001 min Scan# 1149
Delta R.T. 0.005 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion:152 Resp: 888
Ion Ratio Lower Upper
152 100
151 21.6 17.5 26.3









#12
2-Methylnaphthalene
Concen: 0.099 ng
RT: 12.072 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:142 Resp: 1266
Ion Ratio Lower Upper
142 100
141 89.9 73.3 109.9
115 50.5 38.4 57.6



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.267 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion:164 Resp: 3705
Ion Ratio Lower Upper
164 100
162 106.3 84.2 126.4
160 53.6 42.6 63.8









#14
2,4,6-Tribromophenol
Concen: 0.099 ng
RT: 15.767 min Scan# 1602
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	330	Resp:	161
Ion	Ratio	Lower	Upper
330	100		
332	87.0	73.8	110.8
141	60.2	43.9	65.9



#15
2-Fluorobiphenyl
Concen: 0.104 ng
RT: 12.889 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion:	172	Resp:	1771
Ion	Ratio	Lower	Upper
172	100		
171	37.5	29.2	43.8
170	26.1	20.5	30.7









#16
Acenaphthylene
Concen: 0.098 ng
RT: 13.989 min Scan# 1439
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:152 Resp: 1765
Ion Ratio Lower Upper
152 100
151 19.6 16.1 24.1
153 12.5 10.5 15.7



#17
Acenaphthene
Concen: 0.099 ng
RT: 14.331 min Scan# 1471
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion:154 Resp: 1162
Ion Ratio Lower Upper
154 100
153 117.2 94.2 141.4
152 62.1 49.4 74.0









#18
Fluorene
Concen: 0.096 ng
RT: 15.325 min Scan# 1564
Delta R.T. 0.011 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:166 Resp: 1484
Ion Ratio Lower Upper
166 100
165 102.0 80.6 120.8
167 13.3 10.6 16.0



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.009 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion:188 Resp: 7492
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 16.9 13.4 20.0









#21
4-Bromophenyl-phenylether
Concen: 0.096 ng
RT: 16.214 min Scan# 1638
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:248 Resp: 455
Ion Ratio Lower Upper
248 100
250 96.8 78.1 117.1
141 78.3 59.7 89.5



#22
Hexachlorobenzene
Concen: 0.099 ng
RT: 16.326 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion:284 Resp: 500
Ion Ratio Lower Upper
284 100
142 52.2 41.2 61.8
249 40.0 28.7 43.1





#23

Atrazine

Concen: 0.090 ng

RT: 16.487 min Scan# 1660

Delta R.T. 0.000 min

Lab File: BN036999.D

Acq: 13 May 2025 17:41

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:200 Resp: 372

Ion Ratio Lower Upper

200 100

173 38.1 25.2 37.8#

215 51.4 39.3 58.9



#24

Pentachlorophenol

Concen: 0.088 ng

RT: 16.674 min Scan# 1675

Delta R.T. 0.000 min

Lab File: BN036999.D

Acq: 13 May 2025 17:41




Tgt Ion:266 Resp: 249


Ion Ratio Lower Upper

266 100

264 59.8 47.9 71.9

268 66.7 50.0 75.0









#25
 Phenanthrene
 Concen: 0.096 ng
 RT: 17.058 min Scan# 1706
 Delta R.T. 0.000 min
 Lab File: BN036999.D
 Acq: 13 May 2025 17:41




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	178	Resp:	2359
Ion	Ratio	Lower	Upper
178	100		
176	20.3	15.7	23.5
179	15.7	12.2	18.2



#26
 Anthracene
 Concen: 0.092 ng
 RT: 17.145 min Scan# 1713
 Delta R.T. 0.000 min
 Lab File: BN036999.D
 Acq: 13 May 2025 17:41

Tgt Ion:	178	Resp:	2059
Ion	Ratio	Lower	Upper
178	100		
176	19.9	15.0	22.6
179	15.3	12.3	18.5









#27
Fluoranthene-d10
Concen: 0.094 ng
RT: 19.049 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:212 Resp: 1934
Ion Ratio Lower Upper
212 100
106 15.0 11.3 16.9
104 9.6 6.7 10.1



#28
Fluoranthene
Concen: 0.094 ng
RT: 19.082 min Scan# 1946
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion:202 Resp: 2737
Ion Ratio Lower Upper
202 100
101 11.4 8.9 13.3
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.216 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:240 Resp: 6297
Ion Ratio Lower Upper
240 100
120 11.0 15.1 22.7#
236 29.2 24.0 36.0



#30
Pyrene
Concen: 0.102 ng
RT: 19.444 min Scan# 2024
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion:202 Resp: 2745
Ion Ratio Lower Upper
202 100
200 21.3 17.1 25.7
203 18.1 14.2 21.4









#31
 Terphenyl-d14
 Concen: 0.105 ng
 RT: 19.658 min Scan# 2070
 Delta R.T. 0.000 min
 Lab File: BN036999.D
 Acq: 13 May 2025 17:41




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244 Resp: 1412
 Ion Ratio Lower Upper
 244 100
 212 14.3 9.7 14.5
 122 20.5 13.4 20.0#



#32
 Benzo(a)anthracene
 Concen: 0.097 ng
 RT: 21.198 min Scan# 2278
 Delta R.T. 0.000 min
 Lab File: BN036999.D
 Acq: 13 May 2025 17:41

Tgt Ion:228 Resp: 2303
 Ion Ratio Lower Upper
 228 100
 226 29.7 22.2 33.4
 229 21.6 16.0 24.0









#33
Chrysene
Concen: 0.104 ng
RT: 21.251 min Scan# 21251
Delta R.T. 0.009 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:228 Resp: 2605
Ion Ratio Lower Upper
228 100
226 31.5 26.3 39.5
229 22.0 16.2 24.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.103 ng
RT: 21.135 min Scan# 2271
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Tgt Ion:149 Resp: 1504
Ion Ratio Lower Upper
149 100
167 25.1 20.6 30.8
279 3.1 2.6 3.8









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.418 min Scan# 21
 Delta R.T. 0.003 min
 Lab File: BN036999.D
 Acq: 13 May 2025 17:41




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	264	Resp:	6037
Ion Ratio	Lower	Upper	
264	100		
260	25.9	21.9	32.9
265	56.5	51.6	77.4



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.092 ng
 RT: 25.629 min Scan# 3515
 Delta R.T. -0.003 min
 Lab File: BN036999.D
 Acq: 13 May 2025 17:41

Tgt Ion:	276	Resp:	2280
Ion Ratio	Lower	Upper	
276	100		
138	25.8	22.7	34.1
277	24.3	20.0	30.0









#37
Benzo(b)fluoranthene
Concen: 0.098 ng
RT: 22.758 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 2461
Ion Ratio Lower Upper
252 100
253 38.2 21.8 32.6#
125 32.2 14.6 21.8#



#38
Benzo(k)fluoranthene
Concen: 0.094 ng
RT: 22.801 min Scan# 2548
Delta R.T. 0.003 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41


Tgt Ion:252 Resp: 2322
Ion Ratio Lower Upper
252 100
253 39.6 21.4 32.2#
125 29.7 13.0 19.4#








#39
Benzo(a)pyrene
Concen: 0.098 ng
RT: 23.322 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1







Tgt Ion:252 Resp: 2083
Ion Ratio Lower Upper
252 100
253 45.5 23.8 35.6#
125 52.0 21.8 32.6#



#40
Dibenzo(a,h)anthracene
Concen: 0.088 ng
RT: 25.646 min Scan# 3521
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41


Tgt Ion:278 Resp: 1685
Ion Ratio Lower Upper
278 100
139 30.0 17.4 26.0#
279 45.1 24.6 37.0#





#41
Benzo(g,h,i)perylene
Concen: 0.094 ng
RT: 26.304 min Scan# 31
Delta R.T. 0.000 min
Lab File: BN036999.D
Acq: 13 May 2025 17:41

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1



Tgt Ion:276 Resp: 1960
Ion Ratio Lower Upper
276 100
277 30.1 20.2 30.4
138 35.5 22.0 33.0#

