

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN051425

Review By	Rahul	Review On	5/15/2025 9:39:49 AM
Supervise By	Jagrut	Supervise On	5/15/2025 3:50:12 PM
SubDirectory	BN051425	HP Acquire Method	BNA_N, 8270_HP Processing Method bn051425

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6779 SP6740,1ul/100ul sample SP6768

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036998.D	13 May 2025 17:02		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN036999.D	13 May 2025 17:41	Compound #02,04,05,20 removed from 0.1ppm	RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN037000.D	13 May 2025 18:17		RC/JU	Ok,M
4	SSTDICCC0.4	SSTDICCC0.4	BN037001.D	13 May 2025 18:53	Compound #20 kept on QR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN037002.D	13 May 2025 19:29		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN037003.D	13 May 2025 20:05		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN037004.D	13 May 2025 20:41		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN037005.D	13 May 2025 21:17		RC/JU	Ok
9	SSTDICV0.4	ICVBN051424	BN037006.D	13 May 2025 22:29		RC/JU	Ok
10	PB167888BL	PB167888BL	BN037007.D	13 May 2025 23:42	Analyzed for contamination check	RC/JU	Not Ok
11	DFTPP	DFTPP	BN037008.D	14 May 2025 09:37		RC/JU	Ok
12	SSTDCCC0.4	SSTDCCC0.4	BN037009.D	14 May 2025 10:31		RC/JU	Ok,M
13	PB167952BL	PB167952BL	BN037010.D	14 May 2025 11:20		RC/JU	Ok
14	Q2000-01	38072-062223	BN037011.D	14 May 2025 12:03	PT Sample, Need 50X Dilution	RC/JU	Dilution
15	Q2000-01DL	38072-062223DL	BN037012.D	14 May 2025 13:29		RC/JU	Ok
16	Q1872-13	HW0425-PT-PAH-SOIL	BN037013.D	14 May 2025 14:46	PT Sample, Need 10X Dilution	RC/JU	Dilution
17	Q1872-13DL	HW0425-PT-PAH-SOIL	BN037014.D	14 May 2025 15:48		RC/JU	Ok,M

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN051425

Review By	Rahul	Review On	5/15/2025 9:39:49 AM
Supervise By	Jagrut	Supervise On	5/15/2025 3:50:12 PM
SubDirectory	BN051425	HP Acquire Method	BNA_N, 8270_HP Processing Method bn051425
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775		
CCC	SP6779		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q1985-01	RW8-BW-20250507	BN037015.D	14 May 2025 16:48		RC/JU	Ok
19	Q2012-01	RW5-SP100-20250509	BN037016.D	14 May 2025 17:24	Need 2X Dilution	RC/JU	Dilution
20	Q2012-02	RW5-SP201-20250509	BN037017.D	14 May 2025 18:00		RC/JU	Ok
21	Q2012-03	RW5-SP303-20250509	BN037018.D	14 May 2025 18:36		RC/JU	Ok,M
22	Q2013-02	BP-TT192D2-GW-2025	BN037019.D	14 May 2025 19:12		RC/JU	Ok
23	Q2013-03	BP-TT192D1-GW-2025	BN037020.D	14 May 2025 19:48		RC/JU	Ok
24	PB167952BS	PB167952BS	BN037021.D	14 May 2025 20:24		RC/JU	Ok,M
25	PB167952BSD	PB167952BSD	BN037022.D	14 May 2025 21:00		RC/JU	Ok,M
26	SSTDCCC0.4	SSTDCCC0.4EC	BN037023.D	14 May 2025 21:36		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN051925

Review By	Rahul	Review On	5/20/2025 12:04:04 PM
Supervise By	Jagrut	Supervise On	5/20/2025 12:14:27 PM
SubDirectory	BN051925	HP Acquire Method	BNA_N, 8270_HP Processing Method bn051425

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6779 SP6740,1ul/100ul sample SP6768

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037037.D	19 May 2025 10:08		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037038.D	19 May 2025 10:47		RC/JU	Ok
3	PB168032BL	PB168032BL	BN037039.D	19 May 2025 11:26	Not Used	RC/JU	Not Ok
4	Q2064-01	RW8-SP100-70-202505	BN037040.D	19 May 2025 12:02		RC/JU	Ok
5	Q2064-02	RW8-SP201-70-202505	BN037041.D	19 May 2025 12:38		RC/JU	Ok
6	Q2064-03	RW8-SP301-70-202505	BN037042.D	19 May 2025 13:14		RC/JU	Ok
7	Q2064-04	RW8-SP303-70-202505	BN037043.D	19 May 2025 13:50		RC/JU	Ok
8	Q2064-05	RW8-SP100-90-202505	BN037044.D	19 May 2025 14:26		RC/JU	Ok
9	Q2064-06	RW8-SP201-90-202505	BN037045.D	19 May 2025 15:03		RC/JU	Ok
10	Q2064-07	RW8-SP301-90-202505	BN037046.D	19 May 2025 15:39		RC/JU	Ok
11	Q2064-08	RW8-SP303-90-202505	BN037047.D	19 May 2025 16:15		RC/JU	Ok
12	Q2050-02	BP-VPB-182-GW-60-62	BN037048.D	19 May 2025 16:51		RC/JU	Ok
13	Q2050-04	BP-VPB-182-GW-150-1	BN037049.D	19 May 2025 17:27		RC/JU	Ok
14	Q2050-06	BP-VPB-182-GW-220-2	BN037050.D	19 May 2025 18:03		RC/JU	Ok
15	Q2050-08	BP-VPB-182-EB-20250	BN037051.D	19 May 2025 18:40		RC/JU	Ok
16	Q2051-01	RW7-SP100-20250514	BN037052.D	19 May 2025 19:16		RC/JU	Ok
17	Q2051-02	RW7-SP201-20250514	BN037053.D	19 May 2025 19:52		RC/JU	Ok
18	Q2051-03	RW7-SP302-20250514	BN037054.D	19 May 2025 20:28		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN051925

Review By	Rahul	Review On	5/20/2025 12:04:04 PM
Supervise By	Jagrut	Supervise On	5/20/2025 12:14:27 PM
SubDirectory	BN051925	HP Acquire Method	BNA_N, 8270_HP Processing Method bn051425
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	SP6757 SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6779 SP6740,1ul/100ul sample SP6768		

19	SSTDCCC0.4	SSTDCCC0.4EC	BN037055.D	19 May 2025 21:05		RC/JU	Ok,M
20	DFTPP	DFTPP	BN037056.D	19 May 2025 22:17	Tune Not OK	RC/JU	Not Ok
21	SSTDCCC0.4	SSTDCCC0.4	BN037057.D	19 May 2025 22:53	Tune Not OK	RC/JU	Not Ok
22	PB168032BL	PB168032BL	BN037058.D	19 May 2025 23:30	Tune Not OK	RC/JU	Not Ok
23	Q2051-04	RW7-SP303-20250514	BN037059.D	20 May 2025 00:06	Tune Not OK	RC/JU	Not Ok
24	Q2063-01	RW5-SP100-20250515	BN037060.D	20 May 2025 00:42	Need 2X Dilution, Tune Not OK	RC/JU	Not Ok
25	Q2063-02	RW5-SP201-20250515	BN037061.D	20 May 2025 01:18	Tune Not OK	RC/JU	Not Ok
26	Q2063-03	RW5-SP303-20250515	BN037062.D	20 May 2025 01:54	Tune Not OK	RC/JU	Not Ok
27	PB168032BS	PB168032BS	BN037063.D	20 May 2025 02:30	Tune Not OK	RC/JU	Not Ok
28	PB168032BSD	PB168032BSD	BN037064.D	20 May 2025 03:06	Tune Not OK	RC/JU	Not Ok
29	SSTDCCC0.4	SSTDCCC0.4EC	BN037065.D	20 May 2025 03:42	Tune Not OK	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN052125

Review By	Rahul	Review On	5/21/2025 2:05:15 PM
Supervise By	Jagrut	Supervise On	5/21/2025 4:43:45 PM
SubDirectory	BN052125	HP Acquire Method	BNA_N, 8270_HP Processing Method bn051425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775
CCC	SP6779
Internal Standard/PEM	SP6740,1ul/100ul sample
ICV/I.BLK	SP6768
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037073.D	20 May 2025 14:30		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037074.D	20 May 2025 15:09		RC/JU	Ok
3	PB168032BL	PB168032BL	BN037075.D	20 May 2025 15:45		RC/JU	Ok
4	Q2051-04	RW7-SP303-20250514	BN037076.D	20 May 2025 16:22		RC/JU	Ok
5	Q2063-01	RW5-SP100-20250515	BN037077.D	20 May 2025 16:59		RC/JU	Ok
6	Q2063-02	RW5-SP201-20250515	BN037078.D	20 May 2025 17:35		RC/JU	Ok
7	Q2063-03	RW5-SP303-20250515	BN037079.D	20 May 2025 18:11		RC/JU	Ok
8	PB168032BS	PB168032BS	BN037080.D	20 May 2025 18:47		RC/JU	Ok,M
9	PB168032BSD	PB168032BSD	BN037081.D	20 May 2025 19:23		RC/JU	Ok,M
10	SSTDCCC0.4	SSTDCCC0.4EC	BN037082.D	20 May 2025 19:59		RC/JU	Ok

M : Manual Integration