

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051525

Review By	Mahesh Dadoda	Review On	5/16/2025 11:49:17 AM				
Supervise By	Semsettin Yesilyurt	Supervise On	5/16/2025 11:51:38 AM				
SubDirectory	VY051525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m		
STD. NAME		STD REF.#					
Tune/Reschk		VP133916					
Initial Calibration Stds		VP133918,VP133919,VP133920,VP133921,VP133922,VP133923					
CCC		VP133925					
Internal Standard/PEM		VP131783					
ICV/I.BLK		VP133924					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022252.D	15 May 2025 07:52		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY022253.D	15 May 2025 09:46		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY022254.D	15 May 2025 10:16		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY022255.D	15 May 2025 10:39		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY022256.D	15 May 2025 11:02		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY022257.D	15 May 2025 11:24		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY022258.D	15 May 2025 11:47		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022259.D	15 May 2025 12:34		SY/MD	Ok
9	VSTDICV050	ICVVY051525	VY022260.D	15 May 2025 13:20		SY/MD	Ok,M
10	VY0515SBL01	VY0515SBL01	VY022261.D	15 May 2025 13:44		SY/MD	Ok
11	VY0515SBS01	VY0515SBS01	VY022262.D	15 May 2025 14:07		SY/MD	Ok,M
12	VY0515SBSD01	VY0515SBSD01	VY022263.D	15 May 2025 14:29		SY/MD	Ok,M
13	Q1984-01	OU4-PCS-TC-33-05072	VY022264.D	15 May 2025 15:12	vial-A	SY/MD	Ok
14	Q1984-03	OU4-PCS-TC-34-05072	VY022265.D	15 May 2025 15:35	vial-A	SY/MD	Ok
15	Q1984-05	OU4-PCS-TC-35-05072	VY022266.D	15 May 2025 15:59	vial-A	SY/MD	Ok
16	Q1984-07	OU4-TS-24-050725	VY022267.D	15 May 2025 16:22	vial-A Surrogate fail	SY/MD	ReRun
17	Q1984-09	OU4-TS-25-050725	VY022268.D	15 May 2025 16:46	vial-A Surrogate fail	SY/MD	Not Ok
18	Q1984-11	OU4-TS-26-050725	VY022269.D	15 May 2025 17:09	vial-A Surrogate fail	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051525

Review By	Mahesh Dadoda	Review On	5/16/2025 11:49:17 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/16/2025 11:51:38 AM		
SubDirectory	VY051525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133916			
Initial Calibration Stds		VP133918,VP133919,VP133920,VP133921,VP133922,VP133923			
CCC		VP133925			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133924			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1984-13	OU4-TS-27-050725	VY022270.D	15 May 2025 17:33	vial-A Surrogate fail	SY/MD	Not Ok
20	Q1984-15	OU4-TS-28-050725	VY022271.D	15 May 2025 17:56	vial-A Surrogate fail	SY/MD	ReRun
21	Q1982-04	TP-4	VY022272.D	15 May 2025 18:19	vial-A	SY/MD	Ok
22	Q1982-05	TP-5	VY022273.D	15 May 2025 18:43	vial-A	SY/MD	Ok
23	Q1982-06	TP-6	VY022274.D	15 May 2025 19:06	vial-A	SY/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VY022275.D	15 May 2025 19:52		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051925

Review By	Mahesh Dadoda	Review On	5/21/2025 3:42:46 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/21/2025 3:54:32 PM		
SubDirectory	VY051925	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133960
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133961,VP133962 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022302.D	19 May 2025 07:57		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022303.D	19 May 2025 08:27		SY/MD	Ok,M
3	VY0519SBL01	VY0519SBL01	VY022304.D	19 May 2025 09:00		SY/MD	Ok
4	VY0519SBS01	VY0519SBS01	VY022305.D	19 May 2025 09:36		SY/MD	Ok,M
5	VY0519SBSD01	VY0519SBSD01	VY022306.D	19 May 2025 09:59		SY/MD	Ok,M
6	Q2055-01	OR-02-051525	VY022307.D	19 May 2025 10:41	vial-B Internal Standard Fail	SY/MD	Ok
7	Q2052-03	TP-B-VOC	VY022308.D	19 May 2025 11:04	vial-B	SY/MD	Ok
8	Q2057-03	MH-L-VOC	VY022309.D	19 May 2025 11:28	vial-B	SY/MD	Ok
9	Q2048-03	L2-WC-1-VOC	VY022310.D	19 May 2025 11:51	vial-B	SY/MD	Ok
10	Q2048-07	L2-WC-2-VOC	VY022311.D	19 May 2025 12:15	vial-B	SY/MD	Ok
11	Q2048-11	L2-WC-3-VOC	VY022312.D	19 May 2025 12:38	vial-B	SY/MD	Ok
12	Q2048-15	L2-WC-4-VOC	VY022313.D	19 May 2025 13:02	vial-B	SY/MD	Ok
13	Q2003-02RE	VOCRE	VY022314.D	19 May 2025 13:25	vial-B, Internal standard fail	SY/MD	Confirms
14	Q2062-03	L1-WC-1-VOC	VY022315.D	19 May 2025 13:49	vial-A	SY/MD	Ok
15	Q2062-07	L1-WC-2-VOC	VY022316.D	19 May 2025 14:12	vial-A	SY/MD	Ok
16	Q2062-11	L1-WC-3-VOC	VY022317.D	19 May 2025 14:35	vial-A	SY/MD	Ok
17	Q2062-15	L1-WC-4-VOC	VY022318.D	19 May 2025 14:59	vial-A	SY/MD	Ok
18	Q2062-19	L1-WC-5-VOC	VY022319.D	19 May 2025 15:22	vial-A Internal Standard Fail	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051925

Review By	Mahesh Dadoda	Review On	5/21/2025 3:42:46 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/21/2025 3:54:32 PM		
SubDirectory	VY051925	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133960			
CCC		VP133961,VP133962			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2062-23	L1-WC-6-VOC	VY022320.D	19 May 2025 15:46	vial-A Internal Standard Fail	SY/MD	Not Ok
20	Q2071-03	L1-WC-7-VOC	VY022321.D	19 May 2025 16:09	vial-A	SY/MD	Ok
21	Q2071-07	L1-WC-8-VOC	VY022322.D	19 May 2025 16:33	vial-A	SY/MD	Ok
22	Q2071-11	L1-WC-9-VOC	VY022323.D	19 May 2025 16:56	vial-A Internal Standard Fail	SY/MD	Not Ok
23	Q2071-15	L2-WC-5-VOC	VY022324.D	19 May 2025 17:20	vial-A	SY/MD	Ok
24	Q2021-02	VOC	VY022325.D	19 May 2025 17:43	vial-B	SY/MD	Ok
25	Q2056-01	GVD11	VY022326.D	19 May 2025 18:06	vial-B	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY022327.D	19 May 2025 18:52		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052025

Review By	Mahesh Dadoda	Review On	5/21/2025 3:43:30 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/23/2025 2:22:55 AM		
SubDirectory	VY052025	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133969
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133970,VP133971 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022328.D	20 May 2025 08:09		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022329.D	20 May 2025 08:40		SY/MD	Ok,M
3	VY0520SBL01	VY0520SBL01	VY022330.D	20 May 2025 09:48		SY/MD	Ok
4	VY0520SBS01	VY0520SBS01	VY022331.D	20 May 2025 10:21		SY/MD	Ok,M
5	VY0520SBSD01	VY0520SBSD01	VY022332.D	20 May 2025 10:43		SY/MD	Ok,M
6	Q2062-19	L1-WC-5-VOC	VY022333.D	20 May 2025 11:28	vial-B	SY/MD	Ok
7	Q2062-23	L1-WC-6-VOC	VY022334.D	20 May 2025 11:51	vial-B	SY/MD	Ok
8	Q2071-11	L1-WC-9-VOC	VY022335.D	20 May 2025 12:15	vial-B	SY/MD	Ok
9	Q2056-02	GVD12	VY022336.D	20 May 2025 12:38	vial-B	SY/MD	Ok
10	Q2056-03	GVD13	VY022337.D	20 May 2025 13:01	vial-B	SY/MD	Ok
11	Q2056-04	GVD14	VY022338.D	20 May 2025 13:25	vial-A	SY/MD	Ok
12	Q2056-05	GVD15	VY022339.D	20 May 2025 13:48	vial-A Internal standard fail	SY/MD	Not Ok
13	Q2080-02	PL-HRH-VOC-01	VY022340.D	20 May 2025 14:12	vial-A	SY/MD	Ok
14	Q2080-08	PL-HRH-VOC-02	VY022341.D	20 May 2025 14:35	vial-A	SY/MD	Ok
15	Q2032-02	TP-29	VY022342.D	20 May 2025 14:59	vial-A	SY/MD	Ok
16	Q2075-01	SS-10	VY022343.D	20 May 2025 15:22	vial-A	SY/MD	Ok
17	Q2075-02	SS-910	VY022344.D	20 May 2025 15:45	vial-A	SY/MD	Ok
18	Q2075-03	SS-11	VY022345.D	20 May 2025 16:09	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052025

Review By	Mahesh Dadoda	Review On	5/21/2025 3:43:30 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/23/2025 2:22:55 AM		
SubDirectory	VY052025	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP133969				
CCC	VP133970,VP133971				
Internal Standard/PEM	VP133934				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2084-02	OR-640-VOC-66	VY022346.D	20 May 2025 16:32	vial-A	SY/MD	Ok
20	Q2032-01	TP-11	VY022347.D	20 May 2025 16:56	vial-A Surrogate fail;Need MeOH	SY/MD	Dilution
21	Q2032-05MS	TP-24MS	VY022348.D	20 May 2025 17:19	vial-B	SY/MD	Ok,M
22	Q2032-06MSD	TP-24MSD	VY022349.D	20 May 2025 17:41	vial-B Internal standard fail	SY/MD	ReRun
23	Q2075-04MS	SS-11MS	VY022350.D	20 May 2025 18:04	vial-A	SY/MD	Ok,M
24	Q2075-05MSD	SS-11MSD	VY022351.D	20 May 2025 18:26	vial-A	SY/MD	Ok,M
25	VSTDCCC050	VSTDCCC050EC	VY022352.D	20 May 2025 19:34	ENDCCC failed for com.#06	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052125

Review By	Maresh Dadoda	Review On	5/22/2025 8:21:53 AM
Supervise By	Semsettin Yesilyurt	Supervise On	5/22/2025 8:26:36 AM
SubDirectory	VY052125	HP Acquire Method	MSVOA_Y HP Processing Method 82y051525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133975
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133976,VP133977 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022353.D	21 May 2025 09:02		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022354.D	21 May 2025 09:33		SY/MD	Ok,M
3	VY0521SBL01	VY0521SBL01	VY022355.D	21 May 2025 10:05		SY/MD	Ok
4	VY0521SBS01	VY0521SBS01	VY022356.D	21 May 2025 10:35		SY/MD	Ok,M
5	VY0521SBSD01	VY0521SBSD01	VY022357.D	21 May 2025 10:58		SY/MD	Ok,M
6	Q2056-05	GVD15	VY022358.D	21 May 2025 11:37	vial-B	SY/MD	Ok
7	Q2071-11	L1-WC-9-VOC	VY022359.D	21 May 2025 12:00	Already Analyzed	SY/MD	Not Ok
8	Q2097-01	ETGI-357	VY022360.D	21 May 2025 12:24	vial-A	SY/MD	Ok
9	Q2097-03	RBR200044	VY022361.D	21 May 2025 12:47	vial-A	SY/MD	Ok
10	Q2097-05	RBR251675	VY022362.D	21 May 2025 13:11	vial-A	SY/MD	Ok
11	Q2097-07	VNJ-244	VY022363.D	21 May 2025 13:34	vial-A	SY/MD	Ok
12	Q2097-09	ETGI-290	VY022364.D	21 May 2025 13:57	vial-A	SY/MD	Ok
13	Q2097-11	RT3419	VY022365.D	21 May 2025 14:21	vial-A Internal standard fail	SY/MD	ReRun
14	Q2097-13	RBR251372	VY022366.D	21 May 2025 14:44	vial-A	SY/MD	Ok
15	Q2097-15	72-12013	VY022367.D	21 May 2025 15:08	vial-A	SY/MD	Ok
16	Q2097-17	RT-3888	VY022368.D	21 May 2025 15:31	vial-A	SY/MD	Ok
17	Q2095-03	WCS-TP1-VOC	VY022369.D	21 May 2025 15:55	vial-A	SY/MD	Ok
18	Q2095-07	WCS-TP2-VOC	VY022370.D	21 May 2025 16:18	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052125

Review By	Mahesh Dadoda	Review On	5/22/2025 8:21:53 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/22/2025 8:26:36 AM		
SubDirectory	VY052125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133975			
CCC		VP133976,VP133977			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2093-01	SOIL-DRUM	VY022371.D	21 May 2025 16:42	vial-A	SY/MD	Ok
20	Q2092-01	VNJ-202	VY022372.D	21 May 2025 17:05	vial-A	SY/MD	Ok
21	Q2087-01	NB-07-05202025	VY022373.D	21 May 2025 17:28	vial-A	SY/MD	Ok
22	Q2087-03	NB-08-05202025	VY022374.D	21 May 2025 17:52	vial-A Internal standard fail; Surrogate fail	SY/MD	ReRun
23	Q2101-03	TP-1-MHE-VOC	VY022375.D	21 May 2025 18:15	vial-A Internal standard fail	SY/MD	ReRun
24	Q2104-01	TR-06-052125	VY022376.D	21 May 2025 18:39	vial-A Internal standard fail; Surrogate fail	SY/MD	ReRun
25	Q2102-01	LAW-25-0077	VY022377.D	21 May 2025 19:02	vial-A Internal standard fail	SY/MD	ReRun
26	VIBLK	VIBLK	VY022378.D	21 May 2025 19:26		SY/MD	Ok

M : Manual Integration