

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2056

Client: G Environmental

Analytical Method: SW8260D

Datafile : VY022356.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0521SBS01	Chloromethane	20	26.9	ug/Kg	135	*		70	130	
	Vinyl chloride	20	27.4	ug/Kg	137	*		72	129	
	Bromomethane	20	37.2	ug/Kg	186	*		58	141	
	Chloroethane	20	32.9	ug/Kg	165	*		69	130	
	Tert butyl alcohol	100	100	ug/Kg	100			24	175	
	1,1-Dichloroethene	20	21.1	ug/Kg	106			79	121	
	Acetone	100	94.0	ug/Kg	94			60	131	
	Carbon disulfide	20	20.3	ug/Kg	102			45	154	
	Methyl tert-butyl Ether	20	21.2	ug/Kg	106			77	129	
	Methylene Chloride	20	19.5	ug/Kg	98			56	174	
	trans-1,2-Dichloroethene	20	20.6	ug/Kg	103			80	123	
	1,1-Dichloroethane	20	20.3	ug/Kg	102			82	123	
	2-Butanone	100	96.3	ug/Kg	96			69	131	
	Carbon Tetrachloride	20	19.8	ug/Kg	99			76	129	
	cis-1,2-Dichloroethene	20	21.0	ug/Kg	105			82	123	
	Chloroform	20	20.8	ug/Kg	104			82	125	
	1,1,1-Trichloroethane	20	20.1	ug/Kg	101			80	126	
	Benzene	20	20.2	ug/Kg	101			84	121	
	1,2-Dichloroethane	20	20.4	ug/Kg	102			81	126	
	Trichloroethene	20	20.2	ug/Kg	101			83	122	
	1,2-Dichloropropane	20	19.8	ug/Kg	99			83	122	
	Bromodichloromethane	20	20.3	ug/Kg	102			82	123	
	4-Methyl-2-Pentanone	100	96.4	ug/Kg	96			70	135	
	Toluene	20	19.9	ug/Kg	100			83	122	
	t-1,3-Dichloropropene	20	19.4	ug/Kg	97			78	124	
	cis-1,3-Dichloropropene	20	19.9	ug/Kg	100			81	122	
	1,1,2-Trichloroethane	20	20.8	ug/Kg	104			82	125	
	2-Hexanone	100	95.9	ug/Kg	96			66	138	
	Dibromochloromethane	20	20.7	ug/Kg	104			79	125	
	Tetrachloroethene	20	20.4	ug/Kg	102			83	125	
	Chlorobenzene	20	20.2	ug/Kg	101			84	122	
	Ethyl Benzene	20	19.3	ug/Kg	97			82	124	
	m/p-Xylenes	40	38.7	ug/Kg	97			83	124	
	o-Xylene	20	19.4	ug/Kg	97			83	123	
	Styrene	20	19.1	ug/Kg	96			82	124	
	Bromoform	20	19.3	ug/Kg	97			75	127	
	1,1,2,2-Tetrachloroethane	20	20.9	ug/Kg	104			77	127	