

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2056 SAS No.: Q2056 SDG NO.: Q2056
 Lab File ID: VY022303.D Date Analyzed: 05/19/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:27
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

| | IS4 AREA # | RT # | | | | |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD | 144499 | 13.346 | | | | |
| UPPER LIMIT | 288998 | 13.846 | | | | |
| LOWER LIMIT | 72249.5 | 12.846 | | | | |
| EPA SAMPLE NO. | | | | | | |
| GVD11 | 89910 | 13.35 | | | | |
| VY0519SBL01 | 158015 | 13.35 | | | | |
| VY0519SBS01 | 133674 | 13.35 | | | | |
| VY0519SBSD01 | 128146 | 13.35 | | | | |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.