

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2056 SAS No.: Q2056 SDG NO.: Q2056
 Lab File ID: VY022329.D Date Analyzed: 05/20/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:40
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	120747	13.346				
UPPER LIMIT	241494	13.846				
LOWER LIMIT	60373.5	12.846				
EPA SAMPLE NO.						
GVD12	128914	13.35				
GVD13	107238	13.35				
GVD14	112421	13.35				
VY0520SBL01	124571	13.35				
VY0520SBS01	111469	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.