

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2058

Project: Pre Treatment Plant 2025

RunNo.: LB135805

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.491	0.50	98	90-110	05/16/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.504	0.5	101	90-110	05/16/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.504	0.5	101	90-110	05/16/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2058

Project: Pre Treatment Plant 2025

RunNo.: LB135808

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.506	0.50	101	90-110	05/16/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.495	0.50	99	90-110	05/16/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.500	0.50	100	90-110	05/16/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2058

Project: Pre Treatment Plant 2025

RunNo.: LB135824

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	05/19/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.99	1	99	90-110	05/19/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	05/19/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	05/19/2025
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	05/19/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2058

Project: Pre Treatment Plant 2025

RunNo.: LB135824

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------