

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2058

Project: Pre Treatment Plant 2025

Run No.: LB135797

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135797BS							
TSS	mg/L	550	532		97	1	90-110	05/16/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2058

Project: Pre Treatment Plant 2025

Run No.: LB135805

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135805BS							
Orthophosphate as P	mg/L	0.5	0.49		99	1	90-110	05/16/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2058

Project: Pre Treatment Plant 2025

Run No.: LB135810

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135810BS							
BOD5	mg/L	198	189		96	1	84.6-115.4	05/16/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2058

Project: Pre Treatment Plant 2025

Run No.: LB135833

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135833BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	05/20/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2058

Project: Pre Treatment Plant 2025

Run No.: LB135808

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168047BS							
Phosphorus, Total	mg/L	0.50	0.50		100	1	90-110	05/16/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2058

Project: Pre Treatment Plant 2025

Run No.: LB135824

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168056BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	05/19/2025