

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q2059
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q2059 **SAS No.:** Q2059
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Silver	53.8	50.0	108	90 - 110	P	05/20/2025	17:32	LB135858

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q2059
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q2059 **SAS No.:** Q2059
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Silver	1.08	1.0	108	80 - 120	P	05/20/2025	17:42	LB135858

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q2059
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q2059 **SAS No.:** Q2059
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Silver	483	500	97	90 - 110	P	05/20/2025	17:59	LB135858
CCV02	Silver	495	500	99	90 - 110	P	05/20/2025	18:39	LB135858
CCV03	Silver	488	500	98	90 - 110	P	05/20/2025	18:51	LB135858