

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q2059
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q2059 **SAS No.:** Q2059
ICS Source: EPA **Instrument ID:** P8

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Silver	0.040			-2	2	05/20/2025	17:50	LB135858
ICSAB01	Silver	18.6	18.0	103	15.3	20.7	05/20/2025	17:53	LB135858