

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q2061	OrderDate:	5/16/2025 10:35:00 AM
Client:	Chemtech Consulting Group	Project:	Weekly Storage Blanks
Contact:	QA Officer	Location:	L31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2061-11	PEST-GPC2-BLANK	SOIL	Pesticide-TCL	SFAM_PEST	05/09/25	05/19/25	05/21/25	05/16/25
Q2061-12	PEST-GPC2-BLANK-SP IKE	SOIL	Pesticide-TCL	SFAM_PEST	05/09/25	05/19/25	05/21/25	05/16/25
Q2061-13	PCB-GPC2-BLANK	SOIL	PCB	SFAM_PCB	05/09/25	05/19/25	05/20/25	05/16/25
Q2061-14	PCB-GCP2-BLANK-SPI KE	SOIL	PCB	SFAM_PCB	05/09/25	05/19/25	05/20/25	05/16/25

Hit Summary Sheet
SW-846

SDG No.: Q2061

Order ID: Q2061

Client: Chemtech Consulting Group

Project ID: Weekly Storage Blanks

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : PEST-GPC2-BLANK-SPIKE								
Q2061-12	PEST-GPC2-BLANK· SOIL		gamma-BHC (Lindane)	21.0		0.48	5.10	ug/kg
Q2061-12	PEST-GPC2-BLANK· SOIL		Heptachlor	20.0		0.57	5.10	ug/kg
Q2061-12	PEST-GPC2-BLANK· SOIL		Aldrin	22.0		0.39	5.10	ug/kg
Q2061-12	PEST-GPC2-BLANK· SOIL		Dieldrin	41.0		0.45	9.90	ug/kg
Q2061-12	PEST-GPC2-BLANK· SOIL		Endrin	40.0		0.63	9.90	ug/kg
Q2061-12	PEST-GPC2-BLANK· SOIL		4,4-DDT	36.0		0.63	9.90	ug/kg
Total Concentration:				180.000				



QC SUMMARY



SAMPLE DATA

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	05/09/25
Project:	Weekly Storage Blanks	Date Received:	05/16/25
Client Sample ID:	PEST-GPC2-BLANK	SDG No.:	Q2061
Lab Sample ID:	Q2061-11	Matrix:	SOIL
Analytical Method:	SFAM_PEST	% Solid:	100
Sample Wt/Vol:	10 Units: g	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	5000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	2.0	Injection Volume :	
PH :			
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088678.D	1	05/19/25 10:53	05/21/25 10:01	PB168055

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.39	U	0.39	5.10	ug/kg
319-85-7	beta-BHC	0.39	U	0.39	5.10	ug/kg
319-86-8	delta-BHC	1.40	U	1.40	5.10	ug/kg
58-89-9	gamma-BHC (Lindane)	0.48	U	0.48	5.10	ug/kg
76-44-8	Heptachlor	0.57	U	0.57	5.10	ug/kg
309-00-2	Aldrin	0.39	U	0.39	5.10	ug/kg
1024-57-3	Heptachlor epoxide	0.42	U	0.42	5.10	ug/kg
959-98-8	Endosulfan I	0.39	U	0.39	5.10	ug/kg
60-57-1	Dieldrin	0.45	U	0.45	9.90	ug/kg
72-55-9	4,4-DDE	0.48	U	0.48	9.90	ug/kg
72-20-8	Endrin	0.63	U	0.63	9.90	ug/kg
33213-65-9	Endosulfan II	0.57	U	0.57	9.90	ug/kg
72-54-8	4,4-DDD	0.57	U	0.57	9.90	ug/kg
1031-07-8	Endosulfan Sulfate	0.60	U	0.60	9.90	ug/kg
50-29-3	4,4-DDT	0.63	U	0.63	9.90	ug/kg
72-43-5	Methoxychlor	0.75	U	0.75	51.0	ug/kg
53494-70-5	Endrin ketone	0.45	U	0.45	9.90	ug/kg
7421-93-4	Endrin Aldehyde	0.54	U	0.54	9.90	ug/kg
5103-71-9	cis-Chlordane	0.39	U	0.39	5.10	ug/kg
5103-74-2	trans-Chlordane	0.39	U	0.39	5.10	ug/kg
8001-35-2	Toxaphene	24.0	U	24.0	510	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	30 - 150	%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	30 - 150	%	SPK: 40

Report of Analysis

Client:	Chemtech Consulting Group		Date Collected:	05/09/25	
Project:	Weekly Storage Blanks		Date Received:	05/16/25	
Client Sample ID:	PEST-GPC2-BLANK		SDG No.:	Q2061	
Lab Sample ID:	Q2061-11		Matrix:	SOIL	
Analytical Method:	SFAM_PEST		% Solid:	100	Decanted:
Sample Wt/Vol:	10	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	2.0	PH :			
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088678.D	1	05/19/25 10:53	05/21/25 10:01	PB168055

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052125\
Data File : PD088678.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 10:01
Operator : AR\AJ
Sample : Q2061-11
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 14:39:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050925CLP.M
Quant Title : GC Extractables
QLast Update : Mon May 12 12:22:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds


Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052125\
Data File : PD088678.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 10:01
Operator : AR\AJ
Sample : Q2061-11
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 14:39:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050925CLP.M
Quant Title : GC Extractables
QLast Update : Mon May 12 12:22:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	05/09/25
Project:	Weekly Storage Blanks	Date Received:	05/16/25
Client Sample ID:	PEST-GPC2-BLANK-SPIKE	SDG No.:	Q2061
Lab Sample ID:	Q2061-12	Matrix:	SOIL
Analytical Method:	SFAM_PEST	% Solid:	100
Sample Wt/Vol:	10 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	2.0	PH :	
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088679.D	1	05/19/25 10:53	05/21/25 10:15	PB168055

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.39	U	0.39	5.10	ug/kg
319-85-7	beta-BHC	0.39	U	0.39	5.10	ug/kg
319-86-8	delta-BHC	1.40	U	1.40	5.10	ug/kg
58-89-9	gamma-BHC (Lindane)	21.0		0.48	5.10	ug/kg
76-44-8	Heptachlor	20.0		0.57	5.10	ug/kg
309-00-2	Aldrin	22.0		0.39	5.10	ug/kg
1024-57-3	Heptachlor epoxide	0.42	U	0.42	5.10	ug/kg
959-98-8	Endosulfan I	0.39	U	0.39	5.10	ug/kg
60-57-1	Dieldrin	41.0		0.45	9.90	ug/kg
72-55-9	4,4-DDE	0.48	U	0.48	9.90	ug/kg
72-20-8	Endrin	40.0		0.63	9.90	ug/kg
33213-65-9	Endosulfan II	0.57	U	0.57	9.90	ug/kg
72-54-8	4,4-DDD	0.57	U	0.57	9.90	ug/kg
1031-07-8	Endosulfan Sulfate	0.60	U	0.60	9.90	ug/kg
50-29-3	4,4-DDT	36.0		0.63	9.90	ug/kg
72-43-5	Methoxychlor	0.75	U	0.75	51.0	ug/kg
53494-70-5	Endrin ketone	0.45	U	0.45	9.90	ug/kg
7421-93-4	Endrin Aldehyde	0.54	U	0.54	9.90	ug/kg
5103-71-9	cis-Chlordane	0.39	U	0.39	5.10	ug/kg
5103-74-2	trans-Chlordane	0.39	U	0.39	5.10	ug/kg
8001-35-2	Toxaphene	24.0	U	24.0	510	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	30 - 150	%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	30 - 150	%	SPK: 40

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	05/09/25
Project:	Weekly Storage Blanks	Date Received:	05/16/25
Client Sample ID:	PEST-GPC2-BLANK-SPIKE	SDG No.:	Q2061
Lab Sample ID:	Q2061-12	Matrix:	SOIL
Analytical Method:	SFAM_PEST	% Solid:	100
Sample Wt/Vol:	10 Units: g	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	5000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	2.0	Injection Volume :	
PH :			
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088679.D	1	05/19/25 10:53	05/21/25 10:15	PB168055

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052125\
 Data File : PD088679.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2025 10:15
 Operator : AR\AJ
 Sample : Q2061-12
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEST-GPC2-BLANK-SPIKE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 14:40:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050925CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon May 12 12:22:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds


3) MA	gamma-BHC...	4.332	3.731	96141021	488.6E6	20.911	21.174
4) MA	Heptachlor	4.931	4.085	92332196	479.1E6	19.520	20.099
5) MB	Aldrin	5.273	4.370	95831677	492.0E6	21.482	21.905
13) MA	Dieldrin	6.348	5.513	175.0E6	887.7E6	40.926	41.809
14) MA	Endrin	6.575	5.790	143.7E6	789.5E6	40.110	41.338
17) MA	4,4'-DDT	7.022	6.185	111.5E6	667.1E6	35.610	37.827

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052125\
Data File : PD088679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 10:15
Operator : AR\AJ
Sample : Q2061-12
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 14:40:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050925CLP.M
Quant Title : GC Extractables
QLast Update : Mon May 12 12:22:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PEST-GPC2-BLANK-SPI

Contract: CHEM02

Lab Code: CHEM Case No.: Q2061 SAS No.: Q2061 SDG NO.: Q2061

Lab Sample ID: Q2061-12 Date(s) Analyzed: 05/21/2025 05/21/2025

Instrument ID (1): ECD_D Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PD088679.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
4,4-DDT COLUMN 1 COLUMN 2	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
Aldrin COLUMN 1 COLUMN 2	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

Dieldrin	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 1	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 2	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
Endrin	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 1	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 2	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
gamma-BHC (Lindane)	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 1	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 2	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
Heptachlor	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 1	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 2	1	0	0	0	0		
	2	0	0	0	0		
	3	0	0	0	0		
	4	0	0	0	0		
	5	0	0	0	0		



QC SAMPLE DATA



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/19/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:40
In Date: 05/16/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 05/17/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135796

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q2061-05	SVOC-GPC-BLANK	1	1.00	1.00	2.00	2.00	100.0	
Q2061-06	PEST-GPC-BLANK	2	1.00	1.00	2.00	2.00	100.0	
Q2061-07	PEST-GPC-BLANK-SPIKE	3	1.00	1.00	2.00	2.00	100.0	
Q2061-08	PCB-GPC-BLANK	4	1.00	1.00	2.00	2.00	100.0	
Q2061-09	PCB-GPC-BLANK-SPIKE	5	1.00	1.00	2.00	2.00	100.0	
Q2061-10	SVOC-GPC2-BLANK	6	1.00	1.00	2.00	2.00	100.0	
Q2061-11	PEST-GPC2-BLANK	7	1.00	1.00	2.00	2.00	100.0	
Q2061-12	PEST-GPC2-BLANK-SPIKE	8	1.00	1.00	2.00	2.00	100.0	
Q2061-13	PCB-GPC2-BLANK	9	1.00	1.00	2.00	2.00	100.0	
Q2061-14	PCB-GPC2-BLANK-SPIKE	10	1.00	1.00	2.00	2.00	100.0	
Q2062-01	L1-WC-1	11	1.14	10.72	11.86	10.3	85.4	
Q2062-02	L1-WC-1-EPH	12	1.13	10.14	11.27	9.85	86.0	
Q2062-03	L1-WC-1-VOC	13	1.15	10.73	11.88	9.00	73.2	
Q2062-05	L1-WC-2	14	1.19	10.75	11.94	9.79	80.0	
Q2062-06	L1-WC-2-EPH	15	1.19	10.38	11.57	10.09	85.7	
Q2062-07	L1-WC-2-VOC	16	1.19	10.01	11.2	9.42	82.2	
Q2062-09	L1-WC-3	17	1.18	10.00	11.18	9.34	81.6	
Q2062-10	L1-WC-3-EPH	18	1.19	10.25	11.44	10.3	88.9	
Q2062-11	L1-WC-3-VOC	19	1.13	10.49	11.62	10.08	85.3	
Q2062-13	L1-WC-4	20	1.19	10.07	11.26	10.09	88.4	
Q2062-14	L1-WC-4-EPH	21	1.16	10.53	11.69	10.4	87.7	
Q2062-15	L1-WC-4-VOC	22	1.19	10.15	11.34	9.68	83.6	
Q2062-17	L1-WC-5	23	1.12	9.80	10.92	9.79	88.5	
Q2062-18	L1-WC-5-EPH	24	1.17	10.30	11.47	10.26	88.3	
Q2062-19	L1-WC-5-VOC	25	1.18	10.51	11.69	10.28	86.6	
Q2062-21	L1-WC-6	26	1.15	10.07	11.22	10.16	89.5	
Q2062-22	L1-WC-6-EPH	27	1.19	10.23	11.42	9.32	79.5	
Q2062-23	L1-WC-6-VOC	28	1.14	10.20	11.34	9.68	83.7	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/19/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:40
In Date: 05/16/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 05/17/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135796

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q2066-01	PAINTED-GP-051625	29	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2068-01	LAW-25-0075	46	1.13	9.76	10.89	6.05	50.4	
Q2068-02	LAW-25-0075	47	1.12	10.05	11.17	7.31	61.6	
Q2069-01	BC205321LOK-1-1	48	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2069-02	BC205321LOK-1-2	49	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2069-03	BC205321LOK-2-1	50	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2069-04	BC205321LOK-2-2	51	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2070-01	CONCRETE-PAD-05162025	52	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
Q2071-01	L1-WC-7	30	1.14	10.14	11.28	9.92	86.6	
Q2071-02	L1-WC-7-EPH	31	1.14	10.73	11.87	10.4	86.3	
Q2071-03	L1-WC-7-VOC	32	1.13	10.74	11.87	10.44	86.7	
Q2071-04	L1-WC-7	33	1.14	10.14	11.28	9.92	86.6	
Q2071-05	L1-WC-8	34	1.18	10.42	11.6	9.66	81.4	
Q2071-06	L1-WC-8-EPH	35	1.14	10.56	11.7	10.21	85.9	
Q2071-07	L1-WC-8-VOC	36	1.19	10.44	11.63	9.65	81.0	
Q2071-09	L1-WC-9	37	1.16	10.42	11.58	9.93	84.2	
Q2071-10	L1-WC-9-EPH	38	1.13	10.42	11.55	10.34	88.4	
Q2071-11	L1-WC-9-VOC	39	1.13	10.24	11.37	10.43	90.8	
Q2071-13	L2-WC-5	40	1.19	10.73	11.92	10.25	84.4	
Q2071-14	L2-WC-5-EPH	41	1.14	10.47	11.61	10.2	86.5	
Q2071-15	L2-WC-5-VOC	42	1.17	10.71	11.88	9.79	80.5	
Q2071-17	L2-WC-6	43	1.16	10.24	11.4	9.99	86.2	
Q2071-18	L2-WC-6-EPH	44	1.16	10.34	11.5	10.16	87.0	
Q2071-19	L2-WC-6-VOC	45	1.18	10.40	11.58	10.52	89.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

135796

WorkList Name : %1-051625

WorkList ID : 189559

Department : Wet-Chemistry

Date : 05-16-2025 09:27:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2061-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	05/09/2025	Chemtech -SO
Q2061-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	05/09/2025	Chemtech -SO
Q2061-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	05/09/2025	Chemtech -SO
Q2061-08	PCB-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	05/09/2025	Chemtech -SO
Q2061-09	PCB-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	05/09/2025	Chemtech -SO
Q2061-10	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	05/09/2025	Chemtech -SO
Q2061-11	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	05/09/2025	Chemtech -SO
Q2061-12	PEST-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	05/09/2025	Chemtech -SO
Q2061-13	PCB-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	05/09/2025	Chemtech -SO
Q2061-14	PCB-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	05/09/2025	Chemtech -SO
Q2062-01	L1-WC-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-02	L1-WC-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-03	L1-WC-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-05	L1-WC-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-06	L1-WC-2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-07	L1-WC-2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-09	L1-WC-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-10	L1-WC-3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-11	L1-WC-3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-13	L1-WC-4	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-14	L1-WC-4-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO

Date/Time 05/16/25 16:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 05-16-25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

VB5496

WorkList Name : %1-051625

WorkList ID : 189559

Department : Wet-Chemistry

Date : 05-16-2025 09:27:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2062-15	L1-WC-4-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-17	L1-WC-5	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-18	L1-WC-5-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-19	L1-WC-5-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-21	L1-WC-6	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-22	L1-WC-6-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2062-23	L1-WC-6-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2066-01	PAINTED-GP-051625	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/15/2025	Chemtech -SO
Q2068-01	LAW-25-0075	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2068-02	LAW-25-0075	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/16/2025	Chemtech -SO
Q2069-01	BC205321LOK-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/16/2025	Chemtech -SO
Q2069-02	BC205321LOK-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/16/2025	Chemtech -SO
Q2069-03	BC205321LOK-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/16/2025	Chemtech -SO
Q2069-04	BC205321LOK-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/16/2025	Chemtech -SO
Q2070-01	CONCRETE-PAD-05162025	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	05/16/2025	Chemtech -SO
Q2071-01	L1-WC-7	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-02	L1-WC-7-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-03	L1-WC-7-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-04	L1-WC-7	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-05	L1-WC-8	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-06	L1-WC-8-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO

Date/Time 05-16-25 16:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 05-16-25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

87135796

WorkList Name : %1-051625

WorkList ID : 189559

Department : Wet-Chemistry

Date : 05-16-2025 09:27:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2071-07	L1-WC-8-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-09	L1-WC-9	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-10	L1-WC-9-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-11	L1-WC-9-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-13	L2-WC-5	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-14	L2-WC-5-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-15	L2-WC-5-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-17	L2-WC-6	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-18	L2-WC-6-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO
Q2071-19	L2-WC-6-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/16/2025	Chemtech -SO

Date/Time 05-16-25 16:20

Raw Sample Received by: 38 wpc

Raw Sample Relinquished by: [Signature]

Date/Time 05-16-25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: 38 wpc

SOP ID: M3620C-Florisil cleanup-6, M3640A-GPC cleanup-6

Clean Up SOP #: GPC Cleanup,Florisil **Extraction Start Date :** 05/19/2025

Matrix : Solid **Extraction Start Time :** 10:53

Weigh By: N/A **Extraction By:** N/A **Extraction End Date :** 05/20/2025

Balance check: N/A **Filter By:** N/A **Extraction End Time :** 10:00

Balance ID: N/A **pH Meter ID:** N/A **Concentration By:** EH

pH Strip Lot#: N/A **Hood ID:** 6,7 **Supervisor By :** RUPESH

Extraction Method: ☐ Separatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
GPC Pest Spike	10.0ML	20/40 PPB	PP24421
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
MeCl2/Acetone/1:1	N/A	EP2612
Baked Na2SO4	N/A	EP2614
Methylene Chloride	N/A	E3930
Hexane	N/A	E3934
Florisil	N/A	E3806
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

No Extraction only GPC Check prepared.

KD Bath ID: WATER BATH-1 **Envap ID:** NEVAP-02

KD Bath Temperature: 60 °C **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
5/20/25	RS(E4-lab)	R. Pest/PCB Lab
10:05	Preparation Group	Analysis Group

Analytical Method: M3620C-Florisil cleanup-6, M3640A-GPC cleanup-6

Concentration Date: 05/20/2025

Sample ID	Client Sample ID	Test	g / (mL)	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
Q2061-11	PEST-GPC2-BLANK	Pesticide-TCL	10.0	N/A	Evelyn	RUPESH	5			
Q2061-12	PEST-GPC2-BLANK-SPIKE	Pesticide-TCL	10.0	N/A	Evelyn	RUPESH	5			



168055
10:53 AM

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2061 WorkList ID : 189599 Department : Extraction Date : 05-19-2025 10:43:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2061-10	SVOC-GPC2-BLANK	Solid	SVOC-Chemtech Full -25	Cool 4 deg C	CHEM02	L31	05/09/2025	SFAM_SVOC
Q2061-11	PEST-GPC2-BLANK	Solid	Pesticide-TCL	Cool 4 deg C	CHEM02	L31	05/09/2025	SFAM_PEST
Q2061-12	PEST-GPC2-BLANK-SPIKE	Solid	Pesticide-TCL	Cool 4 deg C	CHEM02	L31	05/09/2025	SFAM_PEST
Q2061-13	PCB-GPC2-BLANK	Solid	PCB	Cool 4 deg C	CHEM02	L31	05/09/2025	SFAM_PCB
Q2061-14	PCB-GCP2-BLANK-SPIKE	Solid	PCB	Cool 4 deg C	CHEM02	L31	05/09/2025	SFAM_PCB

Date/Time 5/19/25 10:45
Raw Sample Received by: RJCEXT-lab
Raw Sample Relinquished by: Cap Sen

Date/Time 5/19/25 11:10
Raw Sample Received by: Cap Sen
Raw Sample Relinquished by: RJCEXT-lab



SHIPPING DOCUMENTS



284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 78-8922

www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q2061

COC Number:

CLIENT INFORMATION

COMPANY: Alliance Technical Group

ADDRESS: 284 Sheffield Street

CITY Mountainside STATE: NJ ZIP: 07092

ATTENTION: QA Officer

PHOI 908-789-8900

FAX: 908-788-9222

PROJECT INFORMATION

PROJECT NAME: Weekly Storage Blanks

PROJECT #: LOCATION:

PROJECT MANAGER:

E-MAIL:

PHONE:

FAX:

BILLING INFORMATION

BILL TO: Same

PO#

ADDRESS:

CITY:

STATE: ZIP:

ATTENTION:

PHONE:

DATA TURNAROUND INFORMATION

FAX: 10 DAYS*

HARD COPY: 10 DAYS*

EDD DAYS*

* TO BE APPROVED BY ALLIANCE

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ RESULTS ONLY (L1) ☐ US EPA CLP (L4)☐ RESULTS + QC (L2) ☐ NYS ASP "A" (L2)☐ NJ REDUCED (L3) ☐ NYS ASP "B" (L4)☐ NJ FULL (L4) ☐ Other☐ EDD Format

ANALYSIS

VOC-Drinking H2O

SVOC-Full+25-8270

PESTICIDE-TCL

PCB

1

2

3

4

5

6

7

8

9

PRESERVATIVES

COMMENTS

ALLIANCE
SAMPLE
IDPROJECT
SAMPLE IDENTIFICATIONSAMPLE
MATRIXSAMPLE
TYPE

COMP

GRAB

SAMPLE
COLLECTION

DATE

TIME

of Bottles

A/E

1

2

3

4

5

6

7

8

9

<-- Specify Preservatives

A-HCl B-HNO3

C-H2SO4 D-NaOH

E-ICE F-Other

1.

PEST-GPC2-BLANK

SO

X

1

X

2.

PEST-GPC2-BLANK-SPIKE

SO

X

1

X

3.

PCB-GPC2-BLANK

SO

X

1

X

4.

PCB-GPC2-BLANK-SPIKE

SO

X

1

X

5.

6.

7.

8.

9.

10.

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER

DATE/TIME

RECEIVED BY

1. Sam

05/16/25

1.

RELINQUISHED BY

DATE/TIME

RECEIVED BY

2.

DATE/TIME

RECEIVED BY

RELINQUISHED BY

DATE/TIME

RECEIVED FOR LAB BY

3.

DATE/TIME

RECEIVED BY

Conditions of bottles or coolers at receipt:

☐ Compliant☐ Non Compliant☐ Cooler Temp 3.0

MeOH extraction requires an additional 4oz. Jar for percent solid

☐ Ice in Cooler?

Comments:

Page 2 of 2

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ OvernightALLIANCE: ☐ Picked Up ☐ Overnight

Shipment Complete

☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT

YELLOW - ALLIANCE COPY

PINK - SAMPLER COPY



284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 78-8922

www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q2061

COC Number:

CLIENT INFORMATION

PROJECT INFORMATION

BILLING INFORMATION

COMPANY: Alliance Technical Group

ADDRESS: 284 Sheffield Street

CITY Mountainside STATE: NJ ZIP: 07092

ATTENTION: QA Officer

PHO: 908-789-8900

FAX: 908-788-9222

PROJECT NAME: Weekly Storage Blanks

PROJECT #: LOCATION:

PROJECT MANAGER:

E-MAIL:

PHONE:

FAX:

BILL TO: Same

PO#

ADDRESS:

CITY:

STATE: ZIP:

ATTENTION:

PHONE:

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX: 10 DAYS*
HARD COPY: 10 DAYS*
EDD DAYS*

* TO BE APPROVED BY ALLIANCE

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

- ☐ RESULTS ONLY (L1) ☐ US EPA CLP (L4)
☐ RESULTS + QC (L2) ☐ NYS ASP "A" (L2)
☐ NJ REDUCED (L3) ☐ NYS ASP "B" (L4)
☐ NJ FULL (L4) ☐ Other _____
☐ EDD Format _____

ANALYSIS

VOC-Drinking H2O

SVOC-Full+25-8270

PESTICIDE-TCL

PCB

1

2

3

4

5

6

7

8

9

PRESERVATIVES

COMMENTS

<- Specify Preservatives
A-HCl B-HNO3
C-H2SO4 D-NaOH
E-ICE F-Other

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles	A/E	1	2	3	4	5	6	7	8	9	
			COMP	GRAB	DATE	TIME												
1.	Storage-Blank-SOIL-REF-6	AQ		X			2	X										
2.	Storage-Blank-WATER-REF-5	AQ		X			2	X										
3.	Storage-Blank-WATER-REF-4	AQ		X			2	X										
4.	Storage-Blank-SAMPLE-MANAGEMENT	AQ		X			2	X										
5.	SVOC-GPC-BLANK	SO		X			1		X									
6.	PEST-GPC-BLANK	SO		X			1			X								
7.	PEST-GPC-BLANK-SPIKE	SO		X			1			X								
8.	PCB-GPC-BLANK	SO		X			1				X							
9.	PCB-GPC-BLANK-SPIKE	SO		X			1				X							
10.	SVOC-GPC2-BLANK	SO		X			1		X									

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp <u>3.0</u> MeOH extraction requires an additional 4oz. Jar for percent solid ☐ Ice in Cooler?: _____
1. <u>Sam</u>	<u>05/16/25</u>	1.	Comments:
RELINQUISHED BY	DATE/TIME	RECEIVED BY	
2.		2.	
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight ALLIANCE: ☐ Picked Up ☐ Overnight
3.		3.	Shipment Complete ☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT

YELLOW - ALLIANCE COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488