

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2061 SAS No.: Q2061 SDG NO.: Q2061

EPA Sample No.: SSTD020369 Date Analyzed: 05/20/2025

Lab File ID: BN037067.D Time Analyzed: 10:16

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	67034	7.605	305940	10.39	212303	14.26
UPPER LIMIT	134068	8.105	611880	10.887	424606	14.763
LOWER LIMIT	33517	7.105	152970	9.887	106152	13.763
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	69237	7.60	292333	10.39	196489	14.26

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.