

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CHEM02

Lab Code: CHEM

SAS No.: Q2061

SDG NO.: Q2061

Lab File ID: BN037066.D

DFTPP Injection Date: 05/20/2025

Instrument ID: BNA_N

DFTPP Injection Time: 09:37

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	71.9
68	Less than 2.0% of mass 69	0.5 (0.8) 1
69	Mass 69 relative abundance	60.4
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	54.8
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
275	10.0 - 60.0% of mass 198	24.5
365	Greater than 1% of mass 198	4.3
441	Present, but less than mass 443	9.2
442	Greater than 50% of mass 198	54.3
443	15.0 - 24.0% of mass 442	11 (20.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTD020369	SSTDCCC020	BN037067.D	05/20/2025	10:16
SVOC-GPC2-BLANK	Q2061-10	BN037071.D	05/20/2025	12:55
SSTD020370	SSTDCCC020EC	BN037072.D	05/20/2025	13:36