

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q2061 SAS No.: Q2061 SDG NO.: Q2061
 Lab File ID: VX046228.D Date Analyzed: 05/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	65634	12.018				
UPPER LIMIT	131268	12.518				
LOWER LIMIT	32817	11.518				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	48381	12.02				
STORAGE-BLANK-WATER-REF-5	47203	12.02				
STORAGE-BLANK-WATER-REF-4	44003	12.02				
STORAGE-BLANK-SAMPLE-MANAGEMENT	50831	12.02				
VX0516WBL01	51388	12.02				
VX0516WBS01	65057	12.02				
VX0516WBSD01	63717	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.