

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2064

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168071BL		WATER		Batch Number:		PB168071		Prep Date:	05/16/2025	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	05/20/2025	11:02	LB135832

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2064

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168036BL	WATER			Batch Number:		PB168036		Prep Date:	05/16/2025	
	Aluminum	50.0	<50.0	U	40.0	50.0	P	05/19/2025	15:09	LB135823
	Antimony	25.0	<25.0	U	6.25	25.0	P	05/19/2025	15:09	LB135823
	Arsenic	10.0	<10.0	U	7.50	10.0	P	05/19/2025	15:09	LB135823
	Barium	50.0	<50.0	U	12.5	50.0	P	05/19/2025	15:09	LB135823
	Beryllium	3.00	<3.00	U	0.75	3.00	P	05/19/2025	15:09	LB135823
	Cadmium	3.00	<3.00	U	0.75	3.00	P	05/19/2025	15:09	LB135823
	Calcium	1000	<1000	U	250	1000	P	05/19/2025	15:09	LB135823
	Chromium	5.00	<5.00	U	2.50	5.00	P	05/19/2025	15:09	LB135823
	Cobalt	15.0	<15.0	U	3.75	15.0	P	05/19/2025	15:09	LB135823
	Copper	10.0	<10.0	U	8.00	10.0	P	05/19/2025	15:09	LB135823
	Iron	50.0	<50.0	U	40.0	50.0	P	05/19/2025	15:09	LB135823
	Lead	6.00	<6.00	U	4.80	6.00	P	05/19/2025	15:09	LB135823
	Magnesium	1000	<1000	U	250	1000	P	05/19/2025	15:09	LB135823
	Manganese	10.0	<10.0	U	7.50	10.0	P	05/19/2025	15:09	LB135823
	Nickel	20.0	<20.0	U	5.00	20.0	P	05/19/2025	15:09	LB135823
	Potassium	1000	<1000	U	800	1000	P	05/19/2025	15:09	LB135823
	Selenium	10.0	<10.0	U	8.00	10.0	P	05/19/2025	15:09	LB135823
	Silver	5.00	<5.00	U	2.50	5.00	P	05/19/2025	15:09	LB135823
	Sodium	1000	<1000	U	500	1000	P	05/19/2025	15:09	LB135823
	Thallium	20.0	<20.0	U	10.0	20.0	P	05/19/2025	15:09	LB135823
	Vanadium	20.0	<20.0	U	10.0	20.0	P	05/19/2025	15:09	LB135823
	Zinc	20.0	<20.0	U	7.50	20.0	P	05/19/2025	15:09	LB135823