

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2064
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2064 **SAS No.:** Q2064

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168036BS							
Aluminum	ug/L	1000	937		94	86 - 115	P
Antimony	ug/L	400	395		99	88 - 113	P
Arsenic	ug/L	400	386		96	87 - 113	P
Barium	ug/L	100	88.9		89	88 - 113	P
Beryllium	ug/L	100	98.6		99	89 - 112	P
Cadmium	ug/L	100	96.1		96	88 - 113	P
Calcium	ug/L	500	476	J	95	87 - 113	P
Chromium	ug/L	200	191		96	90 - 113	P
Cobalt	ug/L	100	93.3		93	89 - 114	P
Copper	ug/L	150	147		98	86 - 114	P
Iron	ug/L	1500	1380		92	87 - 115	P
Lead	ug/L	500	469		94	86 - 113	P
Magnesium	ug/L	1000	912	J	91	85 - 113	P
Manganese	ug/L	100	94.2		94	90 - 114	P
Nickel	ug/L	250	236		94	88 - 113	P
Potassium	ug/L	5000	4430		89	86 - 114	P
Selenium	ug/L	1000	983		98	83 - 114	P
Silver	ug/L	37.5	35.9		96	84 - 115	P
Sodium	ug/L	1500	1410		94	87 - 115	P
Thallium	ug/L	1000	968		97	85 - 114	P
Vanadium	ug/L	150	139		93	90 - 111	P
Zinc	ug/L	100	97.0		97	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2064

Contract: TETR06

Lab Code: CHEM

Case No.: Q2064

SAS No.: Q2064

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168071BS Mercury	ug/L	4.0	3.61		90	82 - 119	CV