

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2064 SAS No.: Q2064 SDG No.: Q2064

Instrument ID: BNA_N Calibration Date/Time: 05/20/2025 19:59

Lab File ID: BN037082.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.581		3.2	50.0
Fluoranthene-d10	1.097	1.017		-7.3	50.0
2-Fluorophenol	1.048	0.946		-9.7	50.0
Phenol-d6	1.311	1.150		-12.3	50.0
Nitrobenzene-d5	0.436	0.417		-4.4	50.0
2-Fluorobiphenyl	1.832	1.796		-2.0	50.0
2,4,6-Tribromophenol	0.176	0.165		-6.3	50.0
Terphenyl-d14	0.856	0.957		11.8	50.0
1,4-Dioxane	0.491	0.473		-3.7	50.0

All other compounds must meet a minimum RRF of 0.010.