

Surrogate Summary

SW-846

SDG No.: Q2064

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168032BL	PB168032BL	2-Methylnaphthalene-d10	0.4	0.36	90		30	150
		Fluoranthene-d10	0.4	0.34	84		30	150
		Nitrobenzene-d5	0.4	0.32	81		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.42	104		58	132
PB168032BS	PB168032BS	2-Methylnaphthalene-d10	0.4	0.39	96		30	150
		Fluoranthene-d10	0.4	0.33	81		30	150
		Nitrobenzene-d5	0.4	0.32	81		55	111
		2-Fluorobiphenyl	0.4	0.33	83		53	106
		Terphenyl-d14	0.4	0.39	97		58	132
PB168032BSD	PB168032BSD	2-Methylnaphthalene-d10	0.4	0.39	98		30	150
		Fluoranthene-d10	0.4	0.33	83		30	150
		Nitrobenzene-d5	0.4	0.35	88		55	111
		2-Fluorobiphenyl	0.4	0.33	83		53	106
		Terphenyl-d14	0.4	0.42	104		58	132
Q2064-01	RW8-SP100-70-20250514	2-Methylnaphthalene-d10	0.4	0.33	82		30	150
		Fluoranthene-d10	0.4	0.36	89		30	150
		Nitrobenzene-d5	0.4	0.31	76		55	111
		2-Fluorobiphenyl	0.4	0.30	76		53	106
		Terphenyl-d14	0.4	0.48	120		58	132
Q2064-02	RW8-SP201-70-20250514	2-Methylnaphthalene-d10	0.4	0.31	78		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.31	77		55	111
		2-Fluorobiphenyl	0.4	0.33	82		53	106
		Terphenyl-d14	0.4	0.48	119		58	132
Q2064-03	RW8-SP301-70-20250514	2-Methylnaphthalene-d10	0.4	0.31	77		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.29	74		55	111
		2-Fluorobiphenyl	0.4	0.29	73		53	106
		Terphenyl-d14	0.4	0.48	120		58	132
Q2064-04	RW8-SP303-70-20250514	2-Methylnaphthalene-d10	0.4	0.31	78		30	150
		Fluoranthene-d10	0.4	0.40	100		30	150
		Nitrobenzene-d5	0.4	0.31	77		55	111
		2-Fluorobiphenyl	0.4	0.31	77		53	106
		Terphenyl-d14	0.4	0.42	105		58	132
Q2064-05	RW8-SP100-90-20250514	2-Methylnaphthalene-d10	0.4	0.34	84		30	150
		Fluoranthene-d10	0.4	0.39	97		30	150
		Nitrobenzene-d5	0.4	0.30	75		55	111
		2-Fluorobiphenyl	0.4	0.33	83		53	106
		Terphenyl-d14	0.4	0.47	117		58	132
Q2064-06	RW8-SP201-90-20250514	2-Methylnaphthalene-d10	0.4	0.31	76		30	150
		Fluoranthene-d10	0.4	0.37	93		30	150
		Nitrobenzene-d5	0.4	0.29	72		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.47	118		58	132
Q2064-07	RW8-SP301-90-20250514	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.31	78		55	111
		2-Fluorobiphenyl	0.4	0.30	75		53	106
		Terphenyl-d14	0.4	0.47	117		58	132
Q2064-08	RW8-SP303-90-20250514	2-Methylnaphthalene-d10	0.4	0.28	70		30	150

Surrogate Summary

SW-846

SDG No.: Q2064

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2064-08	RW8-SP303-90-20250514	Fluoranthene-d10	0.4	0.34	86		30	150
		Nitrobenzene-d5	0.4	0.28	70		55	111
		2-Fluorobiphenyl	0.4	0.28	71		53	106
		Terphenyl-d14	0.4	0.52	130		58	132