

Surrogate Summary

SW-846

SDG No.: Q2064

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168048BL	PB168048BL	2-Fluorophenol	150	133	88		19	119
		Phenol-d6	150	122	81		10	130
		Nitrobenzene-d5	100	76.1	76		44	120
		2-Fluorobiphenyl	100	76.8	77		44	119
		2,4,6-Tribromophenol	150	145	97		43	140
		Terphenyl-d14	100	90.1	90		50	134
PB168048BS	PB168048BS	2-Fluorophenol	150	122	82		19	119
		Phenol-d6	150	117	78		10	130
		Nitrobenzene-d5	100	69.5	69		44	120
		2-Fluorobiphenyl	100	74.5	74		44	119
		2,4,6-Tribromophenol	150	139	93		43	140
		Terphenyl-d14	100	76.9	77		50	134
PB168048BSD	PB168048BSD	2-Fluorophenol	150	128	85		19	119
		Phenol-d6	150	120	80		10	130
		Nitrobenzene-d5	100	75.8	76		44	120
		2-Fluorobiphenyl	100	78.1	78		44	119
		2,4,6-Tribromophenol	150	146	97		43	140
		Terphenyl-d14	100	79.4	79		50	134
Q2064-01	RW8-SP100-70-20250514	2-Fluorophenol	150	48.9	33		19	119
		Phenol-d6	150	25.8	17		10	130
		Nitrobenzene-d5	100	68.6	69		44	120
		2-Fluorobiphenyl	100	71.1	71		44	119
		2,4,6-Tribromophenol	150	149	100		43	140
		Terphenyl-d14	100	83.3	83		50	134
Q2064-04	RW8-SP303-70-20250514	2-Fluorophenol	150	49.7	33		19	119
		Phenol-d6	150	27.0	18		10	130
		Nitrobenzene-d5	100	67.4	67		44	120
		2-Fluorobiphenyl	100	71.4	71		44	119
		2,4,6-Tribromophenol	150	155	103		43	140
		Terphenyl-d14	100	94.1	94		50	134
Q2064-05	RW8-SP100-90-20250514	2-Fluorophenol	150	52.3	35		19	119
		Phenol-d6	150	28.7	19		10	130
		Nitrobenzene-d5	100	70.0	70		44	120
		2-Fluorobiphenyl	100	70.3	70		44	119
		2,4,6-Tribromophenol	150	161	108		43	140
		Terphenyl-d14	100	97.0	97		50	134
Q2064-08	RW8-SP303-90-20250514	2-Fluorophenol	150	48.5	32		19	119
		Phenol-d6	150	26.5	18		10	130
		Nitrobenzene-d5	100	67.0	67		44	120
		2-Fluorobiphenyl	100	71.7	72		44	119
		2,4,6-Tribromophenol	150	159	106		43	140
		Terphenyl-d14	100	95.6	96		50	134