

Continuing Calibration Report for SequenceID : FC052025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068930.D

Aliphatic C9-C12	5307130.000	60.000	3.341	6.640	88452.167	99133.167	10.774
Aliphatic C12-C16	3329114.000	40.000	6.641	10.040	83227.850	90984.404	8.525
Aliphatic C16-C21	4663876.000	60.000	10.041	13.408	77731.267	85760.459	9.362
Aliphatic C21-C28	6018133.000	80.000	13.409	17.070	75226.663	83091.075	9.465
Aliphatic C28-C40	9163208.000	120.000	17.071	22.065	76360.067	84622.692	9.764
Aliphatic EPH	28481461.000	360.000	3.341	22.065	79115.169	87597.230	9.683