

Continuing Calibration Report for SequenceID : FC052025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068938.D

Aliphatic C9-C12	5276650.000	60.000	3.341	6.640	87944.167	99133.167	11.287
Aliphatic C12-C16	3356370.000	40.000	6.641	10.040	83909.250	90984.404	7.776
Aliphatic C16-C21	4751063.000	60.000	10.041	13.408	79184.383	85760.459	7.668
Aliphatic C21-C28	6466121.000	80.000	13.409	17.070	80826.513	83091.075	2.725
Aliphatic C28-C40	11636358.000	120.000	17.071	22.065	96969.650	84622.692	-14.591
Aliphatic EPH	31486562.000	360.000	3.341	22.065	87462.672	87597.230	0.154