

Continuing Calibration Report for SequenceID : FD052025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049350.D

Aromatic C10-C12	5138070.000	40.000	4.424	6.169	128451.750	128531.556	0.062
Aromatic C12-C16	8580296.000	60.000	6.170	8.799	143004.933	135727.245	-5.362
Aromatic C16-C21	10789324.000	80.000	8.800	13.093	134866.550	131120.372	-2.857
Aromatic C21-C36	20634309.000	180.000	13.094	18.511	114635.050	117353.852	2.317
Aromatic EPH	45141999.000	360.000	4.424	18.511	125394.442	124717.278	-0.543