

Continuing Calibration Report for SequenceID : FD052025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049358.D

Aromatic C10-C12	5073449.000	40.000	4.424	6.169	126836.225	128531.556	1.319
Aromatic C12-C16	8368792.000	60.000	6.170	8.799	139479.867	135727.245	-2.765
Aromatic C16-C21	10676502.000	80.000	8.800	13.093	133456.275	131120.372	-1.781
Aromatic C21-C36	20607450.000	180.000	13.094	18.511	114485.833	117353.852	2.444
Aromatic EPH	44726193.000	360.000	4.424	18.511	124239.425	124717.278	0.383