

Continuing Calibration Report for SequenceID : FD052025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049363.D

Aromatic C10-C12	4873320.000	40.000	4.424	6.169	121833.000	128531.556	5.212
Aromatic C12-C16	7942671.000	60.000	6.170	8.799	132377.850	135727.245	2.468
Aromatic C16-C21	10163951.000	80.000	8.800	13.093	127049.388	131120.372	3.105
Aromatic C21-C36	20332750.000	180.000	13.094	18.511	112959.722	117353.852	3.744
Aromatic EPH	43312692.000	360.000	4.424	18.511	120313.033	124717.278	3.531