

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** PSEG
Lab Code: CHEM **Cas No:** Q2067 **SAS No :** Q2067 **SDG No:** Q2067
Sample No : PB168080BS **Datafile:** FD049360.D **Client ID :** PB168080BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	150	75		(40-140)
Aromatic C12-C16	300	260	87		(40-140)
Aromatic C16-C21	500	517	103		(40-140)
Aromatic C21-C36	800	799	100		(40-140)

COMPOUND	SPIKE ADDED	LCS/LCSD CONCENTRATION	% REC	Qual	QC LIMITS
					(40-140)