

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** PSEG
Lab Code: CHEM **Cas No:** Q2067 **SAS No :** Q2067 **SDG No:** Q2067
Sample No : PB168080BSD **Datafile:** FD049361.D **Client ID :** PB168080BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aromatic C10-C12	200	147	74		2 (40-140)	25
Aromatic C12-C16	300	252	84		3.2 (40-140)	25
Aromatic C16-C21	500	499	100		3.2 (40-140)	25
Aromatic C21-C36	800	781	98		2.3 (40-140)	25

COMPOUND	SPIKE ADDED	LCS/LCSD CONCENTRATION	% REC	Qual	QC LIMITS
					(40-140)