

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: First Environment, Inc. **SDG No.:** Q2072
Contract: FIRS02 **Lab Code:** CHEM **Case No.:** Q2072 **SAS No.:** Q2072
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	91300	101000	90	85600	116500	05/21/2025	13:52	LB135868
	Sodium	4.13			0	0	05/21/2025	13:52	LB135868
ICSAB01	Iron	91200	99300	92	84400	114500	05/21/2025	13:56	LB135868
	Sodium	-38.7			0	0	05/21/2025	13:56	LB135868
ICSA01	Iron	109000	101000	108	85600	116500	05/28/2025	13:03	LB135935
	Sodium	8.77			0	0	05/28/2025	13:03	LB135935
ICSAB01	Iron	93500	99300	94	84400	114500	05/28/2025	13:08	LB135935
	Sodium	-21.6			0	0	05/28/2025	13:08	LB135935