

LAB CHRONICLE

OrderID:	Q2072	OrderDate:	5/16/2025 2:45:18 PM
Client:	First Environment, Inc.	Project:	Somerset Hills Memorial Park
Contact:	Ken Cwieka	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2072-01	MW-2	Water	VOCMS Group1	8260-Low	05/16/25		05/16/25	05/16/25
Q2072-02	MW-3	Water	VOCMS Group1	8260-Low	05/16/25		05/16/25	05/16/25
Q2072-03	MW-4	Water	VOCMS Group1	8260-Low	05/16/25		05/16/25	05/16/25
Q2072-04	MW-6	Water	VOCMS Group1	8260-Low	05/16/25		05/16/25	05/16/25
Q2072-05	FB	Water	VOCMS Group1	8260-Low	05/16/25		05/16/25	05/16/25
Q2072-06	TB	Water	VOCMS Group1	8260-Low	05/16/25		05/16/25	05/16/25

Hit Summary Sheet SW-846

SDG No.: Q2072

Client: First Environment, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	MW-4							
Q2072-03	MW-4	Water	Acetone	9.40		1.50	5.00	ug/L
Q2072-03	MW-4	Water	Isopropylbenzene	1.10		0.12	1.00	ug/L
Q2072-03	MW-4	Water	1,2,4-Trimethylbenzene	0.27	J	0.14	1.00	ug/L
Total Voc :				10.8				
Q2072-03	MW-4	Water	Butane, 2-methyl-	* 11.0	J	0	0	ug/L
Q2072-03	MW-4	Water	Butane, 2,3-dimethyl-	* 6.70	J	0	0	ug/L
Q2072-03	MW-4	Water	Pentane, 2,3,4-trimethyl-	* 5.30	J	0	0	ug/L
Q2072-03	MW-4	Water	Butane, 2,2,3,3-tetramethyl-	* 10.2	J	0	0	ug/L
Q2072-03	MW-4	Water	2-Pentene, 4-methyl-, (Z)-	* 6.90	J	0	0	ug/L
Q2072-03	MW-4	Water	Cyclopentene, 4,4-dimethyl-	* 5.10	J	0	0	ug/L
Q2072-03	MW-4	Water	sec-Butylbenzene	* 2.20	J	0.13	1.00	ug/L
Total Tics :				47.4				
Total Concentration:				58.2				



SAMPLE DATA

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	MW-2	SDG No.:	Q2072
Lab Sample ID:	Q2072-01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046248.D	1		05/16/25 17:15	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	MW-2	SDG No.:	Q2072
Lab Sample ID:	Q2072-01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046248.D	1		05/16/25 17:15	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.14	U	0.14	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.6		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	50.9		70 (75) - 130 (124)	102%	SPK: 50
2037-26-5	Toluene-d8	49.8		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.8		70 (77) - 130 (121)	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	60900	5.55			
540-36-3	1,4-Difluorobenzene	122000	6.763			
3114-55-4	Chlorobenzene-d5	116000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	48800	12.018			

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	MW-3	SDG No.:	Q2072
Lab Sample ID:	Q2072-02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046249.D	1		05/16/25 17:39	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	MW-3	SDG No.:	Q2072
Lab Sample ID:	Q2072-02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046249.D	1		05/16/25 17:39	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.14	U	0.14	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.7		70 (74) - 130 (125)	109%	SPK: 50
1868-53-7	Dibromofluoromethane	52.2		70 (75) - 130 (124)	104%	SPK: 50
2037-26-5	Toluene-d8	50.9		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.9		70 (77) - 130 (121)	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	60500	5.544			
540-36-3	1,4-Difluorobenzene	119000	6.757			
3114-55-4	Chlorobenzene-d5	112000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	48400	12.018			

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	MW-4	SDG No.:	Q2072
Lab Sample ID:	Q2072-03	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046250.D	1		05/16/25 18:02	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	9.40		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	MW-4	SDG No.:	Q2072
Lab Sample ID:	Q2072-03	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046250.D	1		05/16/25 18:02	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	1.10		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.27	J	0.14	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.2		70 (74) - 130 (125)	110%	SPK: 50
1868-53-7	Dibromofluoromethane	51.7		70 (75) - 130 (124)	103%	SPK: 50
2037-26-5	Toluene-d8	50.7		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.6		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	59300	5.55			
540-36-3	1,4-Difluorobenzene	119000	6.757			
3114-55-4	Chlorobenzene-d5	114000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	49000	12.018			

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	MW-4	SDG No.:	Q2072
Lab Sample ID:	Q2072-03	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046250.D	1		05/16/25 18:02	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TENTATIVE IDENTIFIED COMPOUNDS						
000078-78-4	Butane, 2-methyl-	11.0	J		1.74	ug/L
000079-29-8	Butane, 2,3-dimethyl-	6.70	J		2.78	ug/L
000691-38-3	2-Pentene, 4-methyl-, (Z)-	6.90	J		2.87	ug/L
000594-82-1	Butane, 2,2,3,3-tetramethyl-	10.2	J		6.24	ug/L
019037-72-0	Cyclopentene, 4,4-dimethyl-	5.10	J		7.17	ug/L
000565-75-3	Pentane, 2,3,4-trimethyl-	5.30	J		7.98	ug/L
135-98-8	sec-Butylbenzene	2.20	J		11.9	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	MW-6	SDG No.:	Q2072
Lab Sample ID:	Q2072-04	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046251.D	1		05/16/25 18:26	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	MW-6	SDG No.:	Q2072
Lab Sample ID:	Q2072-04	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046251.D	1		05/16/25 18:26	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.14	U	0.14	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	56.5		70 (74) - 130 (125)	113%	SPK: 50
1868-53-7	Dibromofluoromethane	52.9		70 (75) - 130 (124)	106%	SPK: 50
2037-26-5	Toluene-d8	51.2		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	56600	5.55			
540-36-3	1,4-Difluorobenzene	113000	6.757			
3114-55-4	Chlorobenzene-d5	110000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	46000	12.024			

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	MW-6	SDG No.:	Q2072
Lab Sample ID:	Q2072-04	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046251.D	1		05/16/25 18:26	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	FB	SDG No.:	Q2072
Lab Sample ID:	Q2072-05	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046246.D	1		05/16/25 16:29	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	FB	SDG No.:	Q2072
Lab Sample ID:	Q2072-05	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046246.D	1		05/16/25 16:29	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.14	U	0.14	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.5		70 (74) - 130 (125)	111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.9		70 (75) - 130 (124)	106%	SPK: 50
2037-26-5	Toluene-d8	50.1		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.3		70 (77) - 130 (121)	97%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	63800	5.55			
540-36-3	1,4-Difluorobenzene	127000	6.763			
3114-55-4	Chlorobenzene-d5	117000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	48600	12.018			

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	FB	SDG No.:	Q2072
Lab Sample ID:	Q2072-05	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046246.D	1		05/16/25 16:29	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	TB	SDG No.:	Q2072
Lab Sample ID:	Q2072-06	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046247.D	1		05/16/25 16:52	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	05/16/25
Project:	Somerset Hills Memorial Park	Date Received:	05/16/25
Client Sample ID:	TB	SDG No.:	Q2072
Lab Sample ID:	Q2072-06	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046247.D	1		05/16/25 16:52	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.14	U	0.14	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	56.0		70 (74) - 130 (125)	112%	SPK: 50
1868-53-7	Dibromofluoromethane	51.3		70 (75) - 130 (124)	103%	SPK: 50
2037-26-5	Toluene-d8	50.7		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.3		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	59300	5.55			
540-36-3	1,4-Difluorobenzene	120000	6.757			
3114-55-4	Chlorobenzene-d5	114000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	49100	12.018			



QC SUMMARY

Surrogate Summary

SDG No.: **Q2072**

Client: **First Environment, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2072-01	MW-2	1,2-Dichloroethane-d4	50	53.6	107	70 (74)	130 (125)
		Dibromofluoromethane	50	50.9	102	70 (75)	130 (124)
		Toluene-d8	50	49.8	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.8	100	70 (77)	130 (121)
Q2072-02	MW-3	1,2-Dichloroethane-d4	50	54.7	109	70 (74)	130 (125)
		Dibromofluoromethane	50	52.2	104	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.9	100	70 (77)	130 (121)
Q2072-03	MW-4	1,2-Dichloroethane-d4	50	55.2	110	70 (74)	130 (125)
		Dibromofluoromethane	50	51.7	103	70 (75)	130 (124)
		Toluene-d8	50	50.7	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101	70 (77)	130 (121)
Q2072-04	MW-6	1,2-Dichloroethane-d4	50	56.5	113	70 (74)	130 (125)
		Dibromofluoromethane	50	52.9	106	70 (75)	130 (124)
		Toluene-d8	50	51.2	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101	70 (77)	130 (121)
Q2072-05	FB	1,2-Dichloroethane-d4	50	55.5	111	70 (74)	130 (125)
		Dibromofluoromethane	50	52.9	106	70 (75)	130 (124)
		Toluene-d8	50	50.1	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.3	97	70 (77)	130 (121)
Q2072-06	TB	1,2-Dichloroethane-d4	50	56.0	112	70 (74)	130 (125)
		Dibromofluoromethane	50	51.3	103	70 (75)	130 (124)
		Toluene-d8	50	50.7	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.3	103	70 (77)	130 (121)
VX0516WBL01	VX0516WBL01	1,2-Dichloroethane-d4	50	55.5	111	70 (74)	130 (125)
		Dibromofluoromethane	50	51.9	104	70 (75)	130 (124)
		Toluene-d8	50	50.7	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.2	102	70 (77)	130 (121)
VX0516WBS01	VX0516WBS01	1,2-Dichloroethane-d4	50	49.9	100	70 (74)	130 (125)
		Dibromofluoromethane	50	51.7	103	70 (75)	130 (124)
		Toluene-d8	50	50.6	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101	70 (77)	130 (121)
VX0516WBSD01	VX0516WBSD01	1,2-Dichloroethane-d4	50	49.5	99	70 (74)	130 (125)
		Dibromofluoromethane	50	51.6	103	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.5	103	70 (77)	130 (121)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2072

Client: First Environment, Inc.

Analytical Method: SW8260-Low

Datafile : VX046231.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0516WBS01	Dichlorodifluoromethane	20	19.2	ug/L	96			40 (69)	160 (116)	
	Chloromethane	20	19.1	ug/L	96			40 (65)	160 (116)	
	Vinyl chloride	20	18.4	ug/L	92			70 (65)	130 (117)	
	Bromomethane	20	18.2	ug/L	91			40 (58)	160 (125)	
	Chloroethane	20	19.8	ug/L	99			40 (56)	160 (128)	
	Trichlorofluoromethane	20	19.7	ug/L	99			40 (73)	160 (115)	
	1,1,2-Trichlorotrifluoroethane	20	19.4	ug/L	97			70 (80)	130 (112)	
	1,1-Dichloroethene	20	18.6	ug/L	93			70 (74)	130 (110)	
	Acetone	100	99.0	ug/L	99			40 (60)	160 (125)	
	Carbon disulfide	20	15.7	ug/L	79			40 (64)	160 (112)	
	Methyl tert-butyl Ether	20	20.0	ug/L	100			70 (78)	130 (114)	
	Methyl Acetate	20	23.3	ug/L	117			70 (67)	130 (125)	
	Methylene Chloride	20	18.7	ug/L	94			70 (72)	130 (114)	
	trans-1,2-Dichloroethene	20	18.8	ug/L	94			70 (75)	130 (108)	
	1,1-Dichloroethane	20	20.2	ug/L	101			70 (78)	130 (112)	
	Cyclohexane	20	18.5	ug/L	93			70 (75)	130 (110)	
	2-Butanone	100	100	ug/L	100			40 (65)	160 (122)	
	Carbon Tetrachloride	20	19.7	ug/L	99			70 (77)	130 (113)	
	cis-1,2-Dichloroethene	20	19.5	ug/L	98			70 (77)	130 (110)	
	Bromochloromethane	20	19.1	ug/L	96			70 (70)	130 (124)	
	Chloroform	20	20.2	ug/L	101			70 (79)	130 (113)	
	1,1,1-Trichloroethane	20	20.1	ug/L	101			70 (80)	130 (108)	
	Methylcyclohexane	20	18.5	ug/L	93			70 (72)	130 (115)	
	Benzene	20	20.1	ug/L	101			70 (82)	130 (109)	
	1,2-Dichloroethane	20	20.5	ug/L	103			70 (80)	130 (115)	
	Trichloroethene	20	19.8	ug/L	99			70 (77)	130 (113)	
	1,2-Dichloropropane	20	21.3	ug/L	106			70 (83)	130 (111)	
	Bromodichloromethane	20	20.7	ug/L	104			70 (83)	130 (110)	
	4-Methyl-2-Pentanone	100	110	ug/L	110			40 (74)	160 (118)	
	Toluene	20	20.2	ug/L	101			70 (82)	130 (110)	
	t-1,3-Dichloropropene	20	19.2	ug/L	96			70 (79)	130 (110)	
	cis-1,3-Dichloropropene	20	19.7	ug/L	99			70 (82)	130 (110)	
	1,1,2-Trichloroethane	20	21.2	ug/L	106			70 (83)	130 (112)	
	2-Hexanone	100	110	ug/L	110			40 (73)	160 (117)	
	Dibromochloromethane	20	20.6	ug/L	103			70 (82)	130 (110)	
	1,2-Dibromoethane	20	20.7	ug/L	104			70 (81)	130 (110)	
	Tetrachloroethene	20	20.9	ug/L	104			70 (67)	130 (123)	
	Chlorobenzene	20	19.4	ug/L	97			70 (82)	130 (109)	
	Ethyl Benzene	20	19.9	ug/L	100			70 (83)	130 (109)	
	m/p-Xylenes	40	39.7	ug/L	99			70 (82)	130 (110)	
	o-Xylene	20	20.0	ug/L	100			70 (83)	130 (109)	
	Styrene	20	20.3	ug/L	102			70 (80)	130 (111)	
	Bromoform	20	19.0	ug/L	95			70 (79)	130 (109)	
	Isopropylbenzene	20	20.0	ug/L	100			70 (83)	130 (112)	
	1,1,2,2-Tetrachloroethane	20	20.1	ug/L	101			70 (76)	130 (118)	
	1,2,4-Trimethylbenzene	20	20.1	ug/L	101			70 (85)	130 (111)	
	1,3-Dichlorobenzene	20	19.6	ug/L	98			70 (82)	130 (108)	
	1,4-Dichlorobenzene	20	19.9	ug/L	100			70 (82)	130 (107)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2072

Client: First Environment, Inc.

Analytical Method: SW8260-Low

Datafile : VX046231.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0516WBS01	1,2-Dichlorobenzene	20	20.4	ug/L	102			70 (82)	130 (109)	
	1,2-Dibromo-3-Chloropropane	20	19.6	ug/L	98			40 (68)	160 (112)	
	1,2,4-Trichlorobenzene	20	17.5	ug/L	88			70 (75)	130 (113)	
	1,2,3-Trichlorobenzene	20	18.8	ug/L	94			70 (76)	130 (114)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2072

Client: First Environment, Inc.

Analytical Method: SW8260-Low

Datafile : VX046232.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0516WBSD01	Dichlorodifluoromethane	20	18.5	ug/L	93	3		40 (69)	160 (116)	20 (20)
	Chloromethane	20	17.4	ug/L	87	10		40 (65)	160 (116)	20 (20)
	Vinyl chloride	20	17.4	ug/L	87	6		70 (65)	130 (117)	20 (20)
	Bromomethane	20	18.1	ug/L	91	0		40 (58)	160 (125)	20 (20)
	Chloroethane	20	19.6	ug/L	98	1		40 (56)	160 (128)	20 (20)
	Trichlorofluoromethane	20	19.2	ug/L	96	3		40 (73)	160 (115)	20 (20)
	1,1,2-Trichlorotrifluoroethane	20	19.2	ug/L	96	1		70 (80)	130 (112)	20 (20)
	1,1-Dichloroethene	20	18.1	ug/L	91	2		70 (74)	130 (110)	20 (20)
	Acetone	100	100	ug/L	100	1		40 (60)	160 (125)	20 (20)
	Carbon disulfide	20	15.6	ug/L	78	1		40 (64)	160 (112)	20 (20)
	Methyl tert-butyl Ether	20	19.9	ug/L	100	0		70 (78)	130 (114)	20 (20)
	Methyl Acetate	20	23.5	ug/L	117	0		70 (67)	130 (125)	20 (20)
	Methylene Chloride	20	18.3	ug/L	92	2		70 (72)	130 (114)	20 (20)
	trans-1,2-Dichloroethene	20	18.1	ug/L	91	3		70 (75)	130 (108)	20 (20)
	1,1-Dichloroethane	20	19.7	ug/L	99	2		70 (78)	130 (112)	20 (20)
	Cyclohexane	20	18.3	ug/L	92	1		70 (75)	130 (110)	20 (20)
	2-Butanone	100	110	ug/L	110	10		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	19.2	ug/L	96	3		70 (77)	130 (113)	20 (20)
	cis-1,2-Dichloroethene	20	19.1	ug/L	96	2		70 (77)	130 (110)	20 (20)
	Bromochloromethane	20	19.0	ug/L	95	1		70 (70)	130 (124)	20 (20)
	Chloroform	20	19.8	ug/L	99	2		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	19.5	ug/L	98	3		70 (80)	130 (108)	20 (20)
	Methylcyclohexane	20	18.5	ug/L	93	0		70 (72)	130 (115)	20 (20)
	Benzene	20	20.0	ug/L	100	1		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	20.9	ug/L	104	1		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.7	ug/L	99	0		70 (77)	130 (113)	20 (20)
	1,2-Dichloropropane	20	20.9	ug/L	104	2		70 (83)	130 (111)	20 (20)
	Bromodichloromethane	20	21.2	ug/L	106	2		70 (83)	130 (110)	20 (20)
	4-Methyl-2-Pentanone	100	110	ug/L	110	0		40 (74)	160 (118)	20 (20)
	Toluene	20	20.3	ug/L	102	1		70 (82)	130 (110)	20 (20)
	t-1,3-Dichloropropene	20	19.5	ug/L	98	2		70 (79)	130 (110)	20 (20)
	cis-1,3-Dichloropropene	20	20.3	ug/L	102	3		70 (82)	130 (110)	20 (20)
	1,1,2-Trichloroethane	20	20.9	ug/L	104	2		70 (83)	130 (112)	20 (20)
	2-Hexanone	100	110	ug/L	110	0		40 (73)	160 (117)	20 (20)
	Dibromochloromethane	20	21.1	ug/L	106	3		70 (82)	130 (110)	20 (20)
	1,2-Dibromoethane	20	21.3	ug/L	106	2		70 (81)	130 (110)	20 (20)
	Tetrachloroethene	20	20.2	ug/L	101	3		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	19.5	ug/L	98	1		70 (82)	130 (109)	20 (20)
	Ethyl Benzene	20	19.7	ug/L	99	1		70 (83)	130 (109)	20 (20)
	m/p-Xylenes	40	40.0	ug/L	100	1		70 (82)	130 (110)	20 (20)
	o-Xylene	20	20.1	ug/L	101	1		70 (83)	130 (109)	20 (20)
	Styrene	20	20.3	ug/L	102	0		70 (80)	130 (111)	20 (20)
	Bromoform	20	19.7	ug/L	99	4		70 (79)	130 (109)	20 (20)
	Isopropylbenzene	20	20.4	ug/L	102	2		70 (83)	130 (112)	20 (20)
	1,1,2,2-Tetrachloroethane	20	20.2	ug/L	101	0		70 (76)	130 (118)	20 (20)
	1,2,4-Trimethylbenzene	20	20.4	ug/L	102	1		70 (85)	130 (111)	20 (20)
	1,3-Dichlorobenzene	20	19.9	ug/L	100	2		70 (82)	130 (108)	20 (20)
	1,4-Dichlorobenzene	20	19.6	ug/L	98	2		70 (82)	130 (107)	20 (20)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2072

Client: First Environment, Inc.

Analytical Method: SW8260-Low

Datafile : VX046232.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0516WBSD01	1,2-Dichlorobenzene	20	20.4	ug/L	102	0		70 (82)	130 (109)	20 (20)
	1,2-Dibromo-3-Chloropropane	20	21.2	ug/L	106	8		40 (68)	160 (112)	20 (20)
	1,2,4-Trichlorobenzene	20	19.0	ug/L	95	8		70 (75)	130 (113)	20 (20)
	1,2,3-Trichlorobenzene	20	19.1	ug/L	96	2		70 (76)	130 (114)	20 (20)

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0516WBL01

Lab Name: CHEMTECH

Contract: FIRS02

Lab Code: CHEM Case No.: Q2072

SAS No.: Q2072 SDG NO.: Q2072

Lab File ID: VX046230.D

Lab Sample ID: VX0516WBL01

Date Analyzed: 05/16/2025

Time Analyzed: 10:12

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0516WBS01	VX0516WBS01	VX046231.D	05/16/2025
VX0516WBSD01	VX0516WBSD01	VX046232.D	05/16/2025
FB	Q2072-05	VX046246.D	05/16/2025
TB	Q2072-06	VX046247.D	05/16/2025
MW-2	Q2072-01	VX046248.D	05/16/2025
MW-3	Q2072-02	VX046249.D	05/16/2025
MW-4	Q2072-03	VX046250.D	05/16/2025
MW-6	Q2072-04	VX046251.D	05/16/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: FIRS02
Lab Code: CHEM Case No.: Q2072 SAS No.: Q2072 SDG NO.: Q2072
Lab File ID: VX046038.D BFB Injection Date: 05/05/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:37
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	68.8
175	5.0 - 9.0% of mass 174	5 (7.3) 1
176	95.0 - 101.0% of mass 174	66.7 (97) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC020	VSTDIC020	VX046041.D	05/05/2025	11:35
VSTDIC050	VSTDIC050	VX046042.D	05/05/2025	11:58
VSTDIC100	VSTDIC100	VX046043.D	05/05/2025	12:21
VSTDIC150	VSTDIC150	VX046044.D	05/05/2025	12:45
VSTDIC005	VSTDIC005	VX046046.D	05/05/2025	16:04
VSTDIC001	VSTDIC001	VX046047.D	05/05/2025	16:27

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: FIRS02
Lab Code: CHEM Case No.: Q2072 SAS No.: Q2072 SDG NO.: Q2072
Lab File ID: VX046227.D BFB Injection Date: 05/16/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:47
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.5
75	30.0 - 60.0% of mass 95	56.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	1 (1.5) 1
174	50.0 - 100.0% of mass 95	65.3
175	5.0 - 9.0% of mass 174	5.2 (8) 1
176	95.0 - 101.0% of mass 174	65.3 (99.9) 1
177	5.0 - 9.0% of mass 176	4.4 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046228.D	05/16/2025	09:22
VX0516WBL01	VX0516WBL01	VX046230.D	05/16/2025	10:12
VX0516WBS01	VX0516WBS01	VX046231.D	05/16/2025	10:36
VX0516WBSD01	VX0516WBSD01	VX046232.D	05/16/2025	11:01
FB	Q2072-05	VX046246.D	05/16/2025	16:29
TB	Q2072-06	VX046247.D	05/16/2025	16:52
MW-2	Q2072-01	VX046248.D	05/16/2025	17:15
MW-3	Q2072-02	VX046249.D	05/16/2025	17:39
MW-4	Q2072-03	VX046250.D	05/16/2025	18:02
MW-6	Q2072-04	VX046251.D	05/16/2025	18:26

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: FIRS02
Lab Code: CHEM Case No.: Q2072 SAS No.: Q2072 SDG NO.: Q2072
Lab File ID: VX046228.D Date Analyzed: 05/16/2025
Instrument ID: MSVOA_X Time Analyzed: 09:22
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	94540	5.54	157316	6.75	136812	10.05
UPPER LIMIT	189080	6.038	314632	7.251	273624	10.549
LOWER LIMIT	47270	5.038	78658	6.251	68406	9.549
EPA SAMPLE NO.						
MW-2	60932	5.55	122154	6.76	115589	10.06
MW-3	60476	5.54	119451	6.76	112435	10.06
MW-4	59308	5.55	118728	6.76	114044	10.06
MW-6	56586	5.55	113386	6.76	109564	10.06
FB	63781	5.55	127117	6.76	117342	10.05
TB	59269	5.55	119588	6.76	114347	10.06
VX0516WBL01	63217	5.54	126396	6.76	119684	10.05
VX0516WBS01	89369	5.54	154326	6.76	137068	10.05
VX0516WBSD01	88353	5.54	151068	6.76	134956	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: FIRS02
 Lab Code: CHEM Case No.: Q2072 SAS No.: Q2072 SDG NO.: Q2072
 Lab File ID: VX046228.D Date Analyzed: 05/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	65634	12.018				
UPPER LIMIT	131268	12.518				
LOWER LIMIT	32817	11.518				
EPA SAMPLE NO.						
MW-2	48818	12.02				
MW-3	48415	12.02				
MW-4	48994	12.02				
MW-6	45951	12.02				
FB	48600	12.02				
TB	49075	12.02				
VX0516WBL01	51388	12.02				
VX0516WBS01	65057	12.02				
VX0516WBSD01	63717	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	Somerset Hills Memorial Park	Date Received:	
Client Sample ID:	VX0516WBL01	SDG No.:	Q2072
Lab Sample ID:	VX0516WBL01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046230.D	1		05/16/25 10:12	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	Somerset Hills Memorial Park	Date Received:	
Client Sample ID:	VX0516WBL01	SDG No.:	Q2072
Lab Sample ID:	VX0516WBL01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046230.D	1		05/16/25 10:12	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.14	U	0.14	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.5		70 (74) - 130 (125)	111%	SPK: 50
1868-53-7	Dibromofluoromethane	51.9		70 (75) - 130 (124)	104%	SPK: 50
2037-26-5	Toluene-d8	50.7		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.2		70 (77) - 130 (121)	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	63200	5.544			
540-36-3	1,4-Difluorobenzene	126000	6.757			
3114-55-4	Chlorobenzene-d5	120000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	51400	12.018			



A
B
C
D
E
F
G

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046230.D	1		05/16/25 10:12	VX051625

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	Somerset Hills Memorial Park	Date Received:	
Client Sample ID:	VX0516WBS01	SDG No.:	Q2072
Lab Sample ID:	VX0516WBS01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046231.D	1		05/16/25 10:36	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	19.2		0.22	1.00	ug/L
74-87-3	Chloromethane	19.1		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	18.4		0.26	1.00	ug/L
74-83-9	Bromomethane	18.2		1.40	5.00	ug/L
75-00-3	Chloroethane	19.8		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.7		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.4		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.6		0.23	1.00	ug/L
67-64-1	Acetone	99.0		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	15.7		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.0		0.16	1.00	ug/L
79-20-9	Methyl Acetate	23.3		0.27	1.00	ug/L
75-09-2	Methylene Chloride	18.7		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.8		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.2		0.23	1.00	ug/L
110-82-7	Cyclohexane	18.5		1.50	5.00	ug/L
78-93-3	2-Butanone	100		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.7		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.5		0.19	1.00	ug/L
74-97-5	Bromochloromethane	19.1		0.22	1.00	ug/L
67-66-3	Chloroform	20.2		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.1		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	18.5		0.16	1.00	ug/L
71-43-2	Benzene	20.1		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.5		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.8		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	21.3		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	20.7		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	5.00	ug/L
108-88-3	Toluene	20.2		0.14	1.00	ug/L

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	Somerset Hills Memorial Park	Date Received:	
Client Sample ID:	VX0516WBS01	SDG No.:	Q2072
Lab Sample ID:	VX0516WBS01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046231.D	1		05/16/25 10:36	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.2		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.7		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.2		0.21	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	20.6		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.7		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	20.9		0.23	1.00	ug/L
108-90-7	Chlorobenzene	19.4		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	19.9		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	39.7		0.24	2.00	ug/L
95-47-6	o-Xylene	20.0		0.12	1.00	ug/L
100-42-5	Styrene	20.3		0.15	1.00	ug/L
75-25-2	Bromoform	19.0		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	20.0		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.1		0.26	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	20.1		0.14	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.6		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.9		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.4		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	19.6		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	17.5		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.8		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.9		70 (74) - 130 (125)	100%	SPK: 50
1868-53-7	Dibromofluoromethane	51.7		70 (75) - 130 (124)	103%	SPK: 50
2037-26-5	Toluene-d8	50.6		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	89400	5.544			
540-36-3	1,4-Difluorobenzene	154000	6.757			
3114-55-4	Chlorobenzene-d5	137000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	65100	12.018			

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	Somerset Hills Memorial Park	Date Received:	
Client Sample ID:	VX0516WBS01	SDG No.:	Q2072
Lab Sample ID:	VX0516WBS01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046231.D	1		05/16/25 10:36	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	Somerset Hills Memorial Park	Date Received:	
Client Sample ID:	VX0516WBSD01	SDG No.:	Q2072
Lab Sample ID:	VX0516WBSD01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046232.D	1		05/16/25 11:01	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	18.5		0.22	1.00	ug/L
74-87-3	Chloromethane	17.4		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	17.4		0.26	1.00	ug/L
74-83-9	Bromomethane	18.1		1.40	5.00	ug/L
75-00-3	Chloroethane	19.6		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.2		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.2		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.1		0.23	1.00	ug/L
67-64-1	Acetone	100		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	15.6		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.9		0.16	1.00	ug/L
79-20-9	Methyl Acetate	23.5		0.27	1.00	ug/L
75-09-2	Methylene Chloride	18.3		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.1		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.7		0.23	1.00	ug/L
110-82-7	Cyclohexane	18.3		1.50	5.00	ug/L
78-93-3	2-Butanone	110		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.2		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.1		0.19	1.00	ug/L
74-97-5	Bromochloromethane	19.0		0.22	1.00	ug/L
67-66-3	Chloroform	19.8		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.5		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	18.5		0.16	1.00	ug/L
71-43-2	Benzene	20.0		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.9		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.7		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.9		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	21.2		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	5.00	ug/L
108-88-3	Toluene	20.3		0.14	1.00	ug/L

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	Somerset Hills Memorial Park	Date Received:	
Client Sample ID:	VX0516WBSD01	SDG No.:	Q2072
Lab Sample ID:	VX0516WBSD01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046232.D	1		05/16/25 11:01	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.5		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.3		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.9		0.21	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	21.1		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	21.3		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	20.2		0.23	1.00	ug/L
108-90-7	Chlorobenzene	19.5		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	19.7		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	40.0		0.24	2.00	ug/L
95-47-6	o-Xylene	20.1		0.12	1.00	ug/L
100-42-5	Styrene	20.3		0.15	1.00	ug/L
75-25-2	Bromoform	19.7		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	20.4		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.2		0.26	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	20.4		0.14	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.9		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.6		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.4		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	21.2		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.0		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	19.1		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.5		70 (74) - 130 (125)	99%	SPK: 50
1868-53-7	Dibromofluoromethane	51.6		70 (75) - 130 (124)	103%	SPK: 50
2037-26-5	Toluene-d8	50.4		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.5		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	88400	5.544			
540-36-3	1,4-Difluorobenzene	151000	6.757			
3114-55-4	Chlorobenzene-d5	135000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	63700	12.018			

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	Somerset Hills Memorial Park	Date Received:	
Client Sample ID:	VX0516WBSD01	SDG No.:	Q2072
Lab Sample ID:	VX0516WBSD01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046232.D	1		05/16/25 11:01	VX051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: FIRS02

Lab Code: CHEM Case No.: Q2072 SAS No.: Q2072 SDG No.: Q2072

Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D	
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD				
Dichlorodifluoromethane	0.697	0.864	0.859	0.875	0.639	0.658	0.765	14.6				
Chloromethane	0.727	0.775	0.787	0.791	0.679	0.694	0.742	6.6				
Vinyl Chloride	0.660	0.710	0.727	0.755	0.619	0.673	0.691	7.2				
Bromomethane	0.296	0.326	0.340	0.334	0.305		0.320	5.8				
Chloroethane	0.354	0.378	0.329	0.317	0.368	0.467	0.369	14.4				
Trichlorofluoromethane	1.035	1.068	0.983	0.985	0.990	1.064	1.021	3.9				
1,1,2-Trichlorotrifluoroethane	0.628	0.641	0.629	0.648	0.610	0.633	0.632	2.1				
1,1-Dichloroethene	0.565	0.601	0.607	0.625	0.567	0.594	0.593	3.9				
Acetone	0.361	0.362	0.361	0.370	0.408	0.380	0.374	4.9				
Carbon Disulfide	1.295	1.455	1.522	1.597	1.141	1.423	1.406	11.7				
Methyl tert-butyl Ether	2.044	2.160	2.172	2.239	1.908	1.949	2.079	6.4				
Methyl Acetate	0.814	0.848	0.845	0.875	0.816	1.006	0.867	8.3				
Methylene Chloride	0.689	0.684	0.691	0.691	0.689	0.853	0.716	9.4				
trans-1,2-Dichloroethene	0.573	0.610	0.612	0.622	0.557	0.604	0.596	4.3				
1,1-Dichloroethane	1.233	1.263	1.263	1.286	1.154	1.116	1.219	5.6				
Cyclohexane	1.090	1.128	1.128	1.150	1.059		1.111	3.3				
2-Butanone	0.540	0.555	0.558	0.569	0.539	0.495	0.543	4.8				
Carbon Tetrachloride	0.528	0.558	0.552	0.577	0.505	0.541	0.544	4.6				
cis-1,2-Dichloroethene	0.716	0.737	0.738	0.755	0.642	0.719	0.718	5.5				
Bromochloromethane	0.628	0.578	0.595	0.590	0.553	0.576	0.587	4.3				
Chloroform	1.287	1.296	1.277	1.300	1.199	1.265	1.271	3				
1,1,1-Trichloroethane	1.106	1.131	1.155	1.188	1.013	1.015	1.101	6.6				
Methylcyclohexane	0.596	0.641	0.627	0.658	0.587	0.627	0.623	4.3				
Benzene	1.426	1.474	1.441	1.477	1.337	1.348	1.417	4.3				
1,2-Dichloroethane	0.632	0.627	0.611	0.625	0.594	0.579	0.612	3.5				
Trichloroethene	0.344	0.355	0.345	0.362	0.315	0.324	0.341	5.3				
1,2-Dichloropropane	0.356	0.371	0.368	0.378	0.324	0.317	0.352	7.4				
Bromodichloromethane	0.557	0.577	0.573	0.594	0.498	0.485	0.547	8.2				
4-Methyl-2-Pentanone	0.620	0.634	0.630	0.631	0.555	0.561	0.605	6				
Toluene	0.884	0.898	0.885	0.904	0.838	0.803	0.869	4.5				

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2072</u>
Instrument ID:	<u>MSVOA_X</u>	SAS No.:	<u>Q2072</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>Q2072</u>
GC Column:	<u>DB-624UI</u>	Calibration Date(s):	<u>05/05/2025</u>
ID:	<u>0.18 (mm)</u>	Calibration Time(s):	<u>11:35</u>
			<u>16:27</u>

LAB FILE ID:		RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D		
		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D		
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD
t-1,3-Dichloropropene	0.468	0.528	0.555	0.591	0.406	0.371	0.487	17.9
cis-1,3-Dichloropropene	0.531	0.578	0.602	0.623	0.469	0.423	0.538	14.6
1,1,2-Trichloroethane	0.349	0.354	0.351	0.356	0.337	0.308	0.343	5.3
2-Hexanone	0.466	0.473	0.477	0.473	0.414	0.385	0.448	8.7
Dibromochloromethane	0.378	0.400	0.415	0.431	0.326	0.306	0.376	13.3
1,2-Dibromoethane	0.359	0.373	0.368	0.381	0.333	0.322	0.356	6.5
Tetrachloroethene	0.390	0.375	0.345	0.344	0.323	0.347	0.354	6.8
Chlorobenzene	1.093	1.098	1.085	1.114	1.046	1.131	1.094	2.7
Ethyl Benzene	1.919	2.022	1.979	2.036	1.816	1.803	1.929	5.2
m/p-Xylenes	0.706	0.740	0.721	0.740	0.678	0.648	0.706	5.2
o-Xylene	0.688	0.727	0.706	0.726	0.639	0.642	0.688	5.7
Styrene	1.135	1.219	1.214	1.230	1.012	0.951	1.127	10.6
Bromoform	0.270	0.304	0.312	0.327	0.236	0.234	0.281	14.2
Isopropylbenzene	3.843	4.130	3.876	4.156	3.562	3.789	3.893	5.7
1,1,2,2-Tetrachloroethane	1.315	1.338	1.284	1.345	1.350	1.552	1.364	7
1,2,4-Trimethylbenzene	3.274	3.522	3.335	3.444	3.034	3.150	3.293	5.5
1,3-Dichlorobenzene	1.633	1.701	1.656	1.730	1.558	1.619	1.649	3.7
1,4-Dichlorobenzene	1.629	1.693	1.639	1.722	1.606	1.817	1.684	4.6
1,2-Dichlorobenzene	1.613	1.696	1.634	1.702	1.577	1.710	1.655	3.3
1,2-Dibromo-3-Chloropropane	0.299	0.322	0.329	0.356	0.248	0.259	0.302	13.9
1,2,4-Trichlorobenzene	0.861	0.981	1.035	1.123	0.842	0.862	0.951	12
1,2,3-Trichlorobenzene	0.921	1.019	1.051	1.107	0.846	0.941	0.981	9.7
1,2-Dichloroethane-d4	0.953	0.910	0.930	0.932	0.935		0.932	1.6
Dibromofluoromethane	0.359	0.355	0.364	0.368	0.354		0.360	1.7
Toluene-d8	1.246	1.223	1.266	1.275	1.221		1.246	2
4-Bromofluorobenzene	0.455	0.470	0.500	0.500	0.464		0.478	4.4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: FIRS02

Lab Code: CHEM Case No.: Q2072 SAS No.: Q2072 SDG No.: Q2072

Instrument ID: MSVOA_X Calibration Date/Time: 05/16/2025 09:22

Lab File ID: VX046228.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.765	0.769		0.52	20
Chloromethane	0.742	0.711	0.1	-4.18	20
Vinyl Chloride	0.691	0.646		-6.51	20
Bromomethane	0.320	0.290		-9.38	20
Chloroethane	0.369	0.361		-2.17	20
Trichlorofluoromethane	1.021	1.027		0.59	20
1,1,2-Trichlorotrifluoroethane	0.632	0.632		0	20
1,1-Dichloroethene	0.593	0.561		-5.4	20
Acetone	0.374	0.382		2.14	20
Carbon Disulfide	1.406	1.242		-11.66	20
Methyl tert-butyl Ether	2.079	2.019		-2.89	20
Methyl Acetate	0.867	0.942		8.65	20
Methylene Chloride	0.716	0.652		-8.94	20
trans-1,2-Dichloroethene	0.596	0.571		-4.2	20
1,1-Dichloroethane	1.219	1.184	0.1	-2.87	20
Cyclohexane	1.111	1.034		-6.93	20
2-Butanone	0.543	0.545		0.37	20
Carbon Tetrachloride	0.544	0.555		2.02	20
cis-1,2-Dichloroethene	0.718	0.684		-4.74	20
Bromochloromethane	0.587	0.596		1.53	20
Chloroform	1.271	1.254		-1.34	20
1,1,1-Trichloroethane	1.101	1.073		-2.54	20
Methylcyclohexane	0.623	0.608		-2.41	20
Benzene	1.417	1.418		0.07	20
1,2-Dichloroethane	0.612	0.614		0.33	20
Trichloroethene	0.341	0.341		0	20
1,2-Dichloropropane	0.352	0.373		5.97	20
Bromodichloromethane	0.547	0.579		5.85	20
4-Methyl-2-Pentanone	0.605	0.629		3.97	20
Toluene	0.869	0.864		-0.57	20
t-1,3-Dichloropropene	0.487	0.517		6.16	20
cis-1,3-Dichloropropene	0.538	0.571		6.13	20
1,1,2-Trichloroethane	0.343	0.353		2.91	20
2-Hexanone	0.448	0.468		4.46	20
Dibromochloromethane	0.376	0.402		6.91	20
1,2-Dibromoethane	0.356	0.364		2.25	20
Tetrachloroethene	0.354	0.372		5.09	20
Chlorobenzene	1.094	1.082	0.3	-1.1	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: FIRS02

Lab Code: CHEM Case No.: Q2072 SAS No.: Q2072 SDG No.: Q2072

Instrument ID: MSVOA_X Calibration Date/Time: 05/16/2025 09:22

Lab File ID: VX046228.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.929	1.962		1.71	20
m/p-Xylenes	0.706	0.716		1.42	20
o-Xylene	0.688	0.698		1.45	20
Styrene	1.127	1.198		6.3	20
Bromoform	0.281	0.296	0.1	5.34	20
Isopropylbenzene	3.893	3.974		2.08	20
1,1,2,2-Tetrachloroethane	1.364	1.296	0.3	-4.99	20
1,2,4-Trimethylbenzene	3.293	3.390		2.95	20
1,3-Dichlorobenzene	1.649	1.661		0.73	20
1,4-Dichlorobenzene	1.684	1.673		-0.65	20
1,2-Dichlorobenzene	1.655	1.667		0.73	20
1,2-Dibromo-3-Chloropropane	0.302	0.301		-0.33	20
1,2,4-Trichlorobenzene	0.951	0.961		1.05	20
1,2,3-Trichlorobenzene	0.981	0.965		-1.63	20
1,2-Dichloroethane-d4	0.932	0.895		-3.97	20
Dibromofluoromethane	0.360	0.380		5.56	20
Toluene-d8	1.246	1.276		2.41	20
4-Bromofluorobenzene	0.478	0.486		1.67	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.