

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: FIRS02
 Lab Code: CHEM Case No.: Q2072 SAS No.: Q2072 SDG NO.: Q2072
 Lab File ID: VX046228.D Date Analyzed: 05/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	65634	12.018				
UPPER LIMIT	131268	12.518				
LOWER LIMIT	32817	11.518				
EPA SAMPLE NO.						
MW-2	48818	12.02				
MW-3	48415	12.02				
MW-4	48994	12.02				
MW-6	45951	12.02				
FB	48600	12.02				
TB	49075	12.02				
VX0516WBL01	51388	12.02				
VX0516WBS01	65057	12.02				
VX0516WBSD01	63717	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.