

Surrogate Summary

SW-846

SDG No.: Q2073

Client: G Environmental

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168100BL	PB168100BL	2-Methylnaphthalene-d10	0.4	0.36	90		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.33	83		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.34	86		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.35	88		30 (25)	130 (149)
		Terphenyl-d14	0.4	0.46	115		30 (54)	130 (175)
PB168100BS	PB168100BS	2-Methylnaphthalene-d10	0.4	0.38	94		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.28	69		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.33	83		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.35	86		30 (25)	130 (149)
		Terphenyl-d14	0.4	0.37	92		30 (54)	130 (175)
PB168100BSD	PB168100BSD	2-Methylnaphthalene-d10	0.4	0.36	90		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.29	73		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.35	86		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.39	98		30 (25)	130 (149)
		Terphenyl-d14	0.4	0.38	95		30 (54)	130 (175)
Q2073-01	GDW1	2-Methylnaphthalene-d10	0.4	0.29	73		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.34	84		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.30	74		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.44	111		30 (25)	130 (149)
		Terphenyl-d14	0.4	0.72	179	*	30 (54)	130 (175)
Q2073-02	GDW2	2-Methylnaphthalene-d10	0.4	0.27	67		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.34	86		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.27	68		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.36	89		30 (25)	130 (149)
		Terphenyl-d14	0.4	0.55	137	*	30 (54)	130 (175)