

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2073

Client: G Environmental

Analytical Method: 8270-Modified DataFile: BN037125.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Limits		
							Qual	Qual	Low	High	RPD
PB168100BSD	Benzo(a)anthracene	0.4	0.35	ug/L	88	3			70 (54)	130 (130)	20 (20)
	Benzo(b)fluoranthene	0.4	0.33	ug/L	83	3			70 (65)	130 (121)	20 (20)
	Benzo(k)fluoranthene	0.4	0.36	ug/L	90	6			70 (72)	130 (119)	20 (20)
	Benzo(a)pyrene	0.4	0.37	ug/L	93	3			70 (68)	130 (120)	20 (20)
	Benzo(g,h,i)perylene	0.4	0.43	ug/L	108	7			70 (76)	130 (117)	20 (20)