

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG NO.: Q2073

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/27/2025

Lab File ID: BN037084.D Time Analyzed: 13:23

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2081	7.611	5420	10.39	2883	14.26
UPPER LIMIT	4162	8.111	10840	10.894	5766	14.756
LOWER LIMIT	1040.5	7.111	2710	9.894	1441.5	13.756
EPA SAMPLE NO.						
01 GDW1	1858	7.61	4733	10.39	2639	14.26
02 GDW2	1828	7.61	4851	10.39	2747	14.26

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.